

Data Path : Z:\VOASRV\HPCHEM1\MSVOA F\DATA\VF011619\
 Data File : VF061426.D
 Acq On : 16 Jan 2019 22:05
 Operator : VA/AP
 Sample : K1167-11RE
 Misc : 5.71g/5mL/MSVOA-F/SOIL
 ALS Vial : 36 Sample Multiplier: 1

Instrument :
 MSVOA_F
 ClientSampleId :
 CLOVER-YARD-BRE

Manual Integrations
 APPROVED

MMDadoda
 1/17/2019 1:24:16 PM

Quant Time: Jan 17 03:29:31 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_F\METHODS\82F011519S.M
 Quant Title : SW846 8260
 QLast Update : Wed Jan 16 04:51:59 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	4.85	168	193658	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	5.58	114	367204	50.00	ug/l	0.00
63) Chlorobenzene-d5	9.77	117	300361	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.54	152	115815	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	4.84	65	127590	53.36	ug/l	0.00
Spiked Amount 50.000			Recovery	=	106.72%	
35) Dibromofluoromethane	4.11	113	145348	50.16	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.32%	
50) Toluene-d8	7.54	98	331822	41.80	ug/l	0.00
Spiked Amount 50.000			Recovery	=	83.60%	
62) 4-Bromofluorobenzene	11.40	95	152212	41.04	ug/l	0.00
Spiked Amount 50.000			Recovery	=	82.08%	
Target Compounds						
16) Acetone	2.23	43	34392	71.702	ug/l	97
20) Methylene Chloride	2.24	84	22140m	12.272	ug/l	
25) 2-Butanone	4.32	43	9931	12.450	ug/l #	79
40) Benzene	4.62	78	23215	2.450	ug/l	99
52) Toluene	7.61	92	36922	5.918	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA F\DATA\VF011619\
Data File : VF061426.D
Acq On : 16 Jan 2019 22:05
Operator : VA/AP
Sample : K1167-11RE
Misc : 5.71g/5mL/MSVOA-F/SOIL
ALS Vial : 36 Sample Multiplier: 1

Instrument :
MSVOA_F
ClientSampleId :
CLOVER-YARD-BRE

Manual Integrations
APPROVED

MMDadoda
1/17/2019 1:24:16 PM

Quant Time: Jan 17 03:29:31 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_F\METHODS\82F011519S.M
Quant Title : SW846 8260
QLast Update : Wed Jan 16 04:51:59 2019
Response via : Initial Calibration

