


Data Path : Z:\VOASRV\HPCHEM1\MSVOA F\DATA\VF011619\
 Data File : VF061412.D
 Acq On : 16 Jan 2019 15:01
 Operator : VA/AP
 Sample : BFB
 Misc : 5.00µ/5mL/MSVOA-F/SOIL
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 MSVOA_F
 ClientSampleId :
 BFB

Integration File: RTEINT.P

Method : Z:\VOASRV\HPCHEM1\MSVOA_F\METHODS\82F011519S.M
 Title : SW846 8260
 Last Update : Wed Jan 16 04:51:59 2019



AutoFind: Scans 1091, 1092, 1093; Background Corrected with Scan 1085

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	18.7	25805	PASS
75	95	30	60	44.8	61754	PASS
95	95	100	100	100.0	137744	PASS
96	95	5	9	6.6	9046	PASS
173	174	0.00	2	0.0	0	PASS
174	95	50	100	71.8	98880	PASS
175	174	5	9	7.3	7261	PASS
176	174	95	101	96.9	95818	PASS
177	176	5	9	6.4	6167	PASS