

Data Path : Z:\VOASRV\HPCHEM1\MSVOA F\DATA\VF011819\
 Data File : VF061465.D
 Acq On : 18 Jan 2019 17:57
 Operator : VA/AP
 Sample : K1013-05
 Misc : 4.72g/5mL/MSVOA-F/SOIL
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_F
 ClientSampleId :
 PL-01-011819-C

Manual Integrations
 APPROVED

MMDadoda
 1/21/2019 11:11:54 AM

Quant Time: Jan 19 04:38:33 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_F\METHODS\82F011519S.M
 Quant Title : SW846 8260
 QLast Update : Wed Jan 16 04:51:59 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	4.87	168	214312	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	5.60	114	377784	50.00	ug/l	0.00
63) Chlorobenzene-d5	9.77	117	319772	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.55	152	152343	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	4.85	65	118800	44.90	ug/l	0.01
Spiked Amount 50.000			Recovery	=	89.80%	
35) Dibromofluoromethane	4.11	113	150569	50.51	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.02%	
50) Toluene-d8	7.55	98	363915	44.56	ug/l	0.00
Spiked Amount 50.000			Recovery	=	89.12%	
62) 4-Bromofluorobenzene	11.41	95	165828	43.45	ug/l	0.00
Spiked Amount 50.000			Recovery	=	86.90%	
Target Compounds						
16) Acetone	2.23	43	13877	26.143	ug/l	92
20) Methylene Chloride	2.17	84	14637m	6.598	ug/l	
52) Toluene	7.62	92	27382	4.266	ug/l	98
95) Naphthalene	14.46	128	28746	4.412	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA F\DATA\VF011819\
Data File : VF061465.D
Acq On : 18 Jan 2019 17:57
Operator : VA/AP
Sample : K1013-05
Misc : 4.72g/5mL/MSVOA-F/SOIL
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_F
ClientSampleId :
PL-01-011819-C

Manual Integrations
APPROVED

MMDadoda
1/21/2019 11:11:54 AM

Quant Time: Jan 19 04:38:33 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_F\METHODS\82F011519S.M
Quant Title : SW846 8260
QLast Update : Wed Jan 16 04:51:59 2019
Response via : Initial Calibration

