

Data Path : Z:\VOASRV\HPCHEM1\MSVOA F\DATA\VF012319\
 Data File : VF061490.D
 Acq On : 23 Jan 2019 18:41
 Operator : VA/AP
 Sample : K1223-08
 Misc : 5.38g/5mL/MSVOA-F/SOIL
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_F
 ClientSampleId :
 TS-T

Quant Time: Jan 24 03:42:17 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_F\METHODS\82F011519S.M
 Quant Title : SW846 8260
 QLast Update : Wed Jan 16 04:51:59 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	4.86	168	142138	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	5.59	114	265843	50.00	ug/l	0.00
63) Chlorobenzene-d5	9.77	117	180744	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.55	152	52409	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	4.84	65	93258	53.14	ug/l	0.01
Spiked Amount 50.000			Recovery	=	106.28%	
35) Dibromofluoromethane	4.11	113	107362	51.18	ug/l	0.01
Spiked Amount 50.000			Recovery	=	102.36%	
50) Toluene-d8	7.55	98	214082	37.25	ug/l	0.00
Spiked Amount 50.000			Recovery	=	74.50%	
62) 4-Bromofluorobenzene	11.40	95	70605	26.29	ug/l	0.00
Spiked Amount 50.000			Recovery	=	52.58%	
Target Compounds						
16) Acetone	2.23	43	36197	102.818	ug/l	94
25) 2-Butanone	4.33	43	99118	169.305	ug/l	96
86) p-Isopropyltoluene	12.46	119	13105	3.302	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA F\DATA\VF012319\
Data File : VF061490.D
Acq On : 23 Jan 2019 18:41
Operator : VA/AP
Sample : K1223-08
Misc : 5.38g/5mL/MSVOA-F/SOIL
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_F
ClientSampleId :
TS-T

Quant Time: Jan 24 03:42:17 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_F\METHODS\82F011519S.M
Quant Title : SW846 8260
QLast Update : Wed Jan 16 04:51:59 2019
Response via : Initial Calibration

