


Data Path : Z:\VOASRV\HPCHEM1\MSVOA F\DATA\VF012319\
Data File : VF061477.D
Acq On : 23 Jan 2019 9:52
Operator : VA/AP
Sample : BFB
Misc : 5.00µ/5mL/MSVOA-F/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_F
ClientSampleId :
BFB

Integration File: RTEINT.P

Method : Z:\VOASRV\HPCHEM1\MSVOA_F\METHODS\82F011519S.M
Title : SW846 8260
Last Update : Wed Jan 16 04:51:59 2019



AutoFind: Scans 1090, 1091, 1092; Background Corrected with Scan 1085

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	19.7	24725	PASS
75	95	30	60	46.2	57970	PASS
95	95	100	100	100.0	125432	PASS
96	95	5	9	6.5	8161	PASS
173	174	0.00	2	0.0	0	PASS
174	95	50	100	64.0	80242	PASS
175	174	5	9	7.4	5965	PASS
176	174	95	101	99.5	79850	PASS
177	176	5	9	6.6	5271	PASS