

Data Path : Z:\VOASRV\HPCHEM1\MSVOA F\DATA\VF020719\
 Data File : VF061555.D
 Acq On : 7 Feb 2019 20:06
 Operator : VA/AP
 Sample : K1404-06
 Misc : 5.42g/5mL/MSVOA-F/SOIL
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_F
 ClientSampleId :
 A0C12-SS2-3.5-4.0

Manual Integrations
 APPROVED

MMDadoda
 2/11/2019 12:36:43 PM

Quant Time: Feb 09 05:25:45 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_F\METHODS\82F011519S.M
 Quant Title : SW846 8260
 QLast Update : Wed Jan 16 04:51:59 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	4.86	168	197979	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	5.59	114	341035	50.00	ug/l	0.00
63) Chlorobenzene-d5	9.78	117	307203	50.00	ug/l	0.01
72) 1,4-Dichlorobenzene-d4	12.55	152	148857	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	4.84	65	120357	49.24	ug/l	0.01
Spiked Amount 50.000			Recovery	=	98.48%	
35) Dibromofluoromethane	4.11	113	137767	51.19	ug/l	0.01
Spiked Amount 50.000			Recovery	=	102.38%	
50) Toluene-d8	7.55	98	321625	43.62	ug/l	0.00
Spiked Amount 50.000			Recovery	=	87.24%	
62) 4-Bromofluorobenzene	11.40	95	147225	42.74	ug/l	0.00
Spiked Amount 50.000			Recovery	=	85.48%	
Target Compounds						
20) Methylene Chloride	2.18	84	9184m	3.897	ug/l	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA F\DATA\VF020719\
Data File : VF061555.D
Acq On : 7 Feb 2019 20:06
Operator : VA/AP
Sample : K1404-06
Misc : 5.42g/5mL/MSVOA-F/SOIL
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_F
ClientSampleId :
A0C12-SS2-3.5-4.0

Manual Integrations
APPROVED

MMDadoda
2/11/2019 12:36:43 PM

Quant Time: Feb 09 05:25:45 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_F\METHODS\82F011519S.M
Quant Title : SW846 8260
QLast Update : Wed Jan 16 04:51:59 2019
Response via : Initial Calibration

