

Data Path : Z:\VOASRV\HPCHEM1\MSVOA F\DATA\VF020819\
 Data File : VF061571.D
 Acq On : 8 Feb 2019 21:37
 Operator : VA/AP
 Sample : K1422-01RE
 Misc : 5.74g/5mL/MSVOA-F/SOIL
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_F
 ClientSampleId :
 TEST-PIT-1RE

Quant Time: Feb 09 04:33:37 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_F\METHODS\82F011519S.M
 Quant Title : SW846 8260
 QLast Update : Wed Jan 16 04:51:59 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	0.00	168	0	0.00	ug/l	-4.86
34) 1,4-Difluorobenzene	0.00	114	0	0.00	ug/l	-5.59
63) Chlorobenzene-d5	0.00	117	0	0.00	ug/l	-9.77
72) 1,4-Dichlorobenzene-d4	0.00	152	0	0.00	ug/l	-12.55
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	4.78	65	60	0.00	ug/l	-0.05
Spiked Amount 50.000			Recovery	=	0.00%	
35) Dibromofluoromethane	0.00	113	0	0.00	ug/l	
Spiked Amount 50.000			Recovery	=	0.00%	
50) Toluene-d8	7.66	98	74	0.00	ug/l	0.11
Spiked Amount 50.000			Recovery	=	0.00%	
62) 4-Bromofluorobenzene	0.00	95	0	0.00	ug/l	
Spiked Amount 50.000			Recovery	=	0.00%	

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA F\DATA\VF020819\
Data File : VF061571.D
Acq On : 8 Feb 2019 21:37
Operator : VA/AP
Sample : K1422-01RE
Misc : 5.74g/5mL/MSVOA-F/SOIL
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_F
ClientSampleId :
TEST-PIT-1RE

Quant Time: Feb 09 04:33:37 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_F\METHODS\82F011519S.M
Quant Title : SW846 8260
QLast Update : Wed Jan 16 04:51:59 2019
Response via : Initial Calibration

