

Data Path : Z:\VOASRV\HPCHEM1\MSVOA F\DATA\VF020819\
 Data File : VF061573.D
 Acq On : 8 Feb 2019 22:35
 Operator : VA/AP
 Sample : K1424-01RE
 Misc : 4.42g/5mL/MSVOA-F/SOIL
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_F
 ClientSampleId :
 SS2-RB200035RE

Manual Integrations
 APPROVED

mohammad
 2/12/2019 7:08:34 AM

Quant Time: Feb 11 06:31:27 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_F\METHODS\82F011519S.M
 Quant Title : SW846 8260
 QLast Update : Wed Jan 16 04:51:59 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	4.86	168	55391	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	5.58	114	86105	50.00	ug/l	-0.01
63) Chlorobenzene-d5	9.76	117	51400	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.54	152	15223	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	4.83	65	20667	30.22	ug/l	0.00
Spiked Amount 50.000			Recovery	=	60.44%	
35) Dibromofluoromethane	4.09	113	25814	37.99	ug/l	-0.01
Spiked Amount 50.000			Recovery	=	75.98%	
50) Toluene-d8	7.54	98	54830	29.45	ug/l	0.00
Spiked Amount 50.000			Recovery	=	58.90%	
62) 4-Bromofluorobenzene	11.40	95	16332	18.78	ug/l	0.00
Spiked Amount 50.000			Recovery	=	37.56%	
Target Compounds						
16) Acetone	2.23	43	9757	71.119	ug/l	94
20) Methylene Chloride	2.19	84	3949m	6.967	ug/l	
84) 1,2,4-Trimethylbenzene	12.20	105	2043	1.985	ug/l	76
95) Naphthalene	14.45	128	988	1.518	ug/l #	71

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA F\DATA\VF020819\
Data File : VF061573.D
Acq On : 8 Feb 2019 22:35
Operator : VA/AP
Sample : K1424-01RE
Misc : 4.42g/5mL/MSVOA-F/SOIL
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_F
ClientSampleId :
SS2-RB200035RE

Manual Integrations
APPROVED

mohammad
2/12/2019 7:08:34 AM

Quant Time: Feb 11 06:31:27 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_F\METHODS\82F011519S.M
Quant Title : SW846 8260
QLast Update : Wed Jan 16 04:51:59 2019
Response via : Initial Calibration

