

Data Path : Z:\VOASRV\HPCHEM1\MSVOA F\DATA\VF021219\
 Data File : VF061618.D
 Acq On : 12 Feb 2019 19:36
 Operator : VA/AP
 Sample : K1343-23
 Misc : 6.37g/5mL/MSVOA-F/SOIL
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_F
 ClientSampleId :
 JC-02-021119-L

Quant Time: Feb 13 03:25:42 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_F\METHODS\82F011519S.M
 Quant Title : SW846 8260
 QLast Update : Wed Jan 16 04:51:59 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	4.87	168	203377	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	5.60	114	359471	50.00	ug/l	0.00
63) Chlorobenzene-d5	9.77	117	298663	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.55	152	143989	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	4.85	65	127634	50.83	ug/l	0.01
Spiked Amount 50.000			Recovery	=	101.66%	
35) Dibromofluoromethane	4.12	113	140878	49.67	ug/l	0.01
Spiked Amount 50.000			Recovery	=	99.34%	
50) Toluene-d8	7.55	98	312683	40.23	ug/l	0.00
Spiked Amount 50.000			Recovery	=	80.46%	
62) 4-Bromofluorobenzene	11.41	95	143845	39.61	ug/l	0.00
Spiked Amount 50.000			Recovery	=	79.22%	
Target Compounds						
16) Acetone	2.24	43	19659	39.027	ug/l	95
20) Methylene Chloride	2.18	84	18329m	9.288	ug/l	
80) 1,3,5-Trimethylbenzene	11.82	105	9835	1.041	ug/l	92
84) 1,2,4-Trimethylbenzene	12.21	105	38273	3.932	ug/l	94
89) n-Butylbenzene	12.84	91	12483	1.110	ug/l #	67
95) Naphthalene	14.46	128	16208	2.632	ug/l #	76

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA F\DATA\VF021219\
Data File : VF061618.D
Acq On : 12 Feb 2019 19:36
Operator : VA/AP
Sample : K1343-23
Misc : 6.37g/5mL/MSVOA-F/SOIL
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_F
ClientSampleId :
JC-02-021119-L

Quant Time: Feb 13 03:25:42 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_F\METHODS\82F011519S.M
Quant Title : SW846 8260
QLast Update : Wed Jan 16 04:51:59 2019
Response via : Initial Calibration

