

Data Path : Z:\VOASRV\HPCHEM1\MSVOA F\DATA\VF021219\
 Data File : VF061619.D
 Acq On : 12 Feb 2019 20:05
 Operator : VA/AP
 Sample : K1343-25
 Misc : 5.67g/5mL/MSVOA-F/SOIL
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_F
 ClientSampleId :
 JC-02-021119-M

Quant Time: Feb 13 03:29:25 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_F\METHODS\82F011519S.M
 Quant Title : SW846 8260
 QLast Update : Wed Jan 16 04:51:59 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	4.87	168	205868	50.00	ug/l	0.01
34) 1,4-Difluorobenzene	5.60	114	376341	50.00	ug/l	0.01
63) Chlorobenzene-d5	9.78	117	312998	50.00	ug/l	0.01
72) 1,4-Dichlorobenzene-d4	12.56	152	141864	50.00	ug/l	0.01
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	4.85	65	132363	52.08	ug/l	0.02
Spiked Amount 50.000			Recovery	=	104.16%	
35) Dibromofluoromethane	4.11	113	139206	46.88	ug/l	0.01
Spiked Amount 50.000			Recovery	=	93.76%	
50) Toluene-d8	7.56	98	328614	40.39	ug/l	0.01
Spiked Amount 50.000			Recovery	=	80.78%	
62) 4-Bromofluorobenzene	11.41	95	140077	36.85	ug/l	0.01
Spiked Amount 50.000			Recovery	=	73.70%	
Target Compounds						
16) Acetone	2.24	43	17252	33.834	ug/l	Qvalue 93
20) Methylene Chloride	2.25	84	19932m	10.114	ug/l	
84) 1,2,4-Trimethylbenzene	12.20	105	9658	1.007	ug/l	90
95) Naphthalene	14.46	128	11488	1.894	ug/l #	63

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA F\DATA\VF021219\
Data File : VF061619.D
Acq On : 12 Feb 2019 20:05
Operator : VA/AP
Sample : K1343-25
Misc : 5.67g/5mL/MSVOA-F/SOIL
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_F
ClientSampleId :
JC-02-021119-M

Quant Time: Feb 13 03:29:25 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_F\METHODS\82F011519S.M
Quant Title : SW846 8260
QLast Update : Wed Jan 16 04:51:59 2019
Response via : Initial Calibration

