

Data Path : Z:\VOASRV\HPCHEM1\MSVOA F\DATA\VF021219\
 Data File : VF061620.D
 Acq On : 12 Feb 2019 20:33
 Operator : VA/AP
 Sample : K1343-27
 Misc : 5.56g/5mL/MSVOA-F/SOIL
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_F
 ClientSampleId :
 JC-02-021119-N

Quant Time: Feb 13 03:32:23 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_F\METHODS\82F011519S.M
 Quant Title : SW846 8260
 QLast Update : Wed Jan 16 04:51:59 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	4.87	168	202619	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	5.60	114	358607	50.00	ug/l	0.00
63) Chlorobenzene-d5	9.78	117	314295	50.00	ug/l	0.02
72) 1,4-Dichlorobenzene-d4	12.55	152	156873	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	4.85	65	124923	49.94	ug/l	0.02
Spiked Amount 50.000			Recovery	=	99.88%	
35) Dibromofluoromethane	4.12	113	147071	51.97	ug/l	0.02
Spiked Amount 50.000			Recovery	=	103.94%	
50) Toluene-d8	7.55	98	335087	43.22	ug/l	0.00
Spiked Amount 50.000			Recovery	=	86.44%	
62) 4-Bromofluorobenzene	11.41	95	151982	41.96	ug/l	0.00
Spiked Amount 50.000			Recovery	=	83.92%	
Target Compounds						
16) Acetone	2.24	43	10807	21.534	ug/l	99
20) Methylene Chloride	2.26	84	20934m	10.915	ug/l	
84) 1,2,4-Trimethylbenzene	12.21	105	13893	1.310	ug/l	95
95) Naphthalene	14.46	128	6759	1.007	ug/l #	84

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA F\DATA\VF021219\
Data File : VF061620.D
Acq On : 12 Feb 2019 20:33
Operator : VA/AP
Sample : K1343-27
Misc : 5.56g/5mL/MSVOA-F/SOIL
ALS Vial : 19 Sample Multiplier: 1

Instrument :
MSVOA_F
ClientSampleId :
JC-02-021119-N

Quant Time: Feb 13 03:32:23 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_F\METHODS\82F011519S.M
Quant Title : SW846 8260
QLast Update : Wed Jan 16 04:51:59 2019
Response via : Initial Calibration

