

Data Path : Z:\VOASRV\HPCHEM1\MSVOA F\DATA\VF022219\
 Data File : VF061711.D
 Acq On : 22 Feb 2019 16:30
 Operator : VA/AP
 Sample : K1595-03
 Misc : 5.46g/5mL/MSVOA-F/SOIL
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_F
 ClientSampleId :
 TP-CBF-B11

Manual Integrations
 APPROVED

MMDadoda
 2/27/2019 5:41:05 PM

Quant Time: Feb 23 03:14:17 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_F\METHODS\82F021319S.M
 Quant Title : SW846 8260
 QLast Update : Thu Feb 14 03:27:56 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	4.86	168	302630	50.00	ug/l	-0.03
34) 1,4-Difluorobenzene	5.58	114	490713	50.00	ug/l	-0.03
63) Chlorobenzene-d5	9.76	117	356527	50.00	ug/l	-0.03
72) 1,4-Dichlorobenzene-d4	12.54	152	151577	50.00	ug/l	-0.02
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	4.83	65	112608	50.74	ug/l	-0.03
Spiked Amount 50.000			Recovery	=	101.48%	
35) Dibromofluoromethane	4.09	113	169069	46.40	ug/l	-0.03
Spiked Amount 50.000			Recovery	=	92.80%	
50) Toluene-d8	7.54	98	358929	34.26	ug/l	-0.02
Spiked Amount 50.000			Recovery	=	68.52%	
62) 4-Bromofluorobenzene	11.40	95	139325	32.87	ug/l	-0.02
Spiked Amount 50.000			Recovery	=	65.74%	
Target Compounds						
16) Acetone	2.23	43	37428	78.606	ug/l	Qvalue 100
20) Methylene Chloride	2.23	84	14717m	3.420	ug/l	
25) 2-Butanone	4.33	43	10650	10.240	ug/l	99
29) Tetrahydrofuran	4.07	42	12389	26.208	ug/l	95
58) 2-Chloroethyl Vinyl ether	7.27	63	16879	38.782	ug/l	94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA F\DATA\VF022219\
Data File : VF061711.D
Acq On : 22 Feb 2019 16:30
Operator : VA/AP
Sample : K1595-03
Misc : 5.46g/5mL/MSVOA-F/SOIL
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_F
ClientSampleId :
TP-CBF-B11

Manual Integrations
APPROVED

MMDadoda
2/27/2019 5:41:05 PM

Quant Time: Feb 23 03:14:17 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_F\METHODS\82F021319S.M
Quant Title : SW846 8260
QLast Update : Thu Feb 14 03:27:56 2019
Response via : Initial Calibration

