

Data Path : Z:\VOASRV\HPCHEM1\MSVOA F\DATA\VF022719\
 Data File : VF061769.D
 Acq On : 27 Feb 2019 16:35
 Operator : VA/AP
 Sample : K1645-07
 Misc : 4.44g/5mL/MSVOA-F/SOIL
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_F
 ClientSampleId :
 SR-9-S

Quant Time: Feb 28 00:59:09 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_F\METHODS\82F021319S.M
 Quant Title : SW846 8260
 QLast Update : Thu Feb 14 03:27:56 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	4.86	168	339074	50.00	ug/l	-0.03
34) 1,4-Difluorobenzene	5.59	114	538702	50.00	ug/l	-0.02
63) Chlorobenzene-d5	9.76	117	464506	50.00	ug/l	-0.03
72) 1,4-Dichlorobenzene-d4	12.54	152	276916	50.00	ug/l	-0.02
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	4.83	65	106831	42.96	ug/l	-0.03
Spiked Amount 50.000			Recovery	=	85.92%	
35) Dibromofluoromethane	4.10	113	163534	40.88	ug/l	-0.02
Spiked Amount 50.000			Recovery	=	81.76%	
50) Toluene-d8	7.54	98	435966	37.91	ug/l	-0.02
Spiked Amount 50.000			Recovery	=	75.82%	
62) 4-Bromofluorobenzene	11.40	95	198717	42.71	ug/l	-0.02
Spiked Amount 50.000			Recovery	=	85.42%	
Target Compounds						
16) Acetone	2.23	43	16477	30.885	ug/l #	83
20) Methylene Chloride	2.25	84	20011m	4.717	ug/l	

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA F\DATA\VF022719\
Data File : VF061769.D
Acq On : 27 Feb 2019 16:35
Operator : VA/AP
Sample : K1645-07
Misc : 4.44g/5mL/MSVOA-F/SOIL
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_F
ClientSampleId :
SR-9-S

Quant Time: Feb 28 00:59:09 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_F\METHODS\82F021319S.M
Quant Title : SW846 8260
QLast Update : Thu Feb 14 03:27:56 2019
Response via : Initial Calibration

