

Data Path : Z:\VOASRV\HPCHEM1\MSVOA F\DATA\VF022719\
 Data File : VF061770.D
 Acq On : 27 Feb 2019 17:04
 Operator : VA/AP
 Sample : K1645-11
 Misc : 4.69g/5mL/MSVOA-F/SOIL
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_F
 ClientSampleId :
 SR-9-D

Quant Time: Feb 28 01:07:32 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_F\METHODS\82F021319S.M
 Quant Title : SW846 8260
 QLast Update : Thu Feb 14 03:27:56 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	4.86	168	188302	50.00	ug/l	-0.03
34) 1,4-Difluorobenzene	5.59	114	308979	50.00	ug/l	-0.02
63) Chlorobenzene-d5	9.76	117	245546	50.00	ug/l	-0.03
72) 1,4-Dichlorobenzene-d4	12.54	152	148247	50.00	ug/l	-0.02
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	4.83	65	71253	51.60	ug/l	-0.03
Spiked Amount 50.000			Recovery	=	103.20%	
35) Dibromofluoromethane	4.10	113	102222	44.55	ug/l	-0.02
Spiked Amount 50.000			Recovery	=	89.10%	
50) Toluene-d8	7.54	98	272677	41.34	ug/l	-0.02
Spiked Amount 50.000			Recovery	=	82.68%	
62) 4-Bromofluorobenzene	11.41	95	118865	44.54	ug/l	0.00
Spiked Amount 50.000			Recovery	=	89.08%	
Target Compounds						
16) Acetone	2.23	43	12708	42.894	ug/l #	86
20) Methylene Chloride	2.18	84	9945m	3.943	ug/l	

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA F\DATA\VF022719\
Data File : VF061770.D
Acq On : 27 Feb 2019 17:04
Operator : VA/AP
Sample : K1645-11
Misc : 4.69g/5mL/MSVOA-F/SOIL
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_F
ClientSampleId :
SR-9-D

Quant Time: Feb 28 01:07:32 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_F\METHODS\82F021319S.M
Quant Title : SW846 8260
QLast Update : Thu Feb 14 03:27:56 2019
Response via : Initial Calibration

