

Data Path : Z:\VOASRV\HPCHEM1\MSVOA F\DATA\VF022819\
 Data File : VF061790.D
 Acq On : 28 Feb 2019 16:28
 Operator : VA/AP
 Sample : VF0228SBSD01
 Misc : 5.00g/5mL/MSVOA-F/SOIL
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_F
 ClientSampleId :
 VF0228SBSD01

Quant Time: Mar 01 00:19:24 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_F\METHODS\82F021319S.M
 Quant Title : SW846 8260
 QLast Update : Thu Feb 14 03:27:56 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	4.85	168	292782	50.00	ug/l	-0.03
34) 1,4-Difluorobenzene	5.58	114	428247	50.00	ug/l	-0.03
63) Chlorobenzene-d5	9.76	117	368919	50.00	ug/l	-0.03
72) 1,4-Dichlorobenzene-d4	12.54	152	210655	50.00	ug/l	-0.02

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	4.82	65	104963	48.89	ug/l	-0.03
Spiked Amount			Recovery	=	97.78%	
35) Dibromofluoromethane	4.09	113	150243	47.24	ug/l	-0.03
Spiked Amount			Recovery	=	94.48%	
50) Toluene-d8	7.54	98	445670	48.75	ug/l	-0.03
Spiked Amount			Recovery	=	97.50%	
62) 4-Bromofluorobenzene	11.40	95	169002	45.69	ug/l	-0.02
Spiked Amount			Recovery	=	91.38%	

Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
2) Dichlorodifluoromethane	0.96	85	83840	24.935	ug/l	95
3) Chloromethane	1.10	50	77504	21.970	ug/l	99
4) Vinyl Chloride	1.14	62	78160	23.833	ug/l	96
5) Bromomethane	1.32	94	51657	22.886	ug/l	99
6) Chloroethane	1.38	64	37245	23.045	ug/l	98
7) Trichlorofluoromethane	1.48	101	93304	23.588	ug/l	100
8) Diethyl Ether	1.63	74	20053	20.117	ug/l	89
9) 1,1,2-Trichlorotrifluoroet	1.83	101	65427	22.928	ug/l	95
10) Methyl Iodide	1.90	142	92268	16.927	ug/l	96
11) Tert butyl alcohol	2.57	59	12006	96.623	ug/l	97
12) 1,1-Dichloroethene	1.79	96	57596	22.957	ug/l	96
13) Acrolein	1.99	56	14098	90.275	ug/l	100
14) Allyl chloride	2.09	41	67319	21.867	ug/l	97
15) Acrylonitrile	2.93	53	43581	98.882	ug/l	92
16) Acetone	2.23	43	53899	117.006	ug/l	94
17) Carbon Disulfide	1.83	76	121355	15.388	ug/l	100
18) Methyl Acetate	2.34	43	24031	23.157	ug/l	97
19) Methyl tert-butyl Ether	2.42	73	96903	22.600	ug/l	94
20) Methylene Chloride	2.17	84	20954	6.285	ug/l	95
21) trans-1,2-Dichloroethene	2.30	96	56593	20.750	ug/l	88
22) Diisopropyl ether	2.79	45	168029	22.737	ug/l	99
23) Vinyl Acetate	3.18	43	373678	103.526	ug/l	97
24) 1,1-Dichloroethane	2.86	63	100121	22.388	ug/l	99
25) 2-Butanone	4.32	43	106591	105.933	ug/l	97
26) 2,2-Dichloropropane	3.58	77	48629	25.671	ug/l	97
27) cis-1,2-Dichloroethene	3.46	96	77993	21.787	ug/l	90
28) Bromochloromethane	3.70	49	45724	21.309	ug/l	88
29) Tetrahydrofuran	4.05	42	42868	93.735	ug/l	100
30) Chloroform	3.85	83	114133	22.441	ug/l	96
31) Cyclohexane	3.68	56	70410	14.543	ug/l	94
32) 1,1,1-Trichloroethane	4.08	97	76815	23.057	ug/l	95
36) 1,1-Dichloropropene	4.27	75	92433	21.788	ug/l	91
37) Ethyl Acetate	4.10	43	38112	19.342	ug/l #	99
38) Carbon Tetrachloride	3.98	117	68124	23.256	ug/l	99

Data Path : Z:\VOASRV\HPCHEM1\MSVOA F\DATA\VF022819\
 Data File : VF061790.D
 Acq On : 28 Feb 2019 16:28
 Operator : VA/AP
 Sample : VF0228SBSD01
 Misc : 5.00g/5mL/MSVOA-F/SOIL
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_F
 ClientSampleId :
 VF0228SBSD01

Quant Time: Mar 01 00:19:24 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_F\METHODS\82F021319S.M
 Quant Title : SW846 8260
 QLast Update : Thu Feb 14 03:27:56 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
39) Methylcyclohexane	5.43	83	115553	23.031	ug/l	98
40) Benzene	4.62	78	248083	22.357	ug/l	97
41) Methacrylonitrile	4.72	41	23353	23.175	ug/l	95
42) 1,2-Dichloroethane	4.93	62	60735	22.882	ug/l	93
43) Isopropyl Acetate	6.94	43	45485	19.501	ug/l #	99
44) Trichloroethene	5.48	130	82523	22.998	ug/l	97
45) 1,2-Dichloropropane	6.22	63	57669	21.501	ug/l	96
46) Dibromomethane	6.08	93	37660	22.804	ug/l	86
47) Bromodichloromethane	6.37	83	83785	23.168	ug/l	98
48) Methyl methacrylate	6.70	41	25947	18.486	ug/l	90
49) 1,4-Dioxane	6.70	88	5409	409.939	ug/l #	86
51) 4-Methyl-2-Pentanone	8.25	43	167611	100.572	ug/l	96
52) Toluene	7.61	92	150193	22.318	ug/l	96
53) t-1,3-Dichloropropene	8.26	75	66610	22.130	ug/l	94
54) cis-1,3-Dichloropropene	7.28	75	90366	22.540	ug/l	92
55) 1,1,2-Trichloroethane	8.47	97	41568	21.023	ug/l	94
56) Ethyl methacrylate	8.60	69	46792	21.255	ug/l	96
57) 1,3-Dichloropropane	8.84	76	72196	21.875	ug/l	92
58) 2-Chloroethyl Vinyl ether	7.28	63	38960	102.573	ug/l	99
59) 2-Hexanone	9.48	43	127052	108.757	ug/l	99
60) Dibromochloromethane	8.70	129	59210	21.514	ug/l	99
61) 1,2-Dibromoethane	8.97	107	47621	22.205	ug/l	91
64) Tetrachloroethene	8.14	164	69682	23.108	ug/l	98
65) Chlorobenzene	9.78	112	166818	22.283	ug/l	97
66) 1,1,1,2-Tetrachloroethane	9.91	131	68662	23.406	ug/l	96
67) Ethyl Benzene	9.88	91	295039	23.699	ug/l	96
68) m/p-Xylenes	10.12	106	224726	47.166	ug/l	100
69) o-Xylene	10.70	106	116697	22.953	ug/l	97
70) Styrene	10.78	104	165720	23.041	ug/l	98
71) Bromoform	10.76	173	34350	21.564	ug/l	98
73) Isopropylbenzene	11.11	105	327475	21.834	ug/l	100
74) N-amyl acetate	11.36	43	75700	21.028	ug/l	98
75) 1,1,2,2-Tetrachloroethane	11.69	83	57401	21.363	ug/l	98
76) 1,2,3-Trichloropropane	11.79	75	37392	20.468	ug/l	98
77) Bromobenzene	11.48	156	82967	22.392	ug/l	89
78) n-propylbenzene	11.58	91	401216	23.312	ug/l	95
79) 2-Chlorotoluene	11.71	91	222531	23.087	ug/l	93
80) 1,3,5-Trimethylbenzene	11.81	105	271244	22.628	ug/l	100
81) trans-1,4-Dichloro-2-buten	11.85	75	22603	25.808	ug/l	92
82) 4-Chlorotoluene	11.88	91	222236	22.213	ug/l	91
83) tert-Butylbenzene	12.12	119	288921	21.761	ug/l	98
84) 1,2,4-Trimethylbenzene	12.19	105	271809	22.824	ug/l	93
85) sec-Butylbenzene	12.30	105	385411	22.448	ug/l	98
86) p-Isopropyltoluene	12.45	119	329238	22.849	ug/l	98
87) 1,3-Dichlorobenzene	12.47	146	158070	21.883	ug/l	98
88) 1,4-Dichlorobenzene	12.55	146	153125	22.042	ug/l	97
89) n-Butylbenzene	12.83	91	321149	23.603	ug/l	100
90) Hexachloroethane	12.91	117	77051	22.992	ug/l	99
91) 1,2-Dichlorobenzene	12.93	146	153704	22.371	ug/l	99
92) 1,2-Dibromo-3-Chloropropan	13.63	75	9158	21.497	ug/l	94


Data Path : Z:\VOASRV\HPCHEM1\MSVOA F\DATA\VF022819\
Data File : VF061790.D
Acq On : 28 Feb 2019 16:28
Operator : VA/AP
Sample : VF0228SBSD01
Misc : 5.00g/5mL/MSVOA-F/SOIL
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_F
ClientSampleId :
VF0228SBSD01

Quant Time: Mar 01 00:19:24 2019
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_F\METHODS\82F021319S.M
Quant Title : SW846 8260
QLast Update : Thu Feb 14 03:27:56 2019
Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
93) 1,2,4-Trichlorobenzene	14.19	180	107528	21.834	ug/l	98
94) Hexachlorobutadiene	14.18	225	68545	22.537	ug/l	97
95) Naphthalene	14.45	128	168965	19.610	ug/l	98
96) 1,2,3-Trichlorobenzene	14.60	180	94934	21.003	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed



#1

Pentafluorobenzene

Concen: 50.000 ug/l

RT: 4.85 min Scan# 428

Delta R.T. -0.03 min

Lab File: VF061790.D

Acq: 28 Feb 2019 16:28

Instrument :

MSVOA_F

ClientSampleId :


VF0228SBSD01

Tot Ion:168 Resp: 292782

Ion Ratio Lower Upper

168 100

99 48.5 34.9 52.3



Abundance Ion 168.00 (167.70 to 168.70): VF0

Ion 99.00 (98.70 to 99.70): VF0

4.85

100000

80000



60000

40000

20000

0

Time-->



#2

Dichlorodifluoromethane

Concen: 24.935 ug/l

RT: 0.96 min Scan# 33

Delta R.T. -0.01 min

Lab File: VF061790.D


Acq: 28 Feb 2019 16:28

Tgt Ion: 85 Resp: 83840

Ion Ratio Lower Upper

85 100

87 29.3 16.1 48.2



Abundance Ion 85.00 (84.70 to 85.70): VF0

Ion 87.00 (86.70 to 87.70): VF0

0.96

30000

25000

20000


15000


10000

5000

0

Time-->





#3

Chloromethane

Concen: 21.970 ug/l

RT: 1.10 min Scan# 47

Delta R.T. -0.01 min

Lab File: VF061790.D

Acq: 28 Feb 2019 16:28

Instrument :

MSVOA_F

ClientSampleId :





VF0228SBSD01

Tot Ion: 50 Resp: 77504

Ion Ratio Lower Upper

50 100

52 32.2 26.2 39.4



#4

Vinyl Chloride

Concen: 23.833 ug/l

RT: 1.14 min Scan# 51

Delta R.T. -0.01 min

Lab File: VF061790.D




Acq: 28 Feb 2019 16:28


Tgt Ion: 62 Resp: 78160

Ion Ratio Lower Upper

62 100

64 31.9 27.5 41.3





#5

Bromomethane

Concen: 22.886 ug/l

RT: 1.32 min Scan# 70

Delta R.T. 0.00 min

Lab File: VF061790.D

Acq: 28 Feb 2019 16:28

Instrument :

MSVOA_F

ClientSampleId :


VF0228SBSD01

Tot Ion: 94 Resp: 51657

Ion Ratio Lower Upper




94 100

96 94.4 74.6 112.0



Abundance Ion 94.00 (93.70 to 94.70): VF0

Ion 96.00 (95.70 to 96.70): VF0



#6

Chloroethane

Concen: 23.045 ug/l

RT: 1.38 min Scan# 76

Delta R.T. -0.02 min

Lab File: VF061790.D


Acq: 28 Feb 2019 16:28

Tgt Ion: 64 Resp: 37245

Ion Ratio Lower Upper



64 100


66 34.0 27.9 41.9



Abundance Ion 64.00 (63.70 to 64.70): VF0

Ion 66.00 (65.70 to 66.70): VF0





#7

Trichlorofluoromethane

Concen: 23.588 ug/l

RT: 1.48 min Scan# 86

Delta R.T. -0.01 min

Lab File: VF061790.D

Acq: 28 Feb 2019 16:28

Instrument :

MSVOA_F

ClientSampleId :


VF0228SBSD01

Tot Ion:101 Resp: 93304

Ion Ratio Lower Upper




101 100

103 61.9 49.3 73.9



Abundance Ion 101.00 (100.70 to 101.70): V

Ion 103.00 (102.70 to 103.70): V



#8

Diethyl Ether

Concen: 20.117 ug/l

RT: 1.63 min Scan# 101

Delta R.T. -0.01 min

Lab File: VF061790.D


Acq: 28 Feb 2019 16:28

Tgt Ion: 74 Resp: 20053

Ion Ratio Lower Upper



74 100


45 109.2 60.8 182.3



Abundance Ion 74.00 (73.70 to 74.70): VF0

Ion 45.00 (44.70 to 45.70): VF0





#9

1,1,2-Trichlorotrifluoroethane

Concen: 22.928 ug/l

RT: 1.83 min Scan# 121

Delta R.T. -0.02 min

Lab File: VF061790.D

Acq: 28 Feb 2019 16:28

Instrument :

MSVOA_F

ClientSampleId :

VF0228SBSD01


Tot Ion:101 Resp: 65427

Ion Ratio Lower Upper

101 100

85 41.4 29.8 44.6




151 66.8 56.3 84.5



Abundance Ion 101.00 (100.70 to 101.70): V

Ion 85.00 (84.70 to 85.70): VF0

Ion 151.00 (150.70 to 151.70): V



#10

Methyl Iodide

Concen: 16.927 ug/l

RT: 1.90 min Scan# 129

Delta R.T. -0.01 min

Lab File: VF061790.D

Acq: 28 Feb 2019 16:28


Tgt Ion:142 Resp: 92268

Ion Ratio Lower Upper

142 100

127 48.5 36.6 55.0



141 14.8 11.0 16.6




Abundance Ion 142.00 (141.70 to 142.70): V

Ion 127.00 (126.70 to 127.70): V

Ion 141.00 (140.70 to 141.70): V









#11

Tert butyl alcohol
 Concen: 96.623 ug/l
 RT: 2.57 min Scan# 196
 Delta R.T. -0.02 min
 Lab File: VF061790.D
 Acq: 28 Feb 2019 16:28

Instrument :
 MSVOA_F
 ClientSampleId :
 VF0228SBSD01




Tot Ion: 59 Resp: 12006
 Ion Ratio Lower Upper
 59 100
 57 17.0 12.5 18.7




#12

1,1-Dichloroethene
 Concen: 22.957 ug/l
 RT: 1.79 min Scan# 117
 Delta R.T. -0.02 min
 Lab File: VF061790.D
 Acq: 28 Feb 2019 16:28

Tgt Ion: 96 Resp: 57596
 Ion Ratio Lower Upper
 96 100
 61 148.8 122.6 184.0
 98 62.6 47.2 70.8





#13

Acrolein

Concen: 90.275 ug/l

RT: 1.99 min Scan# 138

Delta R.T. -0.02 min

Lab File: VF061790.D

Acq: 28 Feb 2019 16:28

Instrument :

MSVOA_F

ClientSampleId :





VF0228SBSD01

Tot Ion: 56 Resp: 14098

Ion Ratio Lower Upper

56 100

55 68.7 55.0 82.6



#14

Allyl chloride

Concen: 21.867 ug/l

RT: 2.09 min Scan# 148

Delta R.T. -0.01 min

Lab File: VF061790.D

Acq: 28 Feb 2019 16:28




Tgt Ion: 41 Resp: 67319


Ion Ratio Lower Upper

41 100

39 79.5 61.6 92.4

76 36.8 30.3 45.5









#15
 Acrylonitrile
 Concen: 98.882 ug/l
 RT: 2.93 min Scan# 233
 Delta R.T. -0.03 min
 Lab File: VF061790.D
 Acq: 28 Feb 2019 16:28




Instrument :
 MSVOA_F
 ClientSampleId :
 VF0228SBS01


Tot Ion	53	Resp	43581
Ion	Ratio	Lower	Upper
53	100		
52	88.6	64.6	97.0
51	40.4	29.8	44.8



#16
 Acetone
 Concen: 117.006 ug/l
 RT: 2.23 min Scan# 162
 Delta R.T. -0.02 min
 Lab File: VF061790.D
 Acq: 28 Feb 2019 16:28

Tgt Ion	43	Resp	53899
Ion	Ratio	Lower	Upper
43	100		
58	25.3	22.9	34.3









#17
Carbon Disulfide
Concen: 15.388 ug/l
RT: 1.83 min Scan# 121
Delta R.T. -0.01 min
Lab File: VF061790.D
Acq: 28 Feb 2019 16:28




Instrument :
MSVOA_F
ClientSampleId :
VF0228SBSD01


Tot Ion: 76 Resp: 121355
Ion Ratio Lower Upper
76 100
78 9.9 7.9 11.9



#18
Methyl Acetate
Concen: 23.157 ug/l
RT: 2.34 min Scan# 173
Delta R.T. -0.02 min
Lab File: VF061790.D
Acq: 28 Feb 2019 16:28

Tgt Ion: 43 Resp: 24031
Ion Ratio Lower Upper
43 100
74 19.4 14.6 21.8









#19

Methyl tert-butyl Ether
 Concen: 22.600 ug/l
 RT: 2.42 min Scan# 181
 Delta R.T. -0.02 min
 Lab File: VF061790.D
 Acq: 28 Feb 2019 16:28

Instrument :
 MSVOA_F
 ClientSampleId :
 VF0228SBSD01




Tot Ion: 73 Resp: 96903
 Ion Ratio Lower Upper
 73 100
 57 22.6 15.8 23.8




#20

Methylene Chloride
 Concen: 6.285 ug/l
 RT: 2.17 min Scan# 156
 Delta R.T. -0.02 min
 Lab File: VF061790.D
 Acq: 28 Feb 2019 16:28

Tgt Ion: 84 Resp: 20954
 Ion Ratio Lower Upper
 84 100
 49 121.7 101.8 152.8
 51 44.4 31.1 46.7
 86 65.4 54.4 81.6





#21

trans-1,2-Dichloroethene

Concen: 20.750 ug/l

RT: 2.30 min Scan# 169

Delta R.T. -0.02 min

Lab File: VF061790.D

Acq: 28 Feb 2019 16:28

Instrument :

MSVOA_F

ClientSampleId :

VF0228SBSD01


Tot Ion: 96 Resp: 56593

Ion Ratio Lower Upper

96 100

61 154.0 108.2 162.4




98 68.6 52.2 78.2



Abundance Ion 96.00 (95.70 to 96.70): VF0

Ion 61.00 (60.70 to 61.70): VF0

Ion 98.00 (97.70 to 98.70): VF0



#22

Diisopropyl ether

Concen: 22.737 ug/l

RT: 2.79 min Scan# 219

Delta R.T. -0.02 min

Lab File: VF061790.D

Acq: 28 Feb 2019 16:28

Tgt Ion: 45 Resp: 168029


Ion Ratio Lower Upper

45 100

43 39.7 31.4 47.2

87 19.0 16.3 24.5

59 10.5 8.7 13.1






Abundance Ion 45.00 (44.70 to 45.70): VF0

Ion 43.00 (42.70 to 43.70): VF0

Ion 87.00 (86.70 to 87.70): VF0

Ion 59.00 (58.70 to 59.70): VF0





#23

Vinyl Acetate

Concen: 103.526 ug/l

RT: 3.18 min Scan# 258

Delta R.T. -0.03 min

Lab File: VF061790.D

Acq: 28 Feb 2019 16:28

Instrument :

MSVOA_F

ClientSampleId :


VF0228SBSD01

Tot Ion: 43 Resp: 373678

Ion Ratio Lower Upper




43 100

86 8.2 7.5 11.3



Abundance Ion 43.00 (42.70 to 43.70): VF0

Ion 86.00 (85.70 to 86.70): VF0



#24

1,1-Dichloroethane

Concen: 22.388 ug/l

RT: 2.86 min Scan# 226

Delta R.T. -0.02 min

Lab File: VF061790.D

Acq: 28 Feb 2019 16:28


Tgt Ion: 63 Resp: 100121

Ion Ratio Lower Upper

63 100

98 5.2 2.8 8.3



100 3.8 1.8 5.4




Abundance Ion 63.00 (62.70 to 63.70): VF0

Ion 98.00 (97.70 to 98.70): VF0

Ion 100.00 (99.70 to 100.70): VF0





#25

2-Butanone

Concen: 105.933 ug/l

RT: 4.32 min Scan# 374

Delta R.T. -0.02 min

Lab File: VF061790.D

Acq: 28 Feb 2019 16:28

Instrument :

MSVOA_F

ClientSampleId :


VF0228SBSD01

Tot Ion: 43 Resp: 106591

Ion Ratio Lower Upper

43 100

72 22.7 17.1 25.7



Abundance Ion 43.00 (42.70 to 43.70): VF0

Ion 72.00 (71.70 to 72.70): VF0

4.32

25000

20000

15000



10000

5000

0

4.20 4.30 4.40 4.50

Time-->



#26

2,2-Dichloropropane

Concen: 25.671 ug/l

RT: 3.58 min Scan# 299

Delta R.T. -0.03 min

Lab File: VF061790.D


Acq: 28 Feb 2019 16:28

Tgt Ion: 77 Resp: 48629

Ion Ratio Lower Upper

77 100

97 21.9 11.7 35.0



Abundance Ion 77.00 (76.70 to 77.70): VF0

Ion 97.00 (96.70 to 97.70): VF0

15000


10000


5000

0

3.40 3.50 3.60 3.70

Time-->









#27

cis-1,2-Dichloroethene
 Concen: 21.787 ug/l
 RT: 3.46 min Scan# 287
 Delta R.T. -0.03 min
 Lab File: VF061790.D
 Acq: 28 Feb 2019 16:28

Instrument :
 MSVOA_F
 ClientSampleId :
 VF0228SBSD01




Tot Ion: 96 Resp: 77993
 Ion Ratio Lower Upper
 96 100
 61 108.8 0.0 241.6
 98 57.5 0.0 130.2




#28

Bromochloromethane
 Concen: 21.309 ug/l
 RT: 3.70 min Scan# 311
 Delta R.T. -0.03 min
 Lab File: VF061790.D
 Acq: 28 Feb 2019 16:28

Tgt Ion: 49 Resp: 45724
 Ion Ratio Lower Upper
 49 100
 129 0.0 0.0 3.4
 130 85.4 77.7 116.5









#29
Tetrahydrofuran
Concen: 93.735 ug/l
RT: 4.05 min Scan# 347
Delta R.T. -0.03 min
Lab File: VF061790.D
Acq: 28 Feb 2019 16:28

Instrument :
MSVOA_F
ClientSampleId :
VF0228SBSD01

Tot Ion: 42 Resp: 42868
Ion Ratio Lower Upper
42 100
72 43.7 35.2 52.8
71 40.9 32.9 49.3




Abundance Ion 42.00 (41.70 to 42.70): VF0
Ion 72.00 (71.70 to 72.70): VF0
Ion 71.00 (70.70 to 71.70): VF0






#30
Chloroform
Concen: 22.441 ug/l
RT: 3.85 min Scan# 326
Delta R.T. -0.03 min
Lab File: VF061790.D
Acq: 28 Feb 2019 16:28

Tgt Ion: 83 Resp: 114133
Ion Ratio Lower Upper
83 100
85 66.9 51.1 76.7



Abundance Ion 83.00 (82.70 to 83.70): VF0
Ion 85.00 (84.70 to 85.70): VF0





#31

Cyclohexane

Concen: 14.543 ug/l

RT: 3.68 min Scan# 309

Delta R.T. -0.03 min

Lab File: VF061790.D

Acq: 28 Feb 2019 16:28

Instrument :

MSVOA_F

ClientSampleId :

VF0228SBS01





Tot Ion: 56 Resp: 70410

Ion Ratio Lower Upper

56 100

69 32.2 23.8 35.8

84 91.6 68.6 103.0



#32

1,1,1-Trichloroethane

Concen: 23.057 ug/l

RT: 4.08 min Scan# 350

Delta R.T. -0.02 min

Lab File: VF061790.D

Acq: 28 Feb 2019 16:28




Tgt Ion: 97 Resp: 76815


Ion Ratio Lower Upper

97 100

99 59.9 52.2 78.2

61 47.5 39.2 58.8





#33

1,2-Dichloroethane-d4

Concen: 48.888 ug/l

RT: 4.82 min Scan# 425

Delta R.T. -0.03 min

Lab File: VF061790.D

Acq: 28 Feb 2019 16:28

Instrument :

MSVOA_F

ClientSampleId :


VF0228SBSD01

Tot Ion: 65 Resp: 104963

Ion Ratio Lower Upper




65 100

67 51.1 0.0 104.0



Abundance Ion 65.00 (64.70 to 65.70): VF0

Ion 67.00 (66.70 to 67.70): VF0



#34

1,4-Difluorobenzene

Concen: 50.000 ug/l

RT: 5.58 min Scan# 502

Delta R.T. -0.03 min

Lab File: VF061790.D

Acq: 28 Feb 2019 16:28


Tgt Ion: 114 Resp: 428247

Ion Ratio Lower Upper

114 100

63 16.9 0.0 36.6



88 17.1 0.0 34.4




Abundance Ion 114.00 (113.70 to 114.70): V

Ion 63.00 (62.70 to 63.70): VF0

Ion 88.00 (87.70 to 88.70): VF0





#35

Dibromofluoromethane

Concen: 47.243 ug/l

RT: 4.09 min Scan# 351

Delta R.T. -0.03 min

Lab File: VF061790.D

Acq: 28 Feb 2019 16:28

Instrument :

MSVOA_F

ClientSampleId :

VF0228SBSD01


Tot Ion:113 Resp: 150243

Ion Ratio Lower Upper

113 100

111 103.2 78.8 118.2




192 18.1 13.4 20.0



Abundance Ion 113.00 (112.70 to 113.70): V

Ion 111.00 (110.70 to 111.70): V

Ion 192.00 (191.70 to 192.70): V



#36

1,1-Dichloropropene

Concen: 21.788 ug/l

RT: 4.27 min Scan# 369

Delta R.T. -0.03 min

Lab File: VF061790.D

Acq: 28 Feb 2019 16:28


Tgt Ion: 75 Resp: 92433

Ion Ratio Lower Upper

75 100

110 34.2 20.5 61.6



77 29.4 27.0 40.4




Abundance Ion 75.00 (74.70 to 75.70): VF0

Ion 110.00 (109.70 to 110.70): V

Ion 77.00 (76.70 to 77.70): VF0









#37
Ethyl Acetate
Concen: 19.342 ug/l
RT: 4.10 min Scan# 352
Delta R.T. -0.03 min
Lab File: VF061790.D
Acq: 28 Feb 2019 16:28




Instrument :
MSVOA_F
ClientSampleId :
VF0228SBSD01


Tot Ion: 43 Resp: 38112
Ion Ratio Lower Upper
43 100
61 57.2 45.7 68.5
70 12.8 8.5 12.7#



#38
Carbon Tetrachloride
Concen: 23.256 ug/l
RT: 3.98 min Scan# 339
Delta R.T. -0.03 min
Lab File: VF061790.D
Acq: 28 Feb 2019 16:28

Tgt Ion: 117 Resp: 68124
Ion Ratio Lower Upper
117 100
119 93.6 75.6 113.4
121 30.0 25.2 37.8









#39
Methvlcvclohexane
Concen: 23.031 ug/l
RT: 5.43 min Scan# 487
Delta R.T. -0.03 min
Lab File: VF061790.D
Acq: 28 Feb 2019 16:28




Instrument :
MSVOA_F
ClientSampleId :
VF0228SBS01


Tot Ion: 83 Resp: 115553
Ion Ratio Lower Upper
83 100
55 83.8 66.6 99.8
98 47.8 40.4 60.6



#40
Benzene
Concen: 22.357 ug/l
RT: 4.62 min Scan# 404
Delta R.T. -0.03 min
Lab File: VF061790.D
Acq: 28 Feb 2019 16:28

Tgt Ion: 78 Resp: 248083
Ion Ratio Lower Upper
78 100
77 22.2 19.0 28.6









#41
Methacrylonitrile
Concen: 23.175 ug/l
RT: 4.72 min Scan# 415
Delta R.T. -0.03 min
Lab File: VF061790.D
Acq: 28 Feb 2019 16:28




Instrument :
MSVOA_F
ClientSampleId :
VF0228SBSD01


Tot Ion:	41	Resp:	23353
Ion	Ratio	Lower	Upper
41	100		
39	59.9	45.3	67.9
67	58.4	49.3	73.9
52	45.8	32.6	49.0



#42
1,2-Dichloroethane
Concen: 22.882 ug/l
RT: 4.93 min Scan# 436
Delta R.T. -0.03 min
Lab File: VF061790.D
Acq: 28 Feb 2019 16:28

Tgt Ion:	62	Resp:	60735
Ion	Ratio	Lower	Upper
62	100		
98	10.4	0.0	26.0





#43

Isopropyl Acetate

Concen: 19.501 ug/l

RT: 6.94 min Scan# 640

Delta R.T. -0.03 min

Lab File: VF061790.D

Acq: 28 Feb 2019 16:28

Instrument :

MSVOA_F

ClientSampleId :

VF0228SBS01


Tot Ion: 43 Resp: 45485

Ion Ratio Lower Upper

43 100

61 32.7 25.8 38.6




87 0.0 0.4 0.6#



Abundance Ion 43.00 (42.70 to 43.70): VF0

Ion 61.00 (60.70 to 61.70): VF0

Ion 87.00 (86.70 to 87.70): VF0



#44

Trichloroethene

Concen: 22.998 ug/l

RT: 5.48 min Scan# 492

Delta R.T. -0.03 min

Lab File: VF061790.D


Acq: 28 Feb 2019 16:28

Tgt Ion:130 Resp: 82523

Ion Ratio Lower Upper



130 100


95 86.4 0.0 179.4



Abundance Ion 129.90 (129.60 to 130.60): V

Ion 94.90 (94.60 to 95.60): VF0





#45

1,2-Dichloropropane

Concen: 21.501 ug/l

RT: 6.22 min Scan# 567

Delta R.T. -0.03 min

Lab File: VF061790.D

Acq: 28 Feb 2019 16:28

Instrument :

MSVOA_F

ClientSampleId :





VF0228SBSD01

Tot Ion: 63 Resp: 57669

Ion Ratio Lower Upper

63 100

65 33.2 24.9 37.3



#46

Dibromomethane

Concen: 22.804 ug/l

RT: 6.08 min Scan# 552

Delta R.T. -0.03 min

Lab File: VF061790.D

Acq: 28 Feb 2019 16:28




Tgt Ion: 93 Resp: 37660


Ion Ratio Lower Upper

93 100

95 71.9 69.0 103.6

174 108.8 99.3 148.9





#47

Bromodichloromethane

Concen: 23.168 ug/l

RT: 6.37 min Scan# 582

Delta R.T. -0.03 min

Lab File: VF061790.D

Acq: 28 Feb 2019 16:28

Instrument :

MSVOA_F

ClientSampleId :

VF0228SBS01


Tot Ion: 83 Resp: 83785

Ion Ratio Lower Upper

83 100

85 68.7 53.4 80.2




127 8.9 7.5 11.3



Abundance Ion 83.00 (82.70 to 83.70): VF0

Ion 85.00 (84.70 to 85.70): VF0

Ion 127.00 (126.70 to 127.70): V



#48

Methyl methacrylate

Concen: 18.486 ug/l

RT: 6.70 min Scan# 615

Delta R.T. -0.03 min

Lab File: VF061790.D

Acq: 28 Feb 2019 16:28


Tgt Ion: 41 Resp: 25947

Ion Ratio Lower Upper

41 100

69 87.7 60.4 90.6



39 58.4 43.7 65.5




Abundance Ion 41.00 (40.70 to 41.70): VF0

Ion 69.00 (68.70 to 69.70): VF0

Ion 39.00 (38.70 to 39.70): VF0





#49

1,4-Dioxane

Concen: 409.939 ug/l

RT: 6.70 min Scan# 615

Delta R.T. -0.02 min

Lab File: VF061790.D

Acq: 28 Feb 2019 16:28

Instrument :

MSVOA_F

ClientSampleId :

VF0228SBSD01


Tot Ion: 88 Resp: 5409

Ion Ratio Lower Upper

88 100

43 40.2 32.3 48.5




58 56.7 59.4 89.2#



Abundance Ion 88.00 (87.70 to 88.70): VF0

Ion 43.00 (42.70 to 43.70): VF0

Ion 58.00 (57.70 to 58.70): VF0



#50

Toluene-d8

Concen: 48.749 ug/l

RT: 7.54 min Scan# 700

Delta R.T. -0.03 min

Lab File: VF061790.D


Acq: 28 Feb 2019 16:28

Tgt Ion: 98 Resp: 445670

Ion Ratio Lower Upper



98 100


100 70.6 55.4 83.2



Abundance Ion 98.00 (97.70 to 98.70): VF0

Ion 100.00 (99.70 to 100.70): VF0





#51

4-Methyl-2-Pentanone

Concen: 100.572 ug/l

RT: 8.25 min Scan# 772

Delta R.T. -0.03 min

Lab File: VF061790.D

Acq: 28 Feb 2019 16:28

Instrument :

MSVOA_F

ClientSampleId :


VF0228SBSD01

Tot Ion: 43 Resp: 167611

Ion Ratio Lower Upper

43 100

58 43.9 33.1 49.7



Abundance Ion 43.00 (42.70 to 43.70): VF0

Ion 58.00 (57.70 to 58.70): VF0

8.25

60000



40000

20000

0

8.10 8.20 8.30 8.40

Time-->



#52

Toluene

Concen: 22.318 ug/l

RT: 7.61 min Scan# 708

Delta R.T. -0.02 min

Lab File: VF061790.D


Acq: 28 Feb 2019 16:28

Tgt Ion: 92 Resp: 150193

Ion Ratio Lower Upper

92 100

91 162.0 133.9 200.9



Abundance Ion 92.00 (91.70 to 92.70): VF0

Ion 91.00 (90.70 to 91.70): VF0

7.61

100000

80000

60000


40000


20000

0

7.50 7.60 7.70 7.80

Time-->









#53
 t-1,3-Dichloropropene
 Concen: 22.130 ug/l
 RT: 8.26 min Scan# 774
 Delta R.T. -0.03 min
 Lab File: VF061790.D
 Acq: 28 Feb 2019 16:28




Instrument :
 MSVOA_F
 ClientSampleId :
 VF0228SBSD01


Tot Ion: 75 Resp: 66610
 Ion Ratio Lower Upper
 75 100
 77 29.1 26.2 39.2



#54
 cis-1,3-Dichloropropene
 Concen: 22.540 ug/l
 RT: 7.28 min Scan# 674
 Delta R.T. -0.03 min
 Lab File: VF061790.D
 Acq: 28 Feb 2019 16:28

Tgt Ion: 75 Resp: 90366
 Ion Ratio Lower Upper
 75 100
 77 29.8 26.4 39.6
 39 51.6 36.0 54.0





#55

1,1,2-Trichloroethane

Concen: 21.023 ug/l

RT: 8.47 min Scan# 795

Delta R.T. -0.03 min

Lab File: VF061790.D

Acq: 28 Feb 2019 16:28

Instrument :

MSVOA_F

ClientSampleId :

VF0228SBSD01

Tot Ion: 97 Resp: 41568


Ion Ratio Lower Upper

97 100

83 81.5 69.2 103.8

85 61.5 43.0 64.4

99 57.9 47.2 70.8






Abundance Ion 97.00 (96.70 to 97.70): VF0

Ion 82.95 (82.65 to 83.65): VF0

Ion 84.95 (84.65 to 85.65): VF0

Ion 98.95 (98.65 to 99.65): VF0



#56

Ethyl methacrylate

Concen: 21.255 ug/l

RT: 8.60 min Scan# 808

Delta R.T. -0.03 min

Lab File: VF061790.D

Acq: 28 Feb 2019 16:28


Tgt Ion: 69 Resp: 46792

Ion Ratio Lower Upper

69 100

41 62.5 52.9 79.3



39 32.7 24.5 36.7




Abundance Ion 69.00 (68.70 to 69.70): VF0

Ion 41.00 (40.70 to 41.70): VF0

Ion 39.00 (38.70 to 39.70): VF0





#57

1,3-Dichloropropane

Concen: 21.875 ug/l

RT: 8.84 min Scan# 832

Delta R.T. -0.03 min

Lab File: VF061790.D

Acq: 28 Feb 2019 16:28

Instrument :

MSVOA_F

ClientSampleId :





VF0228SBSD01

Tot Ion: 76 Resp: 72196

Ion Ratio Lower Upper

76 100

78 30.3 27.8 41.6



#58

2-Chloroethyl Vinyl ether

Concen: 102.573 ug/l

RT: 7.28 min Scan# 674

Delta R.T. -0.03 min

Lab File: VF061790.D




Acq: 28 Feb 2019 16:28


Tgt Ion: 63 Resp: 38960

Ion Ratio Lower Upper

63 100

106 31.4 24.6 36.8









#59
2-Hexanone
Concen: 108.757 ug/l
RT: 9.48 min Scan# 897
Delta R.T. -0.03 min
Lab File: VF061790.D
Acq: 28 Feb 2019 16:28




Instrument :
MSVOA_F
ClientSampleId :
VF0228SBSD01


Tot Ion: 43 Resp: 127052
Ion Ratio Lower Upper
43 100
58 57.6 28.5 85.5



#60
Dibromochloromethane
Concen: 21.514 ug/l
RT: 8.70 min Scan# 818
Delta R.T. -0.03 min
Lab File: VF061790.D
Acq: 28 Feb 2019 16:28

Tgt Ion:129 Resp: 59210
Ion Ratio Lower Upper
129 100
127 76.2 37.8 113.4





#61

1,2-Dibromoethane

Concen: 22.205 ug/l

RT: 8.97 min Scan# 846

Delta R.T. -0.03 min

Lab File: VF061790.D

Acq: 28 Feb 2019 16:28

Instrument :

MSVOA_F

ClientSampleId :


VF0228SBSD01

Tot Ion:107 Resp: 47621

Ion Ratio Lower Upper

107 100

109 88.1 77.4 116.0



Abundance Ion 107.00 (106.70 to 107.70): V

Ion 109.00 (108.70 to 109.70): V

8.97

20000

15000



10000

5000

0

Time-->

8.90 9.00 9.10



#62

4-Bromofluorobenzene

Concen: 45.688 ug/l

RT: 11.40 min Scan# 1092

Delta R.T. -0.02 min

Lab File: VF061790.D

Acq: 28 Feb 2019 16:28


Tgt Ion: 95 Resp: 169002

Ion Ratio Lower Upper

95 100

174 88.0 0.0 165.8

176 87.8 0.0 167.0



Abundance Ion 95.00 (94.70 to 95.70): VF0

Ion 174.00 (173.70 to 174.70): V

Ion 176.00 (175.70 to 176.70): V

11.40


100000


50000

0

Time-->

11.30 11.40 11.50





#63

Chlorobenzene-d5

Concen: 50.000 ug/l

RT: 9.76 min Scan# 926

Delta R.T. -0.03 min

Lab File: VF061790.D

Acq: 28 Feb 2019 16:28

Instrument :

MSVOA_F

ClientSampleId :

VF0228SBSD01


Tot Ion:117 Resp: 368919

Ion Ratio Lower Upper

117 100

82 47.7 39.3 58.9




119 33.2 26.6 40.0



Abundance Ion 117.00 (116.70 to 117.70): V

Ion 82.00 (81.70 to 82.70): VF0

Ion 119.00 (118.70 to 119.70): V



#64

Tetrachloroethene

Concen: 23.108 ug/l

RT: 8.14 min Scan# 761

Delta R.T. -0.03 min

Lab File: VF061790.D

Acq: 28 Feb 2019 16:28

Tgt Ion:164 Resp: 69682


Ion Ratio Lower Upper

164 100

166 131.0 103.9 155.9

129 89.0 71.3 106.9

131 89.8 67.1 100.7






Abundance Ion 163.85 (163.55 to 164.55): V

Ion 165.80 (165.50 to 166.50): V

Ion 128.90 (128.60 to 129.60): V

Ion 130.90 (130.60 to 131.60): V





#65

Chlorobenzene

Concen: 22.283 ug/l

RT: 9.78 min Scan# 928

Delta R.T. -0.03 min

Lab File: VF061790.D

Acq: 28 Feb 2019 16:28

Instrument :

MSVOA_F

ClientSampleId :


VF0228SBSD01

Tot Ion:112 Resp: 166818

Ion Ratio Lower Upper


112 100

114 31.1 26.2 39.2





Abundance Ion 112.00 (111.70 to 112.70): V

Ion 114.00 (113.70 to 114.70): V



Time-->



#66

1,1,1,2-Tetrachloroethane

Concen: 23.406 ug/l

RT: 9.91 min Scan# 941

Delta R.T. -0.03 min

Lab File: VF061790.D

Acq: 28 Feb 2019 16:28


Tgt Ion:131 Resp: 68662

Ion Ratio Lower Upper

131 100

133 99.0 48.9 146.6


119 63.1 34.8 104.3




Abundance Ion 131.00 (130.70 to 131.70): V


Ion 133.00 (132.70 to 133.70): V

Ion 119.00 (118.70 to 119.70): V



Time-->





#67

Ethyl Benzene

Concen: 23.699 ug/l

RT: 9.88 min Scan# 938

Delta R.T. -0.03 min

Lab File: VF061790.D

Acq: 28 Feb 2019 16:28

Instrument :

MSVOA_F

ClientSampleId :


VF0228SBSD01

Tot Ion: 91 Resp: 295039




Ion Ratio Lower Upper

91 100

106 32.2 24.1 36.1



Abundance Ion 91.00 (90.70 to 91.70): VF0



#68

m/p-Xylenes

Concen: 47.166 ug/l

RT: 10.12 min Scan# 962

Delta R.T. -0.03 min

Lab File: VF061790.D


Acq: 28 Feb 2019 16:28

Tgt Ion: 106 Resp: 224726



Ion Ratio Lower Upper


106 100

91 191.5 153.0 229.6



Abundance Ion 106.00 (105.70 to 106.70): V









#69
o-Xylene
Concen: 22.953 ug/l
RT: 10.70 min Scan# 1021
Delta R.T. -0.03 min
Lab File: VF061790.D
Acq: 28 Feb 2019 16:28




Instrument :
MSVOA_F
ClientSampleId :
VF0228SBSD01


Tot Ion:106 Resp: 116697
Ion Ratio Lower Upper
106 100
91 203.9 100.1 300.1



#70
Styrene
Concen: 23.041 ug/l
RT: 10.78 min Scan# 1029
Delta R.T. -0.02 min
Lab File: VF061790.D
Acq: 28 Feb 2019 16:28

Tgt Ion:104 Resp: 165720
Ion Ratio Lower Upper
104 100
78 39.9 33.2 49.8
103 45.7 37.2 55.8





#71

Bromoform

Concen: 21.564 ug/l

RT: 10.76 min Scan# 1027

Delta R.T. -0.03 min

Lab File: VF061790.D

Acq: 28 Feb 2019 16:28

Instrument :

MSVOA_F

ClientSampleId :

VF0228SBSD01


Tot Ion:173 Resp: 34350

Ion Ratio Lower Upper

173 100

175 49.6 24.0 72.0




254 9.9 8.2 12.2



Abundance Ion 173.00 (172.70 to 173.70): V

Ion 175.00 (174.70 to 175.70): V

Ion 254.00 (253.70 to 254.70): V



#72

1,4-Dichlorobenzene-d4

Concen: 50.000 ug/l

RT: 12.54 min Scan# 1208

Delta R.T. -0.02 min

Lab File: VF061790.D

Acq: 28 Feb 2019 16:28


Tgt Ion:152 Resp: 210655

Ion Ratio Lower Upper

152 100

115 48.3 24.3 72.8



150 151.6 0.0 312.0




Abundance Ion 152.00 (151.70 to 152.70): V

Ion 115.00 (114.70 to 115.70): V

Ion 150.00 (149.70 to 150.70): V





#73

Isopropylbenzene

Concen: 21.834 ug/l

RT: 11.11 min Scan# 1063

Delta R.T. -0.02 min

Lab File: VF061790.D

Acq: 28 Feb 2019 16:28

Instrument :

MSVOA_F

ClientSampleId :


VF0228SBSD01

Tot Ion:105 Resp: 327475

Ion Ratio Lower Upper

105 100

120 26.4 13.4 40.1



Abundance Ion 105.00 (104.70 to 105.70): V

Ion 120.00 (119.70 to 120.70): V

200000

150000



100000

50000

0

11.00 11.10 11.20 11.30

Time-->



#74

N-ethyl acetate

Concen: 21.028 ug/l

RT: 11.36 min Scan# 1088

Delta R.T. -0.02 min

Lab File: VF061790.D

Acq: 28 Feb 2019 16:28

Tgt Ion: 43 Resp: 75700


Ion Ratio Lower Upper

43 100

70 38.7 31.5 47.3

55 22.1 16.7 25.1

61 27.8 23.0 34.6



Abundance Ion 43.00 (42.70 to 43.70): VF0

Ion 70.00 (69.70 to 70.70): VF0

Ion 55.00 (54.70 to 55.70): VF0

Ion 61.00 (60.70 to 61.70): VF0

60000


40000


20000

0

11.30 11.40 11.50

Time-->





#75

1,1,2,2-Tetrachloroethane

Concen: 21.363 ug/l

RT: 11.69 min Scan# 1121

Delta R.T. -0.02 min

Lab File: VF061790.D

Acq: 28 Feb 2019 16:28

Instrument :

MSVOA_F

ClientSampleId :

VF0228SBSD01





Tot Ion: 83 Resp: 57401

Ion Ratio Lower Upper

83 100

131 11.2 4.4 13.2

85 63.9 31.5 94.5



#76

1,2,3-Trichloropropane

Concen: 20.468 ug/l

RT: 11.79 min Scan# 1131

Delta R.T. -0.02 min

Lab File: VF061790.D




Acq: 28 Feb 2019 16:28


Tgt Ion: 75 Resp: 37392

Ion Ratio Lower Upper

75 100

77 34.5 16.6 49.7





#77

Bromobenzene

Concen: 22.392 ug/l

RT: 11.48 min Scan# 1100

Delta R.T. -0.03 min

Lab File: VF061790.D

Acq: 28 Feb 2019 16:28

Instrument :

MSVOA_F

ClientSampleId :

VF0228SBSD01


Tot Ion:156 Resp: 82967

Ion Ratio Lower Upper

156 100

77 144.5 64.3 192.7




158 94.4 50.9 152.7



Abundance Ion 156.00 (155.70 to 156.70): V

Ion 77.00 (76.70 to 77.70): VF0

Ion 158.00 (157.70 to 158.70): V



#78

n-propylbenzene

Concen: 23.312 ug/l

RT: 11.58 min Scan# 1110

Delta R.T. -0.03 min

Lab File: VF061790.D


Acq: 28 Feb 2019 16:28

Tgt Ion: 91 Resp: 401216

Ion Ratio Lower Upper



91 100


120 21.8 12.1 36.3



Abundance Ion 91.00 (90.70 to 91.70): VF0

Ion 120.00 (119.70 to 120.70): V









#79
 2-Chlorotoluene
 Concen: 23.087 ug/l
 RT: 11.71 min Scan# 1123
 Delta R.T. -0.03 min
 Lab File: VF061790.D
 Acq: 28 Feb 2019 16:28




Instrument :
 MSVOA_F
 ClientSampleId :
 VF0228SBSD01


Tot Ion: 91 Resp: 222531
 Ion Ratio Lower Upper
 91 100
 126 35.5 19.8 59.4



#80
 1,3,5-Trimethylbenzene
 Concen: 22.628 ug/l
 RT: 11.81 min Scan# 1134
 Delta R.T. -0.02 min
 Lab File: VF061790.D
 Acq: 28 Feb 2019 16:28

Tgt Ion: 105 Resp: 271244
 Ion Ratio Lower Upper
 105 100
 120 55.6 27.9 83.6





#81

trans-1,4-Dichloro-2-butene

Concen: 25.808 ug/l

RT: 11.85 min Scan# 1138

Delta R.T. -0.03 min

Lab File: VF061790.D

Acq: 28 Feb 2019 16:28

Instrument :


MSVOA_F

ClientSampleId :

VF0228SBS01

Tot Ion: 75 Resp: 22603

Ion	Ratio	Lower	Upper
75	100		
53	74.0	53.7	80.5
89	46.3	41.6	62.4






Abundance

Ion 75.00 (74.70 to 75.70): VF0

Ion 53.00 (52.70 to 53.70): VF0

Ion 89.00 (88.70 to 89.70): VF0



#82

4-Chlorotoluene

Concen: 22.213 ug/l

RT: 11.88 min Scan# 1141


Delta R.T. -0.03 min

Lab File: VF061790.D

Acq: 28 Feb 2019 16:28

Tgt Ion: 91 Resp: 222236



Ion	Ratio	Lower	Upper
91	100		
126	32.4	19.1	57.1




Abundance

Ion 91.00 (90.70 to 91.70): VF0

Ion 126.00 (125.70 to 126.70): V









#83
tert-Butylbenzene
Concen: 21.761 ug/l
RT: 12.12 min Scan# 1165
Delta R.T. -0.02 min
Lab File: VF061790.D
Acq: 28 Feb 2019 16:28




Instrument :
MSVOA_F
ClientSampleId :
VF0228SBS01


Tot Ion:119 Resp: 288921
Ion Ratio Lower Upper
119 100
91 54.5 26.3 78.8
134 25.9 12.9 38.7



#84
1,2,4-Trimethylbenzene
Concen: 22.824 ug/l
RT: 12.19 min Scan# 1172
Delta R.T. -0.03 min
Lab File: VF061790.D
Acq: 28 Feb 2019 16:28

Tgt Ion:105 Resp: 271809
Ion Ratio Lower Upper
105 100
120 47.0 26.1 78.3





#85

sec-Butylbenzene

Concen: 22.448 ug/l

RT: 12.30 min Scan# 1183

Delta R.T. -0.02 min

Lab File: VF061790.D

Acq: 28 Feb 2019 16:28

Instrument :

MSVOA_F

ClientSampleId :


VF0228SBS01

Tot Ion:105 Resp: 385411

Ion Ratio Lower Upper




105 100

134 20.8 10.8 32.4



Abundance Ion 105.00 (104.70 to 105.70): V

Ion 134.00 (133.70 to 134.70): V



#86

p-Isopropyltoluene

Concen: 22.849 ug/l

RT: 12.45 min Scan# 1198

Delta R.T. -0.02 min

Lab File: VF061790.D

Acq: 28 Feb 2019 16:28


Tgt Ion:119 Resp: 329238

Ion Ratio Lower Upper

119 100

134 25.5 13.7 41.1



91 21.2 10.8 32.4




Abundance Ion 119.00 (118.70 to 119.70): V

Ion 134.00 (133.70 to 134.70): V

Ion 91.00 (90.70 to 91.70): V





#87

1,3-Dichlorobenzene

Concen: 21.883 ug/l

RT: 12.47 min Scan# 1200

Delta R.T. -0.02 min





Lab File: VF061790.D

Acq: 28 Feb 2019 16:28

Instrument :
MSVOA_F
ClientSampleId :
VF0228SBSD01

Tot Ion:146 Resp: 158070

Ion	Ratio	Lower	Upper
146	100		
111	35.7	17.5	52.6
148	65.3	31.4	94.3



#88

1,4-Dichlorobenzene

Concen: 22.042 ug/l

RT: 12.55 min Scan# 1209




Delta R.T. -0.02 min


Lab File: VF061790.D

Acq: 28 Feb 2019 16:28

Tgt Ion:146 Resp: 153125

Ion	Ratio	Lower	Upper
146	100		
111	34.7	18.4	55.4
148	66.5	32.2	96.6









#89
n-Butylbenzene
Concen: 23.603 ug/l
RT: 12.83 min Scan# 1237
Delta R.T. -0.02 min
Lab File: VF061790.D
Acq: 28 Feb 2019 16:28




Instrument :
MSVOA_F
ClientSampleId :
VF0228SBSD01


Tot Ion: 91 Resp: 321149
Ion Ratio Lower Upper
91 100
92 63.0 31.4 94.3
134 25.1 12.3 36.9



#90
Hexachloroethane
Concen: 22.992 ug/l
RT: 12.91 min Scan# 1245
Delta R.T. -0.02 min
Lab File: VF061790.D
Acq: 28 Feb 2019 16:28

Tgt Ion: 117 Resp: 77051
Ion Ratio Lower Upper
117 100
201 71.6 36.2 108.6









#91
 1,2-Dichlorobenzene
 Concen: 22.371 ug/l
 RT: 12.93 min Scan# 1247
 Delta R.T. -0.02 min
 Lab File: VF061790.D
 Acq: 28 Feb 2019 16:28




Instrument :
 MSVOA_F
 ClientSampleId :
 VF0228SBSD01


Tot Ion:146 Resp: 153704
 Ion Ratio Lower Upper
 146 100
 111 39.5 20.2 60.5
 148 65.1 32.3 96.9



#92
 1,2-Dibromo-3-Chloropropane
 Concen: 21.497 ug/l
 RT: 13.63 min Scan# 1318
 Delta R.T. -0.02 min
 Lab File: VF061790.D
 Acq: 28 Feb 2019 16:28

Tgt Ion: 75 Resp: 9158
 Ion Ratio Lower Upper
 75 100
 155 107.8 48.9 146.8
 157 119.9 58.3 174.8









#93
 1,2,4-Trichlorobenzene
 Concen: 21.834 ug/l
 RT: 14.19 min Scan# 1375
 Delta R.T. -0.02 min
 Lab File: VF061790.D
 Acq: 28 Feb 2019 16:28




Instrument :
 MSVOA_F
 ClientSampleId :
 VF0228SBSD01


Tot Ion	Ratio	Lower	Upper
180	100		
182	99.4	48.6	145.9
145	31.8	15.3	45.9



#94
 Hexachlorobutadiene
 Concen: 22.537 ug/l
 RT: 14.18 min Scan# 1374
 Delta R.T. -0.02 min
 Lab File: VF061790.D
 Acq: 28 Feb 2019 16:28

Tgt Ion	Ratio	Lower	Upper
225	100		
223	64.4	30.9	92.7
227	61.9	32.1	96.3





#95

Naphthalene

Concen: 19.610 ug/l

RT: 14.45 min Scan# 1401

Delta R.T. -0.02 min

Lab File: VF061790.D

Acq: 28 Feb 2019 16:28

Instrument :

MSVOA_F

ClientSampleId :

VF0228SBSD01


Tot Ion:128 Resp: 168965

Ion Ratio Lower Upper

128 100

127 11.1 9.8 14.6




129 11.2 9.4 14.2



Abundance Ion 128.00 (127.70 to 128.70): V

Ion 127.00 (126.70 to 127.70): V

Ion 129.00 (128.70 to 129.70): V



#96

1,2,3-Trichlorobenzene

Concen: 21.003 ug/l

RT: 14.60 min Scan# 1416

Delta R.T. -0.02 min

Lab File: VF061790.D

Acq: 28 Feb 2019 16:28


Tgt Ion:180 Resp: 94934

Ion Ratio Lower Upper

180 100

182 95.8 49.0 147.2

145 29.7 15.4 46.1



Abundance Ion 180.00 (179.70 to 180.70): V

Ion 182.00 (181.70 to 182.70): V

Ion 145.00 (144.70 to 145.70): V

