

Data Path : Z:\VOASRV\HPCHEM1\MSVOA F\DATA\VF030519\
 Data File : VF061851.D
 Acq On : 5 Mar 2019 18:03
 Operator : VA/AP
 Sample : K1712-15
 Misc : 6.13g/5mL/MSVOA-F/SOIL
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_F
 ClientSampleId :
 Z-2-3

Manual Integrations
 APPROVED

MMDadoda
 3/8/2019 12:12:19 AM

Quant Time: Mar 06 06:16:34 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_F\METHODS\82F030419S.M
 Quant Title : SW846 8260
 QLast Update : Tue Mar 05 04:50:44 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	4.87	168	269733	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	5.60	114	421815	50.00	ug/l	0.00
63) Chlorobenzene-d5	9.78	117	299311	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.55	152	107414	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	4.85	65	76546	38.97	ug/l	0.00
Spiked Amount 50.000			Recovery	=	77.94%	
35) Dibromofluoromethane	4.12	113	114111	38.76	ug/l	0.00
Spiked Amount 50.000			Recovery	=	77.52%	
50) Toluene-d8	7.55	98	268489	30.69	ug/l	0.00
Spiked Amount 50.000			Recovery	=	61.38%	
62) 4-Bromofluorobenzene	11.41	95	87201	25.44	ug/l	0.00
Spiked Amount 50.000			Recovery	=	50.88%	
Target Compounds						
16) Acetone	2.24	43	17200	41.802	ug/l #	79
20) Methylene Chloride	2.25	84	14679m	1.292	ug/l	
86) p-Isopropyltoluene	12.46	119	17610	2.525	ug/l	93
95) Naphthalene	14.46	128	5378	1.345	ug/l #	91

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA F\DATA\VF030519\
Data File : VF061851.D
Acq On : 5 Mar 2019 18:03
Operator : VA/AP
Sample : K1712-15
Misc : 6.13g/5mL/MSVOA-F/SOIL
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_F
ClientSampleId :
Z-2-3

Manual Integrations
APPROVED

MMDadoda
3/8/2019 12:12:19 AM

Quant Time: Mar 06 06:16:34 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_F\METHODS\82F030419S.M
Quant Title : SW846 8260
QLast Update : Tue Mar 05 04:50:44 2019
Response via : Initial Calibration

