

Data Path : Z:\VOASRV\HPCHEM1\MSVOA F\DATA\VF031519\
 Data File : VF061964.D
 Acq On : 15 Mar 2019 16:41
 Operator : VA/AP
 Sample : K1928-07
 Misc : 6.68g/5mL/MSVOA-F/SOIL
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_F
 ClientSampleId :
 A1-2-H2(2.5-3)

Manual Integrations
 APPROVED

MMDadoda
 3/17/2019 5:22:27 AM

Quant Time: Mar 16 00:20:15 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_F\METHODS\82F030419S.M
 Quant Title : SW846 8260
 QLast Update : Tue Mar 05 04:50:44 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	4.87	168	234603	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	5.60	114	401366	50.00	ug/l	0.00
63) Chlorobenzene-d5	9.78	117	330381	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.56	152	190317	50.00	ug/l	0.01
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	4.84	65	107482	62.91	ug/l	0.00
Spiked Amount 50.000			Recovery	=	125.82%	
35) Dibromofluoromethane	4.11	113	141636	50.57	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.14%	
50) Toluene-d8	7.56	98	380456	45.70	ug/l	0.00
Spiked Amount 50.000			Recovery	=	91.40%	
62) 4-Bromofluorobenzene	11.42	95	165615	50.78	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.56%	
Target Compounds						
16) Acetone	2.24	43	16335	45.644	ug/l #	88
20) Methylene Chloride	2.26	84	23988m	7.601	ug/l	

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA F\DATA\VF031519\
Data File : VF061964.D
Acq On : 15 Mar 2019 16:41
Operator : VA/AP
Sample : K1928-07
Misc : 6.68g/5mL/MSVOA-F/SOIL
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_F
ClientSampleId :
A1-2-H2(2.5-3)

Manual Integrations
APPROVED

MMDadoda
3/17/2019 5:22:27 AM

Quant Time: Mar 16 00:20:15 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_F\METHODS\82F030419S.M
Quant Title : SW846 8260
QLast Update : Tue Mar 05 04:50:44 2019
Response via : Initial Calibration

