

Data Path : Z:\VOASRV\HPCHEM1\MSVOA F\DATA\VF031519\
 Data File : VF061968.D
 Acq On : 15 Mar 2019 18:37
 Operator : VA/AP
 Sample : K1928-11
 Misc : 6.51g/5mL/MSVOA-F/SOIL
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_F
 ClientSampleId :
 SI-2(1.5-2)

Quant Time: Mar 16 00:36:21 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_F\METHODS\82F030419S.M
 Quant Title : SW846 8260
 QLast Update : Tue Mar 05 04:50:44 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	4.88	168	163666	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	5.61	114	255428	50.00	ug/l	0.00
63) Chlorobenzene-d5	9.79	117	215280	50.00	ug/l	0.01
72) 1,4-Dichlorobenzene-d4	12.56	152	110596	50.00	ug/l	0.01
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	4.86	65	62857	52.74	ug/l	0.01
Spiked Amount 50.000			Recovery	=	105.48%	
35) Dibromofluoromethane	4.13	113	77729	43.61	ug/l	0.01
Spiked Amount 50.000			Recovery	=	87.22%	
50) Toluene-d8	7.57	98	228813	43.19	ug/l	0.00
Spiked Amount 50.000			Recovery	=	86.38%	
62) 4-Bromofluorobenzene	11.42	95	94960	45.75	ug/l	0.00
Spiked Amount 50.000			Recovery	=	91.50%	
Target Compounds						
52) Toluene	7.63	92	44857	12.023	ug/l	90
67) Ethyl Benzene	9.91	91	38958	5.683	ug/l	98
68) m/p-Xylenes	10.15	106	48527	18.478	ug/l	92
69) o-Xylene	10.72	106	19560	6.917	ug/l	83
78) n-propylbenzene	11.61	91	14957	1.740	ug/l	98
80) 1,3,5-Trimethylbenzene	11.84	105	16455	2.692	ug/l	90
84) 1,2,4-Trimethylbenzene	12.21	105	53820	8.981	ug/l	93
95) Naphthalene	14.47	128	55739	13.541	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA F\DATA\VF031519\
Data File : VF061968.D
Acq On : 15 Mar 2019 18:37
Operator : VA/AP
Sample : K1928-11
Misc : 6.51g/5mL/MSVOA-F/SOIL
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_F
ClientSampleId :
SI-2(1.5-2)

Quant Time: Mar 16 00:36:21 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_F\METHODS\82F030419S.M
Quant Title : SW846 8260
QLast Update : Tue Mar 05 04:50:44 2019
Response via : Initial Calibration

