

Data Path : Z:\VOASRV\HPCHEM1\MSVOA F\DATA\VF031819\
 Data File : VF061989.D
 Acq On : 18 Mar 2019 10:55
 Operator : VA/AP
 Sample : VSTDCCC050
 Misc : 5.00g/5mL/MSVOA-F/SOIL
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_F
 ClientSampleId :

Quant Time: Mar 19 01:18:09 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_F\METHODS\82F031619S.M
 Quant Title : SW846 8260
 QLast Update : Mon Mar 18 07:28:40 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	4.85	168	280551	50.00	ug/l	-0.02
34) 1,4-Difluorobenzene	5.58	114	414133	50.00	ug/l	-0.02
63) Chlorobenzene-d5	9.75	117	355066	50.00	ug/l	-0.02
72) 1,4-Dichlorobenzene-d4	12.54	152	184857	50.00	ug/l	-0.01

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	4.82	65	99746	49.39	ug/l	-0.02
Spiked Amount			Recovery	=	98.78%	
35) Dibromofluoromethane	4.09	113	145852	52.45	ug/l	-0.02
Spiked Amount			Recovery	=	104.90%	
50) Toluene-d8	7.53	98	430312	52.93	ug/l	-0.02
Spiked Amount			Recovery	=	105.86%	
62) 4-Bromofluorobenzene	11.40	95	165018	51.10	ug/l	-0.01
Spiked Amount			Recovery	=	102.20%	

Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
2) Dichlorodifluoromethane	0.95	85	146420	48.957	ug/l	93
3) Chloromethane	1.09	50	153578	50.531	ug/l #	87
4) Vinyl Chloride	1.13	62	147302	50.112	ug/l	99
5) Bromomethane	1.31	94	105604	51.141	ug/l	99
6) Chloroethane	1.39	64	75448	49.720	ug/l	93
7) Trichlorofluoromethane	1.48	101	198601	50.998	ug/l	100
8) Diethyl Ether	1.62	74	41137	47.517	ug/l	77
9) 1,1,2-Trichlorotrifluoroet	1.83	101	135521	49.393	ug/l	93
10) Methyl Iodide	1.90	142	182276	37.957	ug/l	97
11) Tert butyl alcohol	2.56	59	25029	202.539	ug/l #	89
12) 1,1-Dichloroethene	1.79	96	116507	49.204	ug/l	96
13) Acrolein	1.99	56	34914	261.717	ug/l	99
14) Allyl chloride	2.09	41	146134	48.348	ug/l	96
15) Acrylonitrile	2.92	53	89947	239.355	ug/l	96
16) Acetone	2.22	43	124604	274.766	ug/l	99
17) Carbon Disulfide	1.81	76	343631	50.180	ug/l #	93
18) Methyl Acetate	2.54	43	2491	2.685	ug/l	96
19) Methyl tert-butyl Ether	2.41	73	205949	48.000	ug/l	97
20) Methylene Chloride	2.24	84	56371	24.145	ug/l	93
21) trans-1,2-Dichloroethene	2.29	96	107446	43.483	ug/l	98
22) Diisopropyl ether	2.78	45	352273	48.454	ug/l	98
23) Vinyl Acetate	3.17	43	718562	250.718	ug/l	97
24) 1,1-Dichloroethane	2.86	63	209588	49.579	ug/l	99
25) 2-Butanone	4.32	43	210001	233.822	ug/l	95
26) 2,2-Dichloropropane	3.58	77	98651	49.474	ug/l	100
27) cis-1,2-Dichloroethene	3.46	96	148861	47.796	ug/l	99
28) Bromochloromethane	3.69	49	78976	45.024	ug/l	96
29) Tetrahydrofuran	4.05	42	80600	206.355	ug/l	100
30) Chloroform	3.83	83	237560	49.598	ug/l	94
31) Cyclohexane	3.67	56	121486	29.332	ug/l	96
32) 1,1,1-Trichloroethane	4.07	97	141011	44.330	ug/l	97
36) 1,1-Dichloropropene	4.26	75	192469	51.465	ug/l	99
37) Ethyl Acetate	4.09	43	71727	42.715	ug/l	94
38) Carbon Tetrachloride	3.96	117	142987	52.315	ug/l	96

Data Path : Z:\VOASRV\HPCHEM1\MSVOA F\DATA\VF031819\
 Data File : VF061989.D
 Acq On : 18 Mar 2019 10:55
 Operator : VA/AP
 Sample : VSTDCCC050
 Misc : 5.00g/5mL/MSVOA-F/SOIL
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_F
 ClientSampleId :

Quant Time: Mar 19 01:18:09 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_F\METHODS\82F031619S.M
 Quant Title : SW846 8260
 QLast Update : Mon Mar 18 07:28:40 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
39) Methylcyclohexane	5.43	83	226710	51.637	ug/l	94
40) Benzene	4.61	78	501534	50.611	ug/l	99
41) Methacrylonitrile	4.72	41	40085	44.680	ug/l	94
42) 1,2-Dichloroethane	4.92	62	122896	49.818	ug/l	98
43) Isopropyl Acetate	6.93	43	88991	44.792	ug/l #	92
44) Trichloroethene	5.48	130	151564	49.928	ug/l	98
45) 1,2-Dichloropropane	6.21	63	112865	48.126	ug/l	95
46) Dibromomethane	6.06	93	70516	50.545	ug/l	89
47) Bromodichloromethane	6.36	83	162096	49.283	ug/l	98
48) Methyl methacrylate	6.69	41	58903	45.873	ug/l	95
49) 1,4-Dioxane	6.68	88	10717	938.311	ug/l #	75
51) 4-Methyl-2-Pentanone	8.23	43	320575	227.223	ug/l	99
52) Toluene	7.60	92	281381	46.472	ug/l	99
53) t-1,3-Dichloropropene	8.25	75	130949	47.159	ug/l	99
54) cis-1,3-Dichloropropene	7.27	75	183519	50.652	ug/l	96
55) 1,1,2-Trichloroethane	8.47	97	82410	48.766	ug/l	97
56) Ethyl methacrylate	8.59	69	92007	47.895	ug/l	97
57) 1,3-Dichloropropane	8.83	76	136092	48.278	ug/l	100
58) 2-Chloroethyl Vinyl ether	7.27	63	73114	210.130	ug/l	91
59) 2-Hexanone	9.47	43	245604	240.545	ug/l	99
60) Dibromochloromethane	8.69	129	121948	51.981	ug/l	93
61) 1,2-Dibromoethane	8.96	107	85459	47.145	ug/l	99
64) Tetrachloroethene	8.13	164	128945	49.960	ug/l	96
65) Chlorobenzene	9.78	112	326350	49.367	ug/l	98
66) 1,1,1,2-Tetrachloroethane	9.91	131	135385	52.835	ug/l	97
67) Ethyl Benzene	9.88	91	565835	49.990	ug/l	98
68) m/p-Xylenes	10.11	106	439727	103.291	ug/l	99
69) o-Xylene	10.69	106	228057	50.064	ug/l	100
70) Styrene	10.76	104	321980	51.020	ug/l	99
71) Bromoform	10.75	173	67150	51.809	ug/l	99
73) Isopropylbenzene	11.11	105	661375	51.787	ug/l	99
74) N-amyl acetate	11.36	43	145114	49.590	ug/l #	93
75) 1,1,2,2-Tetrachloroethane	11.68	83	100564	47.706	ug/l	96
76) 1,2,3-Trichloropropane	11.78	75	67431	45.746	ug/l	96
77) Bromobenzene	11.47	156	155971	52.737	ug/l	92
78) n-propylbenzene	11.58	91	737097	50.194	ug/l	99
79) 2-Chlorotoluene	11.70	91	415863	50.171	ug/l	98
80) 1,3,5-Trimethylbenzene	11.81	105	519751	50.382	ug/l	97
81) trans-1,4-Dichloro-2-buten	11.85	75	45117	69.187	ug/l	97
82) 4-Chlorotoluene	11.88	91	406863	48.539	ug/l	96
83) tert-Butylbenzene	12.11	119	567115	52.565	ug/l	99
84) 1,2,4-Trimethylbenzene	12.18	105	507618	49.125	ug/l	97
85) sec-Butylbenzene	12.29	105	715719	49.100	ug/l	98
86) p-Isopropyltoluene	12.44	119	631721	51.937	ug/l	96
87) 1,3-Dichlorobenzene	12.46	146	289403	49.788	ug/l	98
88) 1,4-Dichlorobenzene	12.55	146	289347	51.717	ug/l	97
89) n-Butylbenzene	12.82	91	585092	51.192	ug/l	97
90) Hexachloroethane	12.90	117	155660	56.078	ug/l	89
91) 1,2-Dichlorobenzene	12.92	146	270456	50.236	ug/l	97
92) 1,2-Dibromo-3-Chloropropan	13.62	75	17081	50.494	ug/l	66


Data Path : Z:\VOASRV\HPCHEM1\MSVOA F\DATA\VF031819\
Data File : VF061989.D
Acq On : 18 Mar 2019 10:55
Operator : VA/AP
Sample : VSTDCCC050
Misc : 5.00g/5mL/MSVOA-F/SOIL
ALS Vial : 2 Sample Multiplier: 1

Instrument :
MSVOA_F
ClientSampleId :

Quant Time: Mar 19 01:18:09 2019
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_F\METHODS\82F031619S.M
Quant Title : SW846 8260
QLast Update : Mon Mar 18 07:28:40 2019
Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
93) 1,2,4-Trichlorobenzene	14.19	180	190692	50.865	ug/l	95
94) Hexachlorobutadiene	14.18	225	129092	51.868	ug/l	98
95) Naphthalene	14.44	128	351028	52.060	ug/l	99
96) 1,2,3-Trichlorobenzene	14.59	180	176363	51.133	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed



#1

Pentafluorobenzene

Concen: 50.000 ug/l

RT: 4.85 min Scan# 428

Delta R.T. -0.02 min

Lab File: VF061989.D

Acq: 18 Mar 2019 10:55


Instrument :
MSVOA_F
ClientSampleId :

Tot Ion:168 Resp: 280551




Ion Ratio Lower Upper

168 100

99 48.4 42.4 63.6



Abundance Ion 168.00 (167.70 to 168.70): VF0



#2

Dichlorodifluoromethane

Concen: 48.957 ug/l

RT: 0.95 min Scan# 33

Delta R.T. -0.01 min

Lab File: VF061989.D


Acq: 18 Mar 2019 10:55

Tgt Ion: 85 Resp: 146420



Ion Ratio Lower Upper


85 100

87 34.8 15.4 46.2



Abundance Ion 85.00 (84.70 to 85.70): VF0





#3

Chloromethane

Concen: 50.531 ug/l

RT: 1.09 min Scan# 47

Delta R.T. -0.00 min

Lab File: VF061989.D

Acq: 18 Mar 2019 10:55

Instrument :

MSVOA_F





ClientSampleId :

Tot Ion: 50 Resp: 153578

Ion Ratio Lower Upper

50 100

52 35.2 22.6 34.0#



#4

Vinyl Chloride

Concen: 50.112 ug/l

RT: 1.13 min Scan# 51

Delta R.T. -0.01 min

Lab File: VF061989.D




Acq: 18 Mar 2019 10:55


Tgt Ion: 62 Resp: 147302

Ion Ratio Lower Upper

62 100

64 33.6 27.4 41.2





#5

Bromomethane

Concen: 51.141 ug/l

RT: 1.31 min Scan# 69

Delta R.T. -0.01 min

Lab File: VF061989.D

Acq: 18 Mar 2019 10:55

Instrument :

MSVOA_F





ClientSampleId :

Tot Ion: 94 Resp: 105604

Ion Ratio Lower Upper

94 100

96 95.0 75.0 112.6



#6

Chloroethane

Concen: 49.720 ug/l

RT: 1.39 min Scan# 77

Delta R.T. -0.00 min

Lab File: VF061989.D




Acq: 18 Mar 2019 10:55


Tgt Ion: 64 Resp: 75448

Ion Ratio Lower Upper

64 100

66 30.5 27.6 41.4









#7

Trichlorofluoromethane
Concen: 50.998 ug/l
RT: 1.48 min Scan# 86
Delta R.T. 0.01 min
Lab File: VF061989.D
Acq: 18 Mar 2019 10:55

Instrument :
MSVOA_F
ClientSampled :




Tot Ion:101 Resp: 198601
Ion Ratio Lower Upper
101 100
103 58.7 47.3 70.9




#8

Diethyl Ether
Concen: 47.517 ug/l
RT: 1.62 min Scan# 101
Delta R.T. -0.00 min
Lab File: VF061989.D
Acq: 18 Mar 2019 10:55

Tgt Ion: 74 Resp: 41137
Ion Ratio Lower Upper
74 100
45 136.7 56.3 168.8





#9

1,1,2-Trichlorotrifluoroethane

Concen: 49.393 ug/l

RT: 1.83 min Scan# 122

Delta R.T. -0.01 min

Lab File: VF061989.D

Acq: 18 Mar 2019 10:55

Instrument :

MSVOA_F

ClientSampleId :


Tot Ion:101 Resp: 135521

Ion Ratio Lower Upper

101 100

85 41.1 36.8 55.2




151 72.0 62.2 93.2



Abundance Ion 101.00 (100.70 to 101.70): V

Ion 85.00 (84.70 to 85.70): VF0

Ion 151.00 (150.70 to 151.70): V



#10

Methyl Iodide

Concen: 37.957 ug/l

RT: 1.90 min Scan# 129

Delta R.T. -0.01 min

Lab File: VF061989.D

Acq: 18 Mar 2019 10:55


Tgt Ion:142 Resp: 182276

Ion Ratio Lower Upper

142 100

127 47.8 39.7 59.5



141 16.1 10.8 16.2




Abundance Ion 142.00 (141.70 to 142.70): V

Ion 127.00 (126.70 to 127.70): V

Ion 141.00 (140.70 to 141.70): V









#11

Tert butyl alcohol
Concen: 202.539 ug/l
RT: 2.56 min Scan# 196
Delta R.T. -0.02 min
Lab File: VF061989.D
Acq: 18 Mar 2019 10:55

Instrument :
MSVOA_F
ClientSampled :




Tot Ion: 59 Resp: 25029
Ion Ratio Lower Upper
59 100
57 15.1 8.7 13.1#




#12

1,1-Dichloroethene
Concen: 49.204 ug/l
RT: 1.79 min Scan# 118
Delta R.T. -0.00 min
Lab File: VF061989.D
Acq: 18 Mar 2019 10:55

Tgt Ion: 96 Resp: 116507
Ion Ratio Lower Upper
96 100
61 165.1 138.6 208.0
98 66.5 53.2 79.8





#13

Acrolein

Concen: 261.717 ug/l

RT: 1.99 min Scan# 138

Delta R.T. -0.01 min

Lab File: VF061989.D

Acq: 18 Mar 2019 10:55

Instrument :

MSVOA_F


ClientSampled :

Tot Ion: 56 Resp: 34914

Ion Ratio Lower Upper

56 100

55 62.3 49.1 73.7



Abundance Ion 56.00 (55.70 to 56.70): VF0

Ion 55.00 (54.70 to 55.70): VF0

1.99

15000



10000

5000

0

Time-->

1.90 2.00 2.10 2.20



#14

Allyl chloride

Concen: 48.348 ug/l

RT: 2.09 min Scan# 148

Delta R.T. -0.01 min

Lab File: VF061989.D

Acq: 18 Mar 2019 10:55


Tgt Ion: 41 Resp: 146134

Ion Ratio Lower Upper

41 100

39 79.0 66.8 100.2

76 39.4 31.1 46.7



Abundance Ion 41.00 (40.70 to 41.70): VF0

Ion 39.00 (38.70 to 39.70): VF0

Ion 76.00 (75.70 to 76.70): VF0

2.09

100000

80000

60000


40000


20000

0

Time-->

2.00 2.05 2.10 2.15





#15

Acrylonitrile

Concen: 239.355 ug/l

RT: 2.92 min Scan# 233





Delta R.T. -0.01 min

Lab File: VF061989.D

Acq: 18 Mar 2019 10:55

Instrument :
MSVOA_F
ClientSampleId :

Tot Ion:	53	Resp:	89947
Ion	Ratio	Lower	Upper
53	100		
52	84.7	64.4	96.6
51	36.0	30.3	45.5



#16

Acetone

Concen: 274.766 ug/l




RT: 2.22 min Scan# 162


Delta R.T. -0.01 min

Lab File: VF061989.D

Acq: 18 Mar 2019 10:55

Tgt Ion:	43	Resp:	124604
Ion	Ratio	Lower	Upper
43	100		
58	25.3	19.9	29.9









#17
Carbon Disulfide
Concen: 50.180 ug/l
RT: 1.81 min Scan# 120
Delta R.T. -0.01 min
Lab File: VF061989.D
Acq: 18 Mar 2019 10:55




Instrument :
MSVOA_F
ClientSampleId :


Tot Ion: 76 Resp: 343631
Ion Ratio Lower Upper
76 100
78 11.1 6.8 10.2#



#18
Methyl Acetate
Concen: 2.685 ug/l
RT: 2.54 min Scan# 194
Delta R.T. 0.11 min
Lab File: VF061989.D
Acq: 18 Mar 2019 10:55

Tgt Ion: 43 Resp: 2491
Ion Ratio Lower Upper
43 100
74 13.9 12.6 18.8









#19
Methyl tert-butyl Ether
Concen: 48.000 ug/l
RT: 2.41 min Scan# 181
Delta R.T. -0.02 min
Lab File: VF061989.D
Acq: 18 Mar 2019 10:55




Instrument :
MSVOA_F
ClientSampleId :


Tot Ion: 73 Resp: 205949
Ion Ratio Lower Upper
73 100
57 17.7 15.3 22.9



#20
Methylene Chloride
Concen: 24.145 ug/l
RT: 2.24 min Scan# 164
Delta R.T. 0.07 min
Lab File: VF061989.D
Acq: 18 Mar 2019 10:55

Tgt Ion: 84 Resp: 56371
Ion Ratio Lower Upper
84 100
49 124.2 107.0 160.6
51 40.0 37.2 55.8
86 71.3 55.3 82.9





#21

trans-1,2-Dichloroethene

Concen: 43.483 ug/l

RT: 2.29 min Scan# 169

Delta R.T. -0.01 min

Lab File: VF061989.D

Acq: 18 Mar 2019 10:55

Instrument :
MSVOA_F
ClientSampleId :


Tot Ion: 96 Resp: 107446

Ion Ratio Lower Upper




96 100

61 149.6 117.4 176.2

98 66.2 51.2 76.8



Abundance Ion 96.00 (95.70 to 96.70): VF0
Ion 61.00 (60.70 to 61.70): VF0
Ion 98.00 (97.70 to 98.70): VF0



#22

Diisopropyl ether

Concen: 48.454 ug/l

RT: 2.78 min Scan# 218

Delta R.T. -0.02 min

Lab File: VF061989.D

Acq: 18 Mar 2019 10:55

Tgt Ion: 45 Resp: 352273


Ion Ratio Lower Upper

45 100



43 40.5 33.7 50.5


87 20.4 17.1 25.7

59 10.8 8.0 12.0



Abundance Ion 45.00 (44.70 to 45.70): VF0
Ion 43.00 (42.70 to 43.70): VF0
Ion 87.00 (86.70 to 87.70): VF0
Ion 59.00 (58.70 to 59.70): VF0









#23
 Vinyl Acetate
 Concen: 250.718 ug/l
 RT: 3.17 min Scan# 258
 Delta R.T. -0.02 min
 Lab File: VF061989.D
 Acq: 18 Mar 2019 10:55




Instrument :
 MSVOA_F
 ClientSampleId :


Tot Ion: 43 Resp: 718562
 Ion Ratio Lower Upper
 43 100
 86 9.2 6.5 9.7



#24
 1,1-Dichloroethane
 Concen: 49.579 ug/l
 RT: 2.86 min Scan# 226
 Delta R.T. -0.01 min
 Lab File: VF061989.D
 Acq: 18 Mar 2019 10:55

Tgt Ion: 63 Resp: 209588
 Ion Ratio Lower Upper
 63 100
 98 5.0 2.7 8.1
 100 3.6 1.7 5.0





#25

2-Butanone

Concen: 233.822 ug/l

RT: 4.32 min Scan# 374

Delta R.T. -0.02 min

Lab File: VF061989.D

Acq: 18 Mar 2019 10:55





Instrument :
MSVOA_F
ClientSampled :

Tot Ion: 43 Resp: 210001

Ion Ratio Lower Upper

43 100

72 22.3 19.8 29.6



#26

2,2-Dichloropropane

Concen: 49.474 ug/l

RT: 3.58 min Scan# 299

Delta R.T. -0.02 min

Lab File: VF061989.D




Acq: 18 Mar 2019 10:55


Tgt Ion: 77 Resp: 98651

Ion Ratio Lower Upper

77 100

97 21.5 10.8 32.4









#27

cis-1,2-Dichloroethene
 Concen: 47.796 ug/l
 RT: 3.46 min Scan# 287
 Delta R.T. -0.01 min
 Lab File: VF061989.D
 Acq: 18 Mar 2019 10:55

Instrument :
 MSVOA_F
 ClientSampleId :




Tot Ion: 96 Resp: 148861
 Ion Ratio Lower Upper
 96 100
 61 123.7 0.0 246.6
 98 65.7 0.0 129.2




#28

Bromochloromethane
 Concen: 45.024 ug/l
 RT: 3.69 min Scan# 311
 Delta R.T. -0.02 min
 Lab File: VF061989.D
 Acq: 18 Mar 2019 10:55

Tgt Ion: 49 Resp: 78976
 Ion Ratio Lower Upper
 49 100
 129 2.9 0.0 5.8
 130 87.8 73.1 109.7









#29
Tetrahydrofuran
Concen: 206.355 ug/l
RT: 4.05 min Scan# 347
Delta R.T. -0.02 min
Lab File: VF061989.D
Acq: 18 Mar 2019 10:55




Instrument :
MSVOA_F
ClientSampleId :


Tot Ion: 42 Resp: 80600
Ion Ratio Lower Upper
42 100
72 42.6 34.0 51.0
71 43.3 34.2 51.4



#30
Chloroform
Concen: 49.598 ug/l
RT: 3.83 min Scan# 325
Delta R.T. -0.03 min
Lab File: VF061989.D
Acq: 18 Mar 2019 10:55

Tgt Ion: 83 Resp: 237560
Ion Ratio Lower Upper
83 100
85 64.7 55.5 83.3





#31

Cyclohexane

Concen: 29.332 ug/l

RT: 3.67 min Scan# 309

Delta R.T. -0.01 min

Lab File: VF061989.D

Acq: 18 Mar 2019 10:55

Instrument :

MSVOA_F

ClientSampled :





Tot Ion: 56 Resp: 121486

Ion Ratio Lower Upper

56 100

69 33.7 25.8 38.8

84 91.5 69.5 104.3



#32

1,1,1-Trichloroethane

Concen: 44.330 ug/l

RT: 4.07 min Scan# 349

Delta R.T. -0.03 min

Lab File: VF061989.D

Acq: 18 Mar 2019 10:55




Tgt Ion: 97 Resp: 141011


Ion Ratio Lower Upper

97 100

99 64.4 49.0 73.4

61 49.9 41.4 62.0









#33
 1,2-Dichloroethane-d4
 Concen: 49.391 ug/l
 RT: 4.82 min Scan# 425
 Delta R.T. -0.02 min
 Lab File: VF061989.D
 Acq: 18 Mar 2019 10:55




Instrument :
 MSVOA_F
 ClientSampleId :


Tot Ion: 65 Resp: 99746
 Ion Ratio Lower Upper
 65 100
 67 58.5 0.0 109.4



#34
 1,4-Difluorobenzene
 Concen: 50.000 ug/l
 RT: 5.58 min Scan# 502
 Delta R.T. -0.02 min
 Lab File: VF061989.D
 Acq: 18 Mar 2019 10:55

Tgt Ion: 114 Resp: 414133
 Ion Ratio Lower Upper
 114 100
 63 17.9 0.0 36.4
 88 18.2 0.0 36.4





#35

Dibromofluoromethane

Concen: 52.446 ug/l

RT: 4.09 min Scan# 351

Delta R.T. -0.02 min

Lab File: VF061989.D

Acq: 18 Mar 2019 10:55

Instrument :
MSVOA_F
ClientSampled :


Tot Ion:113 Resp: 145852

Ion Ratio Lower Upper




113 100

111 100.4 80.8 121.2

192 17.1 13.0 19.6



Abundance Ion 113.00 (112.70 to 113.70): V
Ion 111.00 (110.70 to 111.70): V
Ion 192.00 (191.70 to 192.70): V



#36

1,1-Dichloropropene

Concen: 51.465 ug/l

RT: 4.26 min Scan# 368

Delta R.T. -0.02 min

Lab File: VF061989.D

Acq: 18 Mar 2019 10:55


Tgt Ion: 75 Resp: 192469

Ion Ratio Lower Upper



75 100


110 35.1 17.8 53.3

77 30.9 24.2 36.2



Abundance Ion 75.00 (74.70 to 75.70): VF0
Ion 110.00 (109.70 to 110.70): V
Ion 77.00 (76.70 to 77.70): VF0









#37
Ethyl Acetate
Concen: 42.715 ug/l
RT: 4.09 min Scan# 351
Delta R.T. -0.02 min
Lab File: VF061989.D
Acq: 18 Mar 2019 10:55




Instrument :
MSVOA_F
ClientSampleId :


Tot Ion: 43 Resp: 71727
Ion Ratio Lower Upper
43 100
61 69.8 59.8 89.6
70 8.0 7.8 11.8



#38
Carbon Tetrachloride
Concen: 52.315 ug/l
RT: 3.96 min Scan# 338
Delta R.T. -0.02 min
Lab File: VF061989.D
Acq: 18 Mar 2019 10:55

Tgt Ion: 117 Resp: 142987
Ion Ratio Lower Upper
117 100
119 93.7 73.0 109.4
121 32.3 23.0 34.6









#39
Methylvlcyclohexane
Concen: 51.637 ug/l
RT: 5.43 min Scan# 487
Delta R.T. -0.02 min
Lab File: VF061989.D
Acq: 18 Mar 2019 10:55




Instrument :
MSVOA_F
ClientSampleId :


Tot Ion: 83 Resp: 226710
Ion Ratio Lower Upper
83 100
55 82.0 69.7 104.5
98 44.2 39.4 59.0



#40
Benzene
Concen: 50.611 ug/l
RT: 4.61 min Scan# 404
Delta R.T. -0.02 min
Lab File: VF061989.D
Acq: 18 Mar 2019 10:55

Tgt Ion: 78 Resp: 501534
Ion Ratio Lower Upper
78 100
77 23.4 19.0 28.4









#41
Methacrylonitrile
Concen: 44.680 ug/l
RT: 4.72 min Scan# 415
Delta R.T. -0.02 min
Lab File: VF061989.D
Acq: 18 Mar 2019 10:55




Instrument :
MSVOA_F
ClientSampleId :


Tot Ion:	41	Resp:	40085
Ion	Ratio	Lower	Upper
41	100		
39	60.1	46.2	69.4
67	68.8	49.8	74.8
52	47.8	42.7	64.1



#42
1,2-Dichloroethane
Concen: 49.818 ug/l
RT: 4.92 min Scan# 435
Delta R.T. -0.02 min
Lab File: VF061989.D
Acq: 18 Mar 2019 10:55

Tgt Ion:	62	Resp:	122896
Ion	Ratio	Lower	Upper
62	100		
98	10.7	0.0	19.8





#43

Isopropyl Acetate

Concen: 44.792 ug/l

RT: 6.93 min Scan# 639

Delta R.T. -0.03 min

Lab File: VF061989.D

Acq: 18 Mar 2019 10:55

Instrument :

MSVOA_F

ClientSampleId :


Tot Ion: 43 Resp: 88991

Ion Ratio Lower Upper

43 100

61 32.5 30.1 45.1




87 0.0 0.4 0.6#



Abundance Ion 43.00 (42.70 to 43.70): VF0

Ion 61.00 (60.70 to 61.70): VF0

Ion 87.00 (86.70 to 87.70): VF0



#44

Trichloroethene

Concen: 49.928 ug/l

RT: 5.48 min Scan# 492

Delta R.T. -0.02 min

Lab File: VF061989.D


Acq: 18 Mar 2019 10:55

Tgt Ion:130 Resp: 151564

Ion Ratio Lower Upper



130 100


95 92.5 0.0 181.4



Abundance Ion 129.90 (129.60 to 130.60): V

Ion 94.90 (94.60 to 95.60): VF0





#45

1,2-Dichloropropane

Concen: 48.126 ug/l

RT: 6.21 min Scan# 566

Delta R.T. -0.02 min

Lab File: VF061989.D

Acq: 18 Mar 2019 10:55

Instrument :

MSVOA_F


ClientSampled :

Tot Ion: 63 Resp: 112865

Ion Ratio Lower Upper




63 100

65 32.1 23.7 35.5



Abundance Ion 63.00 (62.70 to 63.70): VF0

Ion 65.00 (64.70 to 65.70): VF0



#46

Dibromomethane

Concen: 50.545 ug/l

RT: 6.06 min Scan# 551

Delta R.T. -0.02 min

Lab File: VF061989.D

Acq: 18 Mar 2019 10:55


Tgt Ion: 93 Resp: 70516

Ion Ratio Lower Upper

93 100

95 87.1 65.7 98.5



174 118.5 82.2 123.4




Abundance Ion 93.00 (92.70 to 93.70): VF0

Ion 95.00 (94.70 to 95.70): VF0

Ion 174.00 (173.70 to 174.70): V





#47

Bromodichloromethane

Concen: 49.283 ug/l

RT: 6.36 min Scan# 581

Delta R.T. -0.03 min

Lab File: VF061989.D

Acq: 18 Mar 2019 10:55

Instrument :

MSVOA_F

ClientSampleId :


Tot Ion: 83 Resp: 162096

Ion Ratio Lower Upper

83 100

85 65.8 51.4 77.2




127 8.7 6.5 9.7



Abundance Ion 83.00 (82.70 to 83.70): VF0

Ion 85.00 (84.70 to 85.70): VF0

Ion 127.00 (126.70 to 127.70): V



#48

Methyl methacrylate

Concen: 45.873 ug/l

RT: 6.69 min Scan# 615

Delta R.T. -0.02 min

Lab File: VF061989.D

Acq: 18 Mar 2019 10:55


Tgt Ion: 41 Resp: 58903

Ion Ratio Lower Upper

41 100

69 74.6 59.0 88.6



39 51.5 35.6 53.4




Abundance Ion 41.00 (40.70 to 41.70): VF0

Ion 69.00 (68.70 to 69.70): VF0

Ion 39.00 (38.70 to 39.70): VF0









#49
1,4-Dioxane
Concen: 938.311 ug/l
RT: 6.68 min Scan# 614
Delta R.T. -0.02 min
Lab File: VF061989.D
Acq: 18 Mar 2019 10:55




Instrument :
MSVOA_F
ClientSampleId :


Tot Ion: 88 Resp: 10717
Ion Ratio Lower Upper
88 100
43 53.0 31.2 46.8#
58 63.3 69.9 104.9#



#50
Toluene-d8
Concen: 52.931 ug/l
RT: 7.53 min Scan# 700
Delta R.T. -0.02 min
Lab File: VF061989.D
Acq: 18 Mar 2019 10:55

Tgt Ion: 98 Resp: 430312
Ion Ratio Lower Upper
98 100
100 70.0 56.8 85.2





#51

4-Methyl-2-Pentanone

Concen: 227.223 ug/l

RT: 8.23 min Scan# 771

Delta R.T. -0.03 min

Lab File: VF061989.D

Acq: 18 Mar 2019 10:55

Instrument :

MSVOA_F


ClientSampleId :

Tot Ion: 43 Resp: 320575

Ion Ratio Lower Upper




43 100

58 39.3 31.1 46.7



Abundance Ion 43.00 (42.70 to 43.70): VF0

Ion 58.00 (57.70 to 58.70): VF0



#52

Toluene

Concen: 46.472 ug/l

RT: 7.60 min Scan# 707

Delta R.T. -0.02 min

Lab File: VF061989.D


Acq: 18 Mar 2019 10:55

Tgt Ion: 92 Resp: 281381

Ion Ratio Lower Upper



92 100


91 166.3 134.0 201.0



Abundance Ion 92.00 (91.70 to 92.70): VF0

Ion 91.00 (90.70 to 91.70): VF0





#53

t-1,3-Dichloropropene

Concen: 47.159 ug/l

RT: 8.25 min Scan# 773

Delta R.T. -0.03 min

Lab File: VF061989.D

Acq: 18 Mar 2019 10:55

Instrument :

MSVOA_F





ClientSampleId :

Tot Ion: 75 Resp: 130949

Ion Ratio Lower Upper

75 100

77 32.7 25.5 38.3



#54

cis-1,3-Dichloropropene

Concen: 50.652 ug/l

RT: 7.27 min Scan# 674

Delta R.T. -0.03 min

Lab File: VF061989.D

Acq: 18 Mar 2019 10:55




Tgt Ion: 75 Resp: 183519


Ion Ratio Lower Upper

75 100

77 32.5 26.4 39.6

39 51.7 38.0 57.0









#55
 1,1,2-Trichloroethane
 Concen: 48.766 ug/l
 RT: 8.47 min Scan# 795
 Delta R.T. -0.02 min
 Lab File: VF061989.D
 Acq: 18 Mar 2019 10:55




Instrument :
 MSVOA_F
 ClientSampleId :


Tot Ion:	97	Resp:	82410
Ion	Ratio	Lower	Upper
97	100		
83	82.2	69.7	104.5
85	56.1	45.8	68.6
99	60.7	50.3	75.5



#56
 Ethyl methacrylate
 Concen: 47.895 ug/l
 RT: 8.59 min Scan# 808
 Delta R.T. -0.02 min
 Lab File: VF061989.D
 Acq: 18 Mar 2019 10:55

Tgt Ion:	69	Resp:	92007
Ion	Ratio	Lower	Upper
69	100		
41	71.2	54.7	82.1
39	36.8	27.9	41.9





#57

1,3-Dichloropropane

Concen: 48.278 ug/l

RT: 8.83 min Scan# 832

Delta R.T. -0.02 min

Lab File: VF061989.D

Acq: 18 Mar 2019 10:55

Instrument :

MSVOA_F





ClientSampleId :

Tot Ion: 76 Resp: 136092

Ion Ratio Lower Upper

76 100

78 31.7 25.4 38.2



#58

2-Chloroethyl Vinyl ether

Concen: 210.130 ug/l

RT: 7.27 min Scan# 674

Delta R.T. -0.02 min

Lab File: VF061989.D




Acq: 18 Mar 2019 10:55


Tgt Ion: 63 Resp: 73114

Ion Ratio Lower Upper

63 100

106 30.5 20.9 31.3









#59
2-Hexanone
Concen: 240.545 ug/l
RT: 9.47 min Scan# 897
Delta R.T. -0.02 min
Lab File: VF061989.D
Acq: 18 Mar 2019 10:55




Instrument :
MSVOA_F
ClientSampleId :


Tot Ion: 43 Resp: 245604
Ion Ratio Lower Upper
43 100
58 56.0 28.4 85.2



#60
Dibromochloromethane
Concen: 51.981 ug/l
RT: 8.69 min Scan# 818
Delta R.T. -0.02 min
Lab File: VF061989.D
Acq: 18 Mar 2019 10:55

Tgt Ion: 129 Resp: 121948
Ion Ratio Lower Upper
129 100
127 68.3 37.3 111.8





#61

1,2-Dibromoethane

Concen: 47.145 ug/l

RT: 8.96 min Scan# 845

Delta R.T. -0.03 min

Lab File: VF061989.D

Acq: 18 Mar 2019 10:55


Instrument :
MSVOA_F
ClientSampleId :

Tot Ion:107 Resp: 85459


Ion Ratio Lower Upper

107 100



109 90.0 72.8 109.2



Abundance Ion 107.00 (106.70 to 107.70): V
Ion 109.00 (108.70 to 109.70): V



Time-->



#62

4-Bromofluorobenzene

Concen: 51.096 ug/l

RT: 11.40 min Scan# 1092

Delta R.T. -0.01 min

Lab File: VF061989.D

Acq: 18 Mar 2019 10:55


Tgt Ion: 95 Resp: 165018

Ion Ratio Lower Upper


95 100

174 86.2 0.0 149.4


176 84.9 0.0 143.4




Abundance Ion 95.00 (94.70 to 95.70): VF0
Ion 174.00 (173.70 to 174.70): V
Ion 176.00 (175.70 to 176.70): V



Time-->









#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 9.75 min Scan# 925
Delta R.T. -0.02 min
Lab File: VF061989.D
Acq: 18 Mar 2019 10:55




Instrument :
MSVOA_F
ClientSampled :


Tot Ion:117 Resp: 355066
Ion Ratio Lower Upper
117 100
82 51.5 42.5 63.7
119 30.8 25.4 38.0



#64
Tetrachloroethene
Concen: 49.960 ug/l
RT: 8.13 min Scan# 761
Delta R.T. -0.02 min
Lab File: VF061989.D
Acq: 18 Mar 2019 10:55

Tgt Ion:164 Resp: 128945
Ion Ratio Lower Upper
164 100
166 122.3 100.0 150.0
129 84.5 72.4 108.6
131 81.1 67.1 100.7





#65

Chlorobenzene

Concen: 49.367 ug/l

RT: 9.78 min Scan# 928

Delta R.T. -0.02 min

Lab File: VF061989.D

Acq: 18 Mar 2019 10:55

Instrument :

MSVOA_F


ClientSampled :

Tot Ion:112 Resp: 326350

Ion Ratio Lower Upper

112 100

114 34.5 26.9 40.3



Abundance Ion 112.00 (111.70 to 112.70): V

Ion 114.00 (113.70 to 114.70): V

9.78



150000

100000

50000

0

Time--> 9.60 9.80 10.00



#66

1,1,1,2-Tetrachloroethane

Concen: 52.835 ug/l

RT: 9.91 min Scan# 941

Delta R.T. -0.02 min

Lab File: VF061989.D

Acq: 18 Mar 2019 10:55


Tgt Ion:131 Resp: 135385

Ion Ratio Lower Upper

131 100

133 95.5 48.0 144.2

119 67.4 35.9 107.7



Abundance Ion 131.00 (130.70 to 131.70): V

Ion 133.00 (132.70 to 133.70): V

Ion 119.00 (118.70 to 119.70): V

9.91

80000


60000


40000

20000

0

Time--> 9.80 9.90 10.00





#67

Ethyl Benzene

Concen: 49.990 ug/l

RT: 9.88 min Scan# 938

Delta R.T. -0.02 min

Lab File: VF061989.D

Acq: 18 Mar 2019 10:55

Instrument :

MSVOA_F





ClientSampled :

Tot Ion: 91 Resp: 565835

Ion Ratio Lower Upper

91 100

106 31.3 24.4 36.6



#68

m/p-Xylenes

Concen: 103.291 ug/l

RT: 10.11 min Scan# 962

Delta R.T. -0.02 min

Lab File: VF061989.D




Acq: 18 Mar 2019 10:55


Tgt Ion: 106 Resp: 439727

Ion Ratio Lower Upper

106 100

91 188.0 151.9 227.9









#69
o-Xylene
Concen: 50.064 ug/l
RT: 10.69 min Scan# 1021
Delta R.T. -0.02 min
Lab File: VF061989.D
Acq: 18 Mar 2019 10:55

Instrument :
MSVOA_F
ClientSampleId :

Tot Ion:106 Resp: 228057
Ion Ratio Lower Upper
106 100
91 204.5 102.3 306.8




Abundance Ion 106.00 (105.70 to 106.70): V
Ion 91.00 (90.70 to 91.70): VF0






#70
Styrene
Concen: 51.020 ug/l
RT: 10.76 min Scan# 1028
Delta R.T. -0.02 min
Lab File: VF061989.D
Acq: 18 Mar 2019 10:55

Tgt Ion:104 Resp: 321980
Ion Ratio Lower Upper
104 100
78 43.1 34.7 52.1
103 44.7 36.6 54.8



Abundance Ion 104.00 (103.70 to 104.70): V
Ion 78.00 (77.70 to 78.70): VF0
Ion 103.00 (102.70 to 103.70): V





#71

Bromoform

Concen: 51.809 ug/l

RT: 10.75 min Scan# 1027

Delta R.T. -0.02 min

Lab File: VF061989.D

Acq: 18 Mar 2019 10:55

Instrument :

MSVOA_F

ClientSampled :


Tot Ion:173 Resp: 67150

Ion Ratio Lower Upper

173 100

175 47.7 23.7 71.1




254 8.8 7.8 11.8



Abundance Ion 173.00 (172.70 to 173.70): V

Ion 175.00 (174.70 to 175.70): V

Ion 254.00 (253.70 to 254.70): V



#72

1,4-Dichlorobenzene-d4

Concen: 50.000 ug/l

RT: 12.54 min Scan# 1208

Delta R.T. -0.01 min

Lab File: VF061989.D

Acq: 18 Mar 2019 10:55


Tgt Ion:152 Resp: 184857

Ion Ratio Lower Upper

152 100

115 50.0 27.3 81.9



150 161.8 0.0 316.2




Abundance Ion 152.00 (151.70 to 152.70): V

Ion 115.00 (114.70 to 115.70): V

Ion 150.00 (149.70 to 150.70): V





#73

Isopropylbenzene

Concen: 51.787 ug/l

RT: 11.11 min Scan# 1063

Delta R.T. -0.01 min

Lab File: VF061989.D

Acq: 18 Mar 2019 10:55

Instrument :

MSVOA_F


ClientSampled :

Tot Ion:105 Resp: 661375

Ion Ratio Lower Upper

105 100

120 28.0 13.8 41.3



Abundance Ion 105.00 (104.70 to 105.70): V

Ion 120.00 (119.70 to 120.70): V

11.11

400000



300000

200000

100000

0

Time--> 11.00 11.20



#74

N-ethyl acetate

Concen: 49.590 ug/l

RT: 11.36 min Scan# 1088

Delta R.T. -0.01 min

Lab File: VF061989.D

Acq: 18 Mar 2019 10:55

Tgt Ion: 43 Resp: 145114


Ion Ratio Lower Upper

43 100

70 39.3 31.9 47.9

55 21.2 17.9 26.9

61 16.9 22.2 33.2#



Abundance Ion 43.00 (42.70 to 43.70): VF0

Ion 70.00 (69.70 to 70.70): VF0

Ion 55.00 (54.70 to 55.70): VF0

Ion 61.00 (60.70 to 61.70): VF0


11.36


100000

50000

0

Time--> 11.30 11.40 11.50





#75

1,1,2,2-Tetrachloroethane

Concen: 47.706 ug/l

RT: 11.68 min Scan# 1121

Delta R.T. -0.01 min

Lab File: VF061989.D

Acq: 18 Mar 2019 10:55

Instrument :

MSVOA_F

ClientSampleId :


Tot Ion: 83 Resp: 100564

Ion Ratio Lower Upper

83 100

131 10.0 4.9 14.7




85 67.3 35.3 105.9



Abundance Ion 83.00 (82.70 to 83.70): VF0

Ion 131.00 (130.70 to 131.70): VF0

Ion 85.00 (84.70 to 85.70): VF0



#76

1,2,3-Trichloropropane

Concen: 45.746 ug/l

RT: 11.78 min Scan# 1131

Delta R.T. -0.01 min

Lab File: VF061989.D


Acq: 18 Mar 2019 10:55

Tgt Ion: 75 Resp: 67431

Ion Ratio Lower Upper



75 100


77 35.1 16.4 49.2



Abundance Ion 75.00 (74.70 to 75.70): VF0

Ion 77.00 (76.70 to 77.70): VF0





#77

Bromobenzene

Concen: 52.737 ug/l

RT: 11.47 min Scan# 1100

Delta R.T. -0.02 min

Lab File: VF061989.D

Acq: 18 Mar 2019 10:55

Instrument :

MSVOA_F

ClientSampleId :


Tot Ion:156 Resp: 155971

Ion Ratio Lower Upper

156 100

77 147.6 67.0 201.0




158 95.7 48.9 146.7



Abundance Ion 156.00 (155.70 to 156.70): V

Ion 77.00 (76.70 to 77.70): VF0

Ion 158.00 (157.70 to 158.70): V



#78

n-propylbenzene

Concen: 50.194 ug/l

RT: 11.58 min Scan# 1111

Delta R.T. -0.01 min

Lab File: VF061989.D


Acq: 18 Mar 2019 10:55

Tgt Ion: 91 Resp: 737097

Ion Ratio Lower Upper



91 100


120 23.8 11.8 35.3



Abundance Ion 91.00 (90.70 to 91.70): VF0

Ion 120.00 (119.70 to 120.70): V





#79

2-Chlorotoluene

Concen: 50.171 ug/l

RT: 11.70 min Scan# 1123

Delta R.T. -0.02 min

Lab File: VF061989.D

Acq: 18 Mar 2019 10:55

Instrument :

MSVOA_F


ClientSampled :

Tot Ion: 91 Resp: 415863

Ion Ratio Lower Upper




91 100

126 35.4 18.4 55.2



Abundance Ion 91.00 (90.70 to 91.70): VF0

Ion 126.00 (125.70 to 126.70): V



#80

1,3,5-Trimethylbenzene

Concen: 50.382 ug/l

RT: 11.81 min Scan# 1134

Delta R.T. -0.01 min

Lab File: VF061989.D


Acq: 18 Mar 2019 10:55

Tgt Ion: 105 Resp: 519751

Ion Ratio Lower Upper



105 100


120 55.0 26.6 79.7



Abundance Ion 105.00 (104.70 to 105.70): V

Ion 120.00 (119.70 to 120.70): V





#81

trans-1,4-Dichloro-2-butene

Concen: 69.187 ug/l

RT: 11.85 min Scan# 1138

Delta R.T. -0.02 min

Lab File: VF061989.D

Acq: 18 Mar 2019 10:55

Instrument :

MSVOA_F

ClientSampleId :





Tot Ion: 75 Resp: 45117

Ion Ratio Lower Upper

75 100

53 63.3 49.0 73.6

89 48.9 37.1 55.7



#82

4-Chlorotoluene

Concen: 48.539 ug/l

RT: 11.88 min Scan# 1141

Delta R.T. -0.02 min

Lab File: VF061989.D




Acq: 18 Mar 2019 10:55


Tgt Ion: 91 Resp: 406863

Ion Ratio Lower Upper

91 100

126 33.7 18.1 54.1









#83
 tert-Butylbenzene
 Concen: 52.565 ug/l
 RT: 12.11 min Scan# 1165
 Delta R.T. -0.01 min
 Lab File: VF061989.D
 Acq: 18 Mar 2019 10:55

Instrument :
 MSVOA_F
 ClientSampleId :

Tot Ion:119 Resp: 567115
 Ion Ratio Lower Upper
 119 100
 91 55.5 27.9 83.6
 134 25.9 12.3 36.8




Abundance Ion 119.00 (118.70 to 119.70): V
 Ion 91.00 (90.70 to 91.70): VF0
 Ion 134.00 (133.70 to 134.70): V






#84
 1,2,4-Trimethylbenzene
 Concen: 49.125 ug/l
 RT: 12.18 min Scan# 1172
 Delta R.T. -0.02 min
 Lab File: VF061989.D
 Acq: 18 Mar 2019 10:55

Tgt Ion:105 Resp: 507618
 Ion Ratio Lower Upper
 105 100
 120 49.6 25.9 77.8



Abundance Ion 105.00 (104.70 to 105.70): V
 Ion 120.00 (119.70 to 120.70): V





#85

sec-Butylbenzene

Concen: 49.100 ug/l

RT: 12.29 min Scan# 1183

Delta R.T. -0.01 min

Lab File: VF061989.D

Acq: 18 Mar 2019 10:55

Instrument :

MSVOA_F


ClientSampleId :

Tot Ion:105 Resp: 715719

Ion Ratio Lower Upper




105 100

134 21.0 10.0 29.8



Abundance Ion 105.00 (104.70 to 105.70): V

Ion 134.00 (133.70 to 134.70): V



#86

p-Isopropyltoluene

Concen: 51.937 ug/l

RT: 12.44 min Scan# 1198

Delta R.T. -0.02 min

Lab File: VF061989.D

Acq: 18 Mar 2019 10:55


Tgt Ion:119 Resp: 631721

Ion Ratio Lower Upper

119 100

134 25.8 13.9 41.7



91 22.3 10.4 31.1




Abundance Ion 119.00 (118.70 to 119.70): V

Ion 134.00 (133.70 to 134.70): V

Ion 91.00 (90.70 to 91.70): V









#87
 1,3-Dichlorobenzene
 Concen: 49.788 ug/l
 RT: 12.46 min Scan# 1200
 Delta R.T. -0.01 min
 Lab File: VF061989.D
 Acq: 18 Mar 2019 10:55




Instrument :
 MSVOA_F
 ClientSampleId :


Tot Ion:146 Resp: 289403
 Ion Ratio Lower Upper
 146 100
 111 35.1 18.6 55.8
 148 63.9 31.3 93.9



#88
 1,4-Dichlorobenzene
 Concen: 51.717 ug/l
 RT: 12.55 min Scan# 1209
 Delta R.T. -0.01 min
 Lab File: VF061989.D
 Acq: 18 Mar 2019 10:55

Tgt Ion:146 Resp: 289347
 Ion Ratio Lower Upper
 146 100
 111 35.5 19.4 58.1
 148 64.0 32.4 97.0





#89

n-Butylbenzene

Concen: 51.192 ug/l

RT: 12.82 min Scan# 1237

Delta R.T. -0.02 min

Lab File: VF061989.D

Acq: 18 Mar 2019 10:55

Instrument :

MSVOA_F

ClientSampleId :





Tot Ion: 91 Resp: 585092

Ion Ratio Lower Upper

91 100

92 63.1 31.7 95.1

134 25.7 15.1 45.3



#90

Hexachloroethane

Concen: 56.078 ug/l

RT: 12.90 min Scan# 1245

Delta R.T. -0.01 min

Lab File: VF061989.D




Acq: 18 Mar 2019 10:55


Tgt Ion: 117 Resp: 155660

Ion Ratio Lower Upper

117 100

201 73.6 32.5 97.5





#91

1,2-Dichlorobenzene

Concen: 50.236 ug/l

RT: 12.92 min Scan# 1247

Delta R.T. -0.02 min

Lab File: VF061989.D

Acq: 18 Mar 2019 10:55

Instrument :

MSVOA_F

ClientSampleId :


Tot Ion:146 Resp: 270456

Ion Ratio Lower Upper

146 100

111 40.9 18.9 56.5




148 64.8 32.9 98.6



Abundance Ion 146.00 (145.70 to 146.70): V

Ion 111.00 (110.70 to 111.70): V

Ion 148.00 (147.70 to 148.70): V



#92

1,2-Dibromo-3-Chloropropane

Concen: 50.494 ug/l

RT: 13.62 min Scan# 1318

Delta R.T. -0.02 min

Lab File: VF061989.D

Acq: 18 Mar 2019 10:55


Tgt Ion: 75 Resp: 17081

Ion Ratio Lower Upper

75 100

155 94.5 65.0 195.1



157 120.4 84.7 254.0




Abundance Ion 75.00 (74.70 to 75.70): VF0

Ion 155.00 (154.70 to 155.70): V

Ion 157.00 (156.70 to 157.70): V









#93
 1,2,4-Trichlorobenzene
 Concen: 50.865 ug/l
 RT: 14.19 min Scan# 1375
 Delta R.T. -0.01 min
 Lab File: VF061989.D
 Acq: 18 Mar 2019 10:55

Instrument :
 MSVOA_F
 ClientSampleId :

Tot Ion:180 Resp: 190692
 Ion Ratio Lower Upper
 180 100
 182 94.3 49.6 149.0
 145 32.4 17.5 52.5




Abundance Ion 180.00 (179.70 to 180.70): V
 Ion 182.00 (181.70 to 182.70): V
 Ion 145.00 (144.70 to 145.70): V






#94
 Hexachlorobutadiene
 Concen: 51.868 ug/l
 RT: 14.18 min Scan# 1374
 Delta R.T. -0.01 min
 Lab File: VF061989.D
 Acq: 18 Mar 2019 10:55

Tgt Ion:225 Resp: 129092
 Ion Ratio Lower Upper
 225 100
 223 61.8 29.9 89.7
 227 66.5 32.4 97.2



Abundance Ion 225.00 (224.70 to 225.70): V
 Ion 223.00 (222.70 to 223.70): V
 Ion 227.00 (226.70 to 227.70): V





#95

Naphthalene

Concen: 52.060 ug/l

RT: 14.44 min Scan# 1401

Delta R.T. -0.01 min

Lab File: VF061989.D

Acq: 18 Mar 2019 10:55

Instrument :

MSVOA_F

ClientSampleId :


Tot Ion:128 Resp: 351028

Ion Ratio Lower Upper

128 100

127 12.0 9.3 13.9




129 11.3 9.1 13.7



Abundance Ion 128.00 (127.70 to 128.70): V

Ion 127.00 (126.70 to 127.70): V

Ion 129.00 (128.70 to 129.70): V



#96

1,2,3-Trichlorobenzene

Concen: 51.133 ug/l

RT: 14.59 min Scan# 1416

Delta R.T. -0.01 min

Lab File: VF061989.D

Acq: 18 Mar 2019 10:55


Tgt Ion:180 Resp: 176363

Ion Ratio Lower Upper

180 100

182 99.5 48.9 146.8

145 30.2 16.7 50.1



Abundance Ion 180.00 (179.70 to 180.70): V

Ion 182.00 (181.70 to 182.70): V

Ion 145.00 (144.70 to 145.70): V

