

Data Path : Z:\VOASRV\HPCHEM1\MSVOA F\DATA\VF031819\
 Data File : VF061990.D
 Acq On : 18 Mar 2019 11:37
 Operator : VA/AP
 Sample : VF0318SBL01
 Misc : 5.00g/5mL/MSVOA-F/SOIL
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_F
 ClientSampleId :
 VF0318SBL01

Quant Time: Mar 19 01:33:55 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_F\METHODS\82F031619S.M
 Quant Title : SW846 8260
 QLast Update : Mon Mar 18 07:28:40 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	4.85	168	297448	50.00	ug/l	-0.02
34) 1,4-Difluorobenzene	5.57	114	445848	50.00	ug/l	-0.03
63) Chlorobenzene-d5	9.75	117	366134	50.00	ug/l	-0.02
72) 1,4-Dichlorobenzene-d4	12.53	152	218059	50.00	ug/l	-0.02

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	4.82	65	111112	51.89	ug/l	-0.02
Spiked Amount				50.000		
			Recovery	=	103.78%	
35) Dibromofluoromethane	4.09	113	151722	50.68	ug/l	-0.02
Spiked Amount				50.000		
			Recovery	=	101.36%	
50) Toluene-d8	7.53	98	399739	45.67	ug/l	-0.02
Spiked Amount				50.000		
			Recovery	=	91.34%	
62) 4-Bromofluorobenzene	11.40	95	170736	49.11	ug/l	0.00
Spiked Amount				50.000		
			Recovery	=	98.22%	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA F\DATA\VF031819\
Data File : VF061990.D
Acq On : 18 Mar 2019 11:37
Operator : VA/AP
Sample : VF0318SBL01
Misc : 5.00µ/5mL/MSVOA-F/SOIL
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_F
ClientSampleId :
VF0318SBL01

Quant Time: Mar 19 01:33:55 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_F\METHODS\82F031619S.M
Quant Title : SW846 8260
QLast Update : Mon Mar 18 07:28:40 2019
Response via : Initial Calibration

