

Data Path : Z:\VOASRV\HPCHEM1\MSVOA F\DATA\VF031919\
 Data File : VF062017.D
 Acq On : 19 Mar 2019 14:45
 Operator : VA/AP
 Sample : K1951-02RE
 Misc : 5.76g/5mL/MSVOA-F/SOIL
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_F
 ClientSampleId :
 FK-GF-03-006-ARE

Quant Time: Mar 20 05:20:55 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_F\METHODS\82F031619S.M
 Quant Title : SW846 8260
 QLast Update : Mon Mar 18 07:28:40 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	0.00	168	0	0.00	ug/l	-4.87
34) 1,4-Difluorobenzene	5.58	114	232	50.00	ug/l	-0.02
63) Chlorobenzene-d5	0.00	117	0	0.00	ug/l	-9.77
72) 1,4-Dichlorobenzene-d4	0.00	152	0	0.00	ug/l	-12.55

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	0.00	65	0	0.00	ug/l	
Spiked Amount				50.000		
			Recovery	=	0.00%	
35) Dibromofluoromethane	0.00	113	0	0.00	ug/l	
Spiked Amount				50.000		
			Recovery	=	0.00%	
50) Toluene-d8	7.55	98	393	86.29	ug/l	0.00
Spiked Amount				50.000		
			Recovery	=	172.58%	
62) 4-Bromofluorobenzene	0.00	95	0	0.00	ug/l	
Spiked Amount				50.000		
			Recovery	=	0.00%	

Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
37) Ethyl Acetate	4.08	43	524	557.027	ug/l #	19
41) Methacrylonitrile	4.73	41	70	139.276	ug/l #	22
43) Isopropyl Acetate	6.94	43	156	140.161	ug/l #	38
47) Bromodichloromethane	6.29	83	66	35.820	ug/l #	25
48) Methyl methacrylate	6.66	41	286	397.592	ug/l #	20
51) 4-Methyl-2-Pentanone	8.22	43	179	226.479	ug/l #	36
55) 1,1,2-Trichloroethane	8.54	97	78	82.393	ug/l #	14
57) 1,3-Dichloropropane	8.80	76	73	46.227	ug/l #	43
58) 2-Chloroethyl Vinyl ether	7.33	63	73	374.510	ug/l #	49
59) 2-Hexanone	9.46	43	72	125.877	ug/l #	23
61) 1,2-Dibromoethane	8.81	107	79	77.797	ug/l #	4

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA F\DATA\VF031919\
Data File : VF062017.D
Acq On : 19 Mar 2019 14:45
Operator : VA/AP
Sample : K1951-02RE
Misc : 5.76g/5mL/MSVOA-F/SOIL
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_F
ClientSampleId :
FK-GF-03-006-ARE

Quant Time: Mar 20 05:20:55 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_F\METHODS\82F031619S.M
Quant Title : SW846 8260
QLast Update : Mon Mar 18 07:28:40 2019
Response via : Initial Calibration

