

Data Path : Z:\VOASRV\HPCHEM1\MSVOA F\DATA\VF031919\
 Data File : VF062029.D
 Acq On : 19 Mar 2019 20:44
 Operator : VA/AP
 Sample : K1971-04
 Misc : 5.56g/5mL/MSVOA-F/SOIL
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 MSVOA_F
 ClientSampleId :
 SB-9(7-8)-031519

Quant Time: Mar 20 06:42:09 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_F\METHODS\82F031619S.M
 Quant Title : SW846 8260
 QLast Update : Mon Mar 18 07:28:40 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	4.84	168	76696	50.00	ug/l	-0.03
34) 1,4-Difluorobenzene	5.57	114	112954	50.00	ug/l	-0.03
63) Chlorobenzene-d5	9.75	117	81584	50.00	ug/l	-0.02
72) 1,4-Dichlorobenzene-d4	12.54	152	28250	50.00	ug/l	-0.02
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	4.82	65	32580	59.01	ug/l	-0.02
Spiked Amount 50.000			Recovery	=	118.02%	
35) Dibromofluoromethane	4.09	113	46992	61.95	ug/l	-0.02
Spiked Amount 50.000			Recovery	=	123.90%	
50) Toluene-d8	7.53	98	102353	46.16	ug/l	-0.03
Spiked Amount 50.000			Recovery	=	92.32%	
62) 4-Bromofluorobenzene	11.39	95	32439	36.83	ug/l	-0.02
Spiked Amount 50.000			Recovery	=	73.66%	
Target Compounds						
44) Trichloroethene	5.48	130	12035	14.535	ug/l	Qvalue 92

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA F\DATA\VF031919\
Data File : VF062029.D
Acq On : 19 Mar 2019 20:44
Operator : VA/AP
Sample : K1971-04
Misc : 5.56g/5mL/MSVOA-F/SOIL
ALS Vial : 21 Sample Multiplier: 1

Instrument :
MSVOA_F
ClientSampleId :
SB-9(7-8)-031519

Quant Time: Mar 20 06:42:09 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_F\METHODS\82F031619S.M
Quant Title : SW846 8260
QLast Update : Mon Mar 18 07:28:40 2019
Response via : Initial Calibration

