


Data Path : Z:\voasrv\HPCHEM1\MSVOA F\Data\VF031919\
 Data File : VF062009.D
 Acq On : 19 Mar 2019 9:47
 Operator : VA/AP
 Sample : BFB
 Misc : 5.00µ/5mL/MSVOA-F/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_F
 ClientSampleId :
 BFB

Integration File: RTEINT.P

Method : Z:\VOASRV\HPCHEM1\MSVOA_F\METHODS\82F031619S.M
 Title : SW846 8260
 Last Update : Mon Mar 18 07:28:40 2019



AutoFind: Scans 1090, 1091, 1092; Background Corrected with Scan 1084

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	15.6	17365	PASS
75	95	30	60	39.5	44016	PASS
95	95	100	100	100.0	111437	PASS
96	95	5	9	6.6	7380	PASS
173	174	0.00	2	0.2	215	PASS
174	95	50	100	83.3	92821	PASS
175	174	5	9	7.2	6671	PASS
176	174	95	101	97.5	90477	PASS
177	176	5	9	6.9	6238	PASS