

Data Path : Z:\VOASRV\HPCHEM1\MSVOA F\DATA\VF032019\
 Data File : VF062048.D
 Acq On : 20 Mar 2019 19:03
 Operator : VA/AP
 Sample : K1971-10RE
 Misc : 5.31g/5mL/MSVOA-F/SOIL
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_F
 ClientSampleId :
 SB-15(6-7)-031519RE

Manual Integrations
 APPROVED

MMDadoda
 3/22/2019 4:52:32 PM

Quant Time: Mar 22 07:38:51 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_F\METHODS\82F031619S.M
 Quant Title : SW846 8260
 QLast Update : Mon Mar 18 07:28:40 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	4.86	168	106104	50.00	ug/l	-0.01
34) 1,4-Difluorobenzene	5.59	114	154633	50.00	ug/l	-0.01
63) Chlorobenzene-d5	9.76	117	44007	50.00	ug/l	-0.01
72) 1,4-Dichlorobenzene-d4	12.55	152	8955	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	4.83	65	62125	81.34	ug/l	-0.01
Spiked Amount 50.000			Recovery	=	162.68%	
35) Dibromofluoromethane	4.10	113	77209	74.35	ug/l	-0.01
Spiked Amount 50.000			Recovery	=	148.70%	
50) Toluene-d8	7.55	98	328206	108.12	ug/l	0.00
Spiked Amount 50.000			Recovery	=	216.24%	
62) 4-Bromofluorobenzene	11.38	95	62767	52.05	ug/l	-0.03
Spiked Amount 50.000			Recovery	=	104.10%	
Target Compounds						
16) Acetone	2.23	43	10188	59.402	ug/l	100
20) Methylene Chloride	2.18	84	11556m	13.087	ug/l	
83) tert-Butylbenzene	12.13	119	40751	77.971	ug/l	81

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA F\DATA\VF032019\
Data File : VF062048.D
Acq On : 20 Mar 2019 19:03
Operator : VA/AP
Sample : K1971-10RE
Misc : 5.31g/5mL/MSVOA-F/SOIL
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_F
ClientSampleId :
SB-15(6-7)-031519RE

Manual Integrations
APPROVED

MMDadoda
3/22/2019 4:52:32 PM

Quant Time: Mar 22 07:38:51 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_F\METHODS\82F031619S.M
Quant Title : SW846 8260
QLast Update : Mon Mar 18 07:28:40 2019
Response via : Initial Calibration

