

Data Path : Z:\VOASRV\HPCHEM1\MSVOA F\DATA\VF032219\
 Data File : VF062101.D
 Acq On : 23 Mar 2019 2:29
 Operator : VA/AP
 Sample : K2040-07
 Misc : 4.52g/5mL/MSVOA-F/SOIL
 ALS Vial : 34 Sample Multiplier: 1

Instrument :
 MSVOA_F
 ClientSampleId :
 S4B(10)

Manual Integrations
 APPROVED

MMDadoda
 3/27/2019 11:57:27 AM

Quant Time: Mar 23 06:57:51 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_F\METHODS\82F032219S.M
 Quant Title : SW846 8260
 QLast Update : Sat Mar 23 01:16:37 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	4.89	168	240118	50.00	ug/l	-0.01
34) 1,4-Difluorobenzene	5.61	114	390485	50.00	ug/l	-0.01
63) Chlorobenzene-d5	9.79	117	323820	50.00	ug/l	-0.01
72) 1,4-Dichlorobenzene-d4	12.56	152	173048	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	4.86	65	80557	45.10	ug/l	-0.01
Spiked Amount 50.000			Recovery	=	90.20%	
35) Dibromofluoromethane	4.13	113	100577	38.08	ug/l	-0.01
Spiked Amount 50.000			Recovery	=	76.16%	
50) Toluene-d8	7.58	98	324590	41.17	ug/l	-0.02
Spiked Amount 50.000			Recovery	=	82.34%	
62) 4-Bromofluorobenzene	11.42	95	128990	42.38	ug/l	-0.01
Spiked Amount 50.000			Recovery	=	84.76%	
Target Compounds						
16) Acetone	2.25	43	19823	52.735	ug/l	92
20) Methylene Chloride	2.19	84	23012m	7.448	ug/l	

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA F\DATA\VF032219\
Data File : VF062101.D
Acq On : 23 Mar 2019 2:29
Operator : VA/AP
Sample : K2040-07
Misc : 4.52g/5mL/MSVOA-F/SOIL
ALS Vial : 34 Sample Multiplier: 1

Instrument :
MSVOA_F
ClientSampleId :
S4B(10)

Manual Integrations
APPROVED

MMDadoda
3/27/2019 11:57:27 AM

Quant Time: Mar 23 06:57:51 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_F\METHODS\82F032219S.M
Quant Title : SW846 8260
QLast Update : Sat Mar 23 01:16:37 2019
Response via : Initial Calibration

