

Data Path : Z:\VOASRV\HPCHEM1\MSVOA F\DATA\VF032219\
 Data File : VF062104.D
 Acq On : 23 Mar 2019 3:56
 Operator : VA/AP
 Sample : K2061-01
 Misc : 6.36g/5mL/MSVOA-F/SOIL
 ALS Vial : 37 Sample Multiplier: 1

Instrument :
 MSVOA_F
 ClientSampleId :
 26524-27245

Manual Integrations
 APPROVED

MMDadoda
 3/27/2019 11:57:29 AM

Quant Time: Mar 23 07:01:06 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_F\METHODS\82F032219S.M
 Quant Title : SW846 8260
 QLast Update : Sat Mar 23 01:16:37 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	4.89	168	232752	50.00	ug/l	-0.02
34) 1,4-Difluorobenzene	5.62	114	372965	50.00	ug/l	-0.01
63) Chlorobenzene-d5	9.79	117	280468	50.00	ug/l	-0.02
72) 1,4-Dichlorobenzene-d4	12.56	152	148371	50.00	ug/l	-0.01
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	4.86	65	90310	52.17	ug/l	-0.02
Spiked Amount 50.000			Recovery	=	104.34%	
35) Dibromofluoromethane	4.13	113	93978	37.25	ug/l	-0.02
Spiked Amount 50.000			Recovery	=	74.50%	
50) Toluene-d8	7.57	98	330277	43.86	ug/l	-0.02
Spiked Amount 50.000			Recovery	=	87.72%	
62) 4-Bromofluorobenzene	11.42	95	131678	45.29	ug/l	-0.02
Spiked Amount 50.000			Recovery	=	90.58%	
Target Compounds						
16) Acetone	2.26	43	27787	76.262	ug/l	97
20) Methylene Chloride	2.19	84	28906m	10.668	ug/l	

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA F\DATA\VF032219\
Data File : VF062104.D
Acq On : 23 Mar 2019 3:56
Operator : VA/AP
Sample : K2061-01
Misc : 6.36g/5mL/MSVOA-F/SOIL
ALS Vial : 37 Sample Multiplier: 1

Instrument :
MSVOA_F
ClientSampleId :
26524-27245

Manual Integrations
APPROVED

MMDadoda
3/27/2019 11:57:29 AM

Quant Time: Mar 23 07:01:06 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_F\METHODS\82F032219S.M
Quant Title : SW846 8260
QLast Update : Sat Mar 23 01:16:37 2019
Response via : Initial Calibration

