

Data Path : Z:\VOASRV\HPCHEM1\MSVOA F\DATA\VF032319\
 Data File : VF062121.D
 Acq On : 23 Mar 2019 17:05
 Operator : VA/AP
 Sample : K2045-09
 Misc : 6.29g/5mL/MSVOA-F/SOIL
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_F
 ClientSampleId :

Quant Time: Mar 25 01:29:24 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_F\METHODS\82F032219S.M
 Quant Title : SW846 8260
 QLast Update : Sat Mar 23 01:16:37 2019
 Response via : Initial Calibration


Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	4.88	168	221791	50.00	ug/l	-0.03
34) 1,4-Difluorobenzene	5.61	114	341829	50.00	ug/l	-0.02
63) Chlorobenzene-d5	9.78	117	266500	50.00	ug/l	-0.03
72) 1,4-Dichlorobenzene-d4	12.55	152	112247	50.00	ug/l	-0.02
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	4.85	65	82813	50.20	ug/l	-0.03
Spiked Amount 50.000			Recovery	=	100.40%	
35) Dibromofluoromethane	4.12	113	111616	48.27	ug/l	-0.03
Spiked Amount 50.000			Recovery	=	96.54%	
50) Toluene-d8	7.56	98	277730	40.24	ug/l	-0.03
Spiked Amount 50.000			Recovery	=	80.48%	
62) 4-Bromofluorobenzene	11.41	95	103932	39.00	ug/l	-0.02
Spiked Amount 50.000			Recovery	=	78.00%	
Target Compounds						
					Qvalue	
4) Vinyl Chloride	1.14	62	5050	1.836	ug/l	92
11) Tert butyl alcohol	2.57	59	759	8.362	ug/l #	66
13) Acrolein	2.02	56	542	4.966	ug/l	96
16) Acetone	2.24	43	19626	56.526	ug/l	93
17) Carbon Disulfide	1.82	76	8569	1.314	ug/l #	54
18) Methyl Acetate	2.38	43	3695	4.461	ug/l #	56
20) Methylene Chloride	2.19	84	2826	Below Cal	#	79
25) 2-Butanone	4.35	43	2013	2.777	ug/l #	55
27) cis-1,2-Dichloroethene	3.48	96	39001	14.308	ug/l	92
29) Tetrahydrofuran	4.07	42	1064	3.376	ug/l #	38
49) 1,4-Dioxane	6.45	88	61	7.085	ug/l #	25
51) 4-Methyl-2-Pentanone	8.29	43	8291	7.188	ug/l #	40
59) 2-Hexanone	9.50	43	885	1.047	ug/l #	22
74) N-amyl acetate	11.40	43	3153	1.675	ug/l #	46


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA F\DATA\VF032319\
Data File : VF062121.D
Acq On : 23 Mar 2019 17:05
Operator : VA/AP
Sample : K2045-09
Misc : 6.29g/5mL/MSVOA-F/SOIL
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_F
ClientSampleId :

Quant Time: Mar 25 01:29:24 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_F\METHODS\82F032219S.M
Quant Title : SW846 8260
QLast Update : Sat Mar 23 01:16:37 2019
Response via : Initial Calibration





#1

Pentafluorobenzene

Concen: 50.000 ug/l

RT: 4.88 min Scan# 431

Delta R.T. -0.03 min

Lab File: VF062121.D

Acq: 23 Mar 2019 17:05


Instrument :
MSVOA_F
ClientSampleId :

Tot Ion:168 Resp: 221791

Ion Ratio Lower Upper

168 100

99 49.3 38.2 57.4



Abundance Ion 168.00 (167.70 to 168.70): VF0

Ion 99.00 (98.70 to 99.70): VF0

4.88

80000

60000



40000

20000

0

Time-->

4.70 4.80 4.90 5.00 5.10



#4

Vinyl Chloride

Concen: 1.836 ug/l

RT: 1.14 min Scan# 52

Delta R.T. -0.01 min

Lab File: VF062121.D


Acq: 23 Mar 2019 17:05

Tgt Ion: 62 Resp: 5050

Ion Ratio Lower Upper

62 100

64 28.1 26.2 39.2



Abundance Ion 62.00 (61.70 to 62.70): VF0

Ion 64.00 (63.70 to 64.70): VF0

1.14

1500


1000


500

0

Time-->

1.00 1.10 1.20 1.30









#11

Tert butyl alcohol
Concen: 8.362 ug/l
RT: 2.57 min Scan# 197
Delta R.T. -0.02 min
Lab File: VF062121.D
Acq: 23 Mar 2019 17:05

Instrument :
MSVOA_F
ClientSampleId :




Tot Ion: 59 Resp: 759
Ion Ratio Lower Upper
59 100
57 0.0 11.0 16.4#




#13

Acrolein
Concen: 4.966 ug/l
RT: 2.02 min Scan# 141
Delta R.T. -0.00 min
Lab File: VF062121.D
Acq: 23 Mar 2019 17:05

Tgt Ion: 56 Resp: 542
Ion Ratio Lower Upper
56 100
55 58.1 49.0 73.4





#16

Acetone

Concen: 56.526 ug/l

RT: 2.24 min Scan# 164

Delta R.T. -0.01 min

Lab File: VF062121.D

Acq: 23 Mar 2019 17:05





Instrument :
MSVOA_F
ClientSampleId :

Tot Ion: 43 Resp: 19626

Ion Ratio Lower Upper

43 100

58 22.4 20.8 31.2



#17

Carbon Disulfide

Concen: 1.314 ug/l

RT: 1.82 min Scan# 121

Delta R.T. -0.03 min

Lab File: VF062121.D




Acq: 23 Mar 2019 17:05


Tgt Ion: 76 Resp: 8569

Ion Ratio Lower Upper

76 100

78 27.1 7.9 11.9#









#18
Methyl Acetate
Concen: 4.461 ug/l
RT: 2.38 min Scan# 178
Delta R.T. 0.02 min
Lab File: VF062121.D
Acq: 23 Mar 2019 17:05




Instrument :
MSVOA_F
ClientSampled :


Tot Ion: 43 Resp: 3695
Ion Ratio Lower Upper
43 100
74 0.0 16.3 24.5#



#20
Methylene Chloride
Concen: Below Cal
RT: 2.19 min Scan# 158
Delta R.T. -0.01 min
Lab File: VF062121.D
Acq: 23 Mar 2019 17:05

Tgt Ion: 84 Resp: 2826
Ion Ratio Lower Upper
84 100
49 146.8 100.4 150.6
51 68.2 31.0 46.6#
86 71.7 51.3 76.9





#25

2-Butanone

Concen: 2.777 ug/l

RT: 4.35 min Scan# 378

Delta R.T. -0.01 min

Lab File: VF062121.D

Acq: 23 Mar 2019 17:05





Instrument :
MSVOA_F
ClientSampleId :

Tot Ion: 43 Resp: 2013

Ion Ratio Lower Upper

43 100

72 0.0 16.7 25.1#



#27

cis-1,2-Dichloroethene

Concen: 14.308 ug/l

RT: 3.48 min Scan# 289

Delta R.T. -0.03 min

Lab File: VF062121.D

Acq: 23 Mar 2019 17:05




Tgt Ion: 96 Resp: 39001


Ion Ratio Lower Upper

96 100

61 129.3 0.0 239.0

98 68.9 0.0 128.2









#29
Tetrahydrofuran
Concen: 3.376 ug/l
RT: 4.07 min Scan# 349
Delta R.T. -0.04 min
Lab File: VF062121.D
Acq: 23 Mar 2019 17:05




Instrument :
MSVOA_F
ClientSampled :


Tot Ion: 42 Resp: 1064
Ion Ratio Lower Upper
42 100
72 0.0 33.3 49.9#
71 6.0 35.0 52.4#



#33
1,2-Dichloroethane-d4
Concen: 50.199 ug/l
RT: 4.85 min Scan# 428
Delta R.T. -0.03 min
Lab File: VF062121.D
Acq: 23 Mar 2019 17:05

Tgt Ion: 65 Resp: 82813
Ion Ratio Lower Upper
65 100
67 51.2 0.0 106.2









#34
1,4-Difluorobenzene
Concen: 50.000 ug/l
RT: 5.61 min Scan# 505
Delta R.T. -0.02 min
Lab File: VF062121.D
Acq: 23 Mar 2019 17:05

Instrument :
MSVOA_F
ClientSampleId :

Tot Ion:114 Resp: 341829
Ion Ratio Lower Upper
114 100
63 17.3 0.0 37.8
88 18.6 0.0 36.4




Abundance Ion 114.00 (113.70 to 114.70): V
Ion 63.00 (62.70 to 63.70): VF0
Ion 88.00 (87.70 to 88.70): VF0






#35
Dibromofluoromethane
Concen: 48.270 ug/l
RT: 4.12 min Scan# 354
Delta R.T. -0.03 min
Lab File: VF062121.D
Acq: 23 Mar 2019 17:05

Tgt Ion:113 Resp: 111616
Ion Ratio Lower Upper
113 100
111 111.3 75.4 113.0
192 18.1 12.3 18.5



Abundance Ion 113.00 (112.70 to 113.70): V
Ion 111.00 (110.70 to 111.70): V
Ion 192.00 (191.70 to 192.70): V









#49
1,4-Dioxane
Concen: 7.085 ug/l
RT: 6.45 min Scan# 590
Delta R.T. -0.30 min
Lab File: VF062121.D
Acq: 23 Mar 2019 17:05

Instrument :
MSVOA_F
ClientSampleId :

Tot Ion: 88 Resp: 61
Ion Ratio Lower Upper
88 100
43 0.0 35.0 52.4#
58 0.0 49.9 74.9#




Abundance Ion 88.00 (87.70 to 88.70): VF0
Ion 43.00 (42.70 to 43.70): VF0
Ion 58.00 (57.70 to 58.70): VF0






#50
Toluene-d8
Concen: 40.237 ug/l
RT: 7.56 min Scan# 703
Delta R.T. -0.03 min
Lab File: VF062121.D
Acq: 23 Mar 2019 17:05

Tgt Ion: 98 Resp: 277730
Ion Ratio Lower Upper
98 100
100 69.0 57.0 85.6



Abundance Ion 98.00 (97.70 to 98.70): VF0
Ion 100.00 (99.70 to 100.70): VF0





#51

4-Methyl-2-Pentanone

Concen: 7.188 ug/l

RT: 8.29 min Scan# 777

Delta R.T. -0.00 min

Lab File: VF062121.D

Acq: 23 Mar 2019 17:05

Instrument :

MSVOA_F


ClientSampleId :

Tot Ion: 43 Resp: 8291

Ion Ratio Lower Upper




43 100

58 4.4 33.8 50.8#



Abundance Ion 43.00 (42.70 to 43.70): VF0

Ion 58.00 (57.70 to 58.70): VF0



#59

2-Hexanone

Concen: 1.047 ug/l

RT: 9.50 min Scan# 900

Delta R.T. -0.02 min

Lab File: VF062121.D


Acq: 23 Mar 2019 17:05

Tgt Ion: 43 Resp: 885

Ion Ratio Lower Upper



43 100


58 0.0 28.8 86.4#



Abundance Ion 43.00 (42.70 to 43.70): VF0

Ion 58.00 (57.70 to 58.70): VF0









#62
4-Bromofluorobenzene
Concen: 39.003 ug/l
RT: 11.41 min Scan# 1094
Delta R.T. -0.02 min
Lab File: VF062121.D
Acq: 23 Mar 2019 17:05

Instrument :
MSVOA_F
ClientSampleId :

Tot Ion: 95 Resp: 103932
Ion Ratio Lower Upper
95 100
174 85.9 0.0 168.2
176 85.2 0.0 164.2




Abundance Ion 95.00 (94.70 to 95.70): VF0
Ion 174.00 (173.70 to 174.70): V
Ion 176.00 (175.70 to 176.70): V






#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 9.78 min Scan# 928
Delta R.T. -0.03 min
Lab File: VF062121.D
Acq: 23 Mar 2019 17:05

Tgt Ion: 117 Resp: 266500
Ion Ratio Lower Upper
117 100
82 50.2 39.2 58.8
119 33.7 25.4 38.2



Abundance Ion 117.00 (116.70 to 117.70): V
Ion 82.00 (81.70 to 82.70): VF0
Ion 119.00 (118.70 to 119.70): V









#72
 1,4-Dichlorobenzene-d4
 Concen: 50.000 ug/l
 RT: 12.55 min Scan# 1209
 Delta R.T. -0.02 min
 Lab File: VF062121.D
 Acq: 23 Mar 2019 17:05

Instrument :
 MSVOA_F
 ClientSampled :

Tot Ion:152 Resp: 112247
 Ion Ratio Lower Upper
 152 100
 115 54.3 27.7 83.1
 150 153.5 0.0 313.2



#74
 N-ethyl acetate
 Concen: 1.675 ug/l
 RT: 11.40 min Scan# 1092
 Delta R.T. -0.00 min
 Lab File: VF062121.D
 Acq: 23 Mar 2019 17:05

Tgt Ion: 43 Resp: 3153
 Ion Ratio Lower Upper
 43 100
 70 64.9 32.9 49.3#
 55 58.3 17.1 25.7#
 61 58.6 22.2 33.2#

