

Data Path : Z:\VOASRV\HPCHEM1\MSVOA F\DATA\VF032519\
 Data File : VF062138.D
 Acq On : 25 Mar 2019 18:18
 Operator : VA/AP
 Sample : K2086-01
 Misc : 6.40g/5mL/MSVOA-F/SOIL
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_F
 ClientSampleId :
 CS-11

Manual Integrations
 APPROVED

MMDadoda
 3/26/2019 3:59:46 PM

Quant Time: Mar 26 05:39:06 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_F\METHODS\82F032219S.M
 Quant Title : SW846 8260
 QLast Update : Sat Mar 23 01:16:37 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	4.88	168	122455	50.00	ug/l	-0.03
34) 1,4-Difluorobenzene	5.60	114	181863	50.00	ug/l	-0.03
63) Chlorobenzene-d5	9.78	117	129001	50.00	ug/l	-0.03
72) 1,4-Dichlorobenzene-d4	12.55	152	50326	50.00	ug/l	-0.02
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	4.84	65	34009	37.34	ug/l	-0.04
Spiked Amount 50.000			Recovery	=	74.68%	
35) Dibromofluoromethane	4.12	113	54858	44.59	ug/l	-0.03
Spiked Amount 50.000			Recovery	=	89.18%	
50) Toluene-d8	7.56	98	132762	36.15	ug/l	-0.03
Spiked Amount 50.000			Recovery	=	72.30%	
62) 4-Bromofluorobenzene	11.40	95	46109	32.52	ug/l	-0.03
Spiked Amount 50.000			Recovery	=	65.04%	
Target Compounds						
20) Methylene Chloride	2.17	84	6347m	2.452	ug/l	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA F\DATA\VF032519\
Data File : VF062138.D
Acq On : 25 Mar 2019 18:18
Operator : VA/AP
Sample : K2086-01
Misc : 6.40g/5mL/MSVOA-F/SOIL
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_F
ClientSampled :
CS-11

Manual Integrations
APPROVED

MMDadoda
3/26/2019 3:59:46 PM

Quant Time: Mar 26 05:39:06 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_F\METHODS\82F032219S.M
Quant Title : SW846 8260
QLast Update : Sat Mar 23 01:16:37 2019
Response via : Initial Calibration

