

Data Path : Z:\VOASRV\HPCHEM1\MSVOA F\DATA\VF032519\
 Data File : VF062139.D
 Acq On : 25 Mar 2019 18:47
 Operator : VA/AP
 Sample : K2086-02
 Misc : 8.62g/5mL/MSVOA-F/SOIL
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_F
 ClientSampleID :
 CS-12

Manual Integrations
 APPROVED

MMDadoda
 3/26/2019 3:59:51 PM

Quant Time: Mar 26 05:40:36 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_F\METHODS\82F032219S.M
 Quant Title : SW846 8260
 QLast Update : Sat Mar 23 01:16:37 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	4.87	168	309999	50.00	ug/l	-0.04
34) 1,4-Difluorobenzene	5.60	114	471762	50.00	ug/l	-0.03
63) Chlorobenzene-d5	9.77	117	363104	50.00	ug/l	-0.04
72) 1,4-Dichlorobenzene-d4	12.55	152	153104	50.00	ug/l	-0.02
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	4.84	65	92715	40.21	ug/l	-0.04
Spiked Amount 50.000			Recovery	=	80.42%	
35) Dibromofluoromethane	4.12	113	140903	44.15	ug/l	-0.03
Spiked Amount 50.000			Recovery	=	88.30%	
50) Toluene-d8	7.55	98	318464	33.43	ug/l	-0.04
Spiked Amount 50.000			Recovery	=	66.86%	
62) 4-Bromofluorobenzene	11.40	95	113382	30.83	ug/l	-0.03
Spiked Amount 50.000			Recovery	=	61.66%	
Target Compounds						
16) Acetone	2.23	43	17814	36.708	ug/l	96
20) Methylene Chloride	2.18	84	15052m	2.080	ug/l	

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA F\DATA\VF032519\
Data File : VF062139.D
Acq On : 25 Mar 2019 18:47
Operator : VA/AP
Sample : K2086-02
Misc : 8.62g/5mL/MSVOA-F/SOIL
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_F
ClientSampleId :
CS-12

Manual Integrations
APPROVED

MMDadoda
3/26/2019 3:59:51 PM

Quant Time: Mar 26 05:40:36 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_F\METHODS\82F032219S.M
Quant Title : SW846 8260
QLast Update : Sat Mar 23 01:16:37 2019
Response via : Initial Calibration

