

Data Path : Z:\VOASRV\HPCHEM1\MSVOA F\DATA\VF032619\
 Data File : VF062146.D
 Acq On : 26 Mar 2019 15:20
 Operator : VA/AP
 Sample : K2004-05
 Misc : 4.83g/5mL/MSVOA-F/SOIL
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 MSVOA_F
 ClientSampleId :
 SB-24(7-8)

Manual Integrations
 APPROVED

MMDadoda
 3/27/2019 12:01:42 PM

Quant Time: Mar 27 09:52:54 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_F\METHODS\82F032219S.M
 Quant Title : SW846 8260
 QLast Update : Sat Mar 23 01:16:37 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	4.85	168	152944	50.00	ug/l	-0.06
34) 1,4-Difluorobenzene	5.58	114	247576	50.00	ug/l	-0.05
63) Chlorobenzene-d5	9.76	117	143132	50.00	ug/l	-0.05
72) 1,4-Dichlorobenzene-d4	12.54	152	41090	50.00	ug/l	-0.03
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	4.83	65	56228	49.43	ug/l	-0.05
Spiked Amount 50.000			Recovery	=	98.86%	
35) Dibromofluoromethane	4.10	113	78052	46.61	ug/l	-0.05
Spiked Amount 50.000			Recovery	=	93.22%	
50) Toluene-d8	7.54	98	152837	30.57	ug/l	-0.05
Spiked Amount 50.000			Recovery	=	61.14%	
62) 4-Bromofluorobenzene	11.40	95	57381m	29.73	ug/l	-0.04
Spiked Amount 50.000			Recovery	=	59.46%	
Target Compounds						
16) Acetone	2.22	43	24332	101.626	ug/l	93
20) Methylene Chloride	2.17	84	11395m	5.026	ug/l	

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA F\DATA\VF032619\
Data File : VF062146.D
Acq On : 26 Mar 2019 15:20
Operator : VA/AP
Sample : K2004-05
Misc : 4.83g/5mL/MSVOA-F/SOIL
ALS Vial : 6 Sample Multiplier: 1

Instrument :
MSVOA_F
ClientSampleId :
SB-24(7-8)

Manual Integrations
APPROVED

MMDadoda
3/27/2019 12:01:42 PM

Quant Time: Mar 27 09:52:54 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_F\METHODS\82F032219S.M
Quant Title : SW846 8260
QLast Update : Sat Mar 23 01:16:37 2019
Response via : Initial Calibration

