


Data Path : Z:\voasrv\HPCHEM1\MSVOA F\Data\VF032619\
Data File : VF062141.D
Acq On : 26 Mar 2019 12:31
Operator : VA/AP
Sample : BFB
Misc : 5.00µ/5mL/MSVOA-F/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_F
ClientSampleId :
BFB

Integration File: RTEINT.P

Method : Z:\VOASRV\HPCHEM1\MSVOA_F\METHODS\82F032219S.M
Title : SW846 8260
Last Update : Sat Mar 23 01:16:37 2019



AutoFind: Scans 1091, 1092, 1093; Background Corrected with Scan 1085

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	15.8	15937	PASS
75	95	30	60	39.0	39442	PASS
95	95	100	100	100.0	101048	PASS
96	95	5	9	6.7	6764	PASS
173	174	0.00	2	0.2	188	PASS
174	95	50	100	83.5	84418	PASS
175	174	5	9	7.1	5964	PASS
176	174	95	101	97.3	82152	PASS
177	176	5	9	6.4	5265	PASS