

Data Path : Z:\VOASRV\HPCHEM1\MSVOA F\DATA\VF041619\
 Data File : VF062180.D
 Acq On : 16 Apr 2019 17:15
 Operator : FY/SY
 Sample : K2345-04
 Misc : 7.62g/5mL/MSVOA F/SOIL
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_F
 ClientSampleId :
 TP-2

Manual Integrations
 APPROVED

MMDadoda
 4/17/2019 11:49:44 AM

Quant Time: Apr 17 08:23:43 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_F\METHODS\82F041619S.M
 Quant Title : SW846 8260
 QLast Update : Wed Apr 17 07:56:48 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.06	168	95423	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	5.78	114	243390	50.00	ug/l	0.00
63) Chlorobenzene-d5	9.91	117	229354	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.61	152	107748	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.03	65	141843	165.07	ug/l	0.00
Spiked Amount 50.000			Recovery	=	330.14%	
35) Dibromofluoromethane	4.31	113	155098	80.78	ug/l	0.00
Spiked Amount 50.000			Recovery	=	161.56%	
50) Toluene-d8	7.71	98	232919	41.51	ug/l	0.00
Spiked Amount 50.000			Recovery	=	83.02%	
62) 4-Bromofluorobenzene	11.50	95	124996	56.42	ug/l	0.00
Spiked Amount 50.000			Recovery	=	112.84%	
Target Compounds						
16) Acetone	2.35	43	43835	226.256	ug/l	94
20) Methylene Chloride	2.29	84	46719m	50.483	ug/l	
40) Benzene	4.82	78	10088	1.412	ug/l #	88
52) Toluene	7.78	92	55617	13.254	ug/l	93
59) 2-Hexanone	9.62	43	21020	29.813	ug/l	91
95) Naphthalene	14.49	128	19588	4.404	ug/l #	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA F\DATA\VF041619\
Data File : VF062180.D
Acq On : 16 Apr 2019 17:15
Operator : FY/SY
Sample : K2345-04
Misc : 7.62g/5mL/MSVOA F/SOIL
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_F
ClientSampleId :
TP-2

Manual Integrations
APPROVED

MMDadoda
4/17/2019 11:49:44 AM

Quant Time: Apr 17 08:23:43 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_F\METHODS\82F041619S.M
Quant Title : SW846 8260
QLast Update : Wed Apr 17 07:56:48 2019
Response via : Initial Calibration

