

Data Path : Z:\VOASRV\HPCHEM1\MSVOA F\DATA\VF042219\
 Data File : VF062251.D
 Acq On : 22 Apr 2019 18:14
 Operator : FY/SY
 Sample : K2514-02
 Misc : 4.70µ/5mL/MSVOA F/SOIL
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_F
 ClientSampleId :
 FK-SF-01-007-ABCD

Quant Time: Apr 23 01:08:30 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82F041719S.M
 Quant Title : SW846 8260
 QLast Update : Fri Apr 19 23:52:01 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.07	168	2545	50.00	ug/l	0.04
34) 1,4-Difluorobenzene	5.78	114	1400	50.00	ug/l	0.02
63) Chlorobenzene-d5	9.90	117	399	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.64	152	63	50.00	ug/l	0.04

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.06	65	193	8.27	ug/l	0.05
Spiked Amount	50.000		Recovery	=	16.54%	
35) Dibromofluoromethane	4.30	113	147	12.00	ug/l	0.00
Spiked Amount	50.000		Recovery	=	24.00%	
50) Toluene-d8	7.73	98	778	25.10	ug/l	0.03
Spiked Amount	50.000		Recovery	=	50.20%	
62) 4-Bromofluorobenzene	11.55	95	749	62.81	ug/l	0.07
Spiked Amount	50.000		Recovery	=	125.62%	

Target Compounds					Qvalue
2) Dichlorodifluoromethane	1.01	85	101	2.611	ug/l # 42
3) Chloromethane	1.14	50	237	5.929	ug/l # 56
4) Vinyl Chloride	1.17	62	69	2.009	ug/l # 42
5) Bromomethane	1.40	94	422	21.151	ug/l # 58
6) Chloroethane	1.42	64	440	30.646	ug/l # 41
7) Trichlorofluoromethane	1.47	101	83	1.902	ug/l # 20
8) Diethyl Ether	1.50	74	183	20.152	ug/l # 75
9) 1,1,2-Trichlorotrifluoroet	1.81	101	64	2.147	ug/l # 21
10) Methyl Iodide	2.10	142	62	1.171	ug/l # 42
11) Tert butyl alcohol	2.72	59	157	96.786	ug/l # 1
12) 1,1-Dichloroethene	1.88	96	501	17.672	ug/l # 1
13) Acrolein	2.14	56	322	438.698	ug/l # 29
14) Allyl chloride	2.18	41	558	17.150	ug/l # 48
15) Acrylonitrile	3.08	53	230	46.553	ug/l # 15
16) Acetone	2.38	43	966	173.181	ug/l # 52
17) Carbon Disulfide	1.93	76	212	2.691	ug/l # 1
18) Methyl Acetate	2.46	43	1239	110.132	ug/l # 59
19) Methyl tert-butyl Ether	2.54	73	440	8.790	ug/l # 67
21) trans-1,2-Dichloroethene	2.46	96	199	7.098	ug/l # 1
22) Diisopropyl ether	2.95	45	400	4.806	ug/l # 46
23) Vinyl Acetate	3.34	43	1016	25.622	ug/l # 79
24) 1,1-Dichloroethane	3.00	63	60	1.318	ug/l # 1
25) 2-Butanone	4.51	43	2268	203.345	ug/l # 56
26) 2,2-Dichloropropane	3.57	77	254	11.050	ug/l # 84
27) cis-1,2-Dichloroethene	3.67	96	279	7.685	ug/l # 1
28) Bromochloromethane	3.93	49	59	2.705	ug/l # 8
29) Tetrahydrofuran	4.27	42	329	65.169	ug/l # 61
30) Chloroform	4.04	83	176	3.146	ug/l # 96
31) Cyclohexane	3.84	56	98	1.796	ug/l # 17
32) 1,1,1-Trichloroethane	4.32	97	206	5.030	ug/l # 17
36) 1,1-Dichloropropene	4.43	75	64	3.813	ug/l # 1
37) Ethyl Acetate	4.31	43	1887	238.239	ug/l # 74
38) Carbon Tetrachloride	4.19	117	73	4.913	ug/l # 13
39) Methylcyclohexane	5.66	83	232	11.544	ug/l # 1

Data Path : Z:\VOASRV\HPCHEM1\MSVOA F\DATA\VF042219\
 Data File : VF062251.D
 Acq On : 22 Apr 2019 18:14
 Operator : FY/SY
 Sample : K2514-02
 Misc : 4.70µ/5mL/MSVOA F/SOIL
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_F
 ClientSampleId :
 FK-SF-01-007-ABCD

Quant Time: Apr 23 01:08:30 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82F041719S.M
 Quant Title : SW846 8260
 QLast Update : Fri Apr 19 23:52:01 2019
 Response via : Initial Calibration


Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
40) Benzene	4.83	78	584	13.265	ug/l #	52
41) Methacrylonitrile	4.90	41	725	158.684	ug/l #	40
42) 1,2-Dichloroethane	5.19	62	70	6.014	ug/l #	1
43) Isopropyl Acetate	7.12	43	1882	225.915	ug/l #	45
44) Trichloroethene	5.68	130	66	5.548	ug/l	2
45) 1,2-Dichloropropane	6.40	63	66	6.806	ug/l #	1
46) Dibromomethane	6.39	93	139	21.543	ug/l #	3
47) Bromodichloromethane	6.59	83	175	12.019	ug/l #	24
48) Methyl methacrylate	6.90	41	503	91.877	ug/l #	15
51) 4-Methyl-2-Pentanone	8.45	43	622	100.385	ug/l #	71
52) Toluene	7.77	92	270	11.634	ug/l #	1
53) t-1,3-Dichloropropene	8.47	75	61	5.578	ug/l #	44
54) cis-1,3-Dichloropropene	7.46	75	72	4.867	ug/l #	39
55) 1,1,2-Trichloroethane	8.57	97	73	11.228	ug/l #	41
56) Ethyl methacrylate	8.75	69	288	38.962	ug/l #	76
57) 1,3-Dichloropropane	8.95	76	76	7.239	ug/l #	1
58) 2-Chloroethyl Vinyl ether	7.30	63	81	37.743	ug/l #	48
59) 2-Hexanone	9.45	43	340	88.083	ug/l	95
60) Dibromochloromethane	9.01	129	86	8.895	ug/l	80
61) 1,2-Dibromoethane	8.88	107	73	9.966	ug/l #	1
64) Tetrachloroethene	8.25	164	70	19.435	ug/l #	1
65) Chlorobenzene	9.68	112	62	7.287	ug/l #	42
66) 1,1,1,2-Tetrachloroethane	9.81	131	60	15.060	ug/l #	1
67) Ethyl Benzene	10.02	91	214	13.833	ug/l #	44
68) m/p-Xylenes	10.30	106	76	13.345	ug/l #	1
69) o-Xylene	10.90	106	62	9.805	ug/l	66
70) Styrene	10.85	104	64	7.172	ug/l #	23
73) Isopropylbenzene	11.24	105	98	18.109	ug/l #	49
74) N-amyl acetate	11.39	43	1823	1482.714	ug/l #	65
75) 1,1,2,2-Tetrachloroethane	11.75	83	140	150.426	ug/l #	66
76) 1,2,3-Trichloropropane	11.88	75	185	288.919	ug/l #	1
77) Bromobenzene	11.53	156	64	55.206	ug/l #	1
78) n-propylbenzene	11.66	91	243	42.565	ug/l #	53
79) 2-Chlorotoluene	11.72	91	289	83.453	ug/l	79
80) 1,3,5-Trimethylbenzene	11.96	105	244	54.134	ug/l	64
81) trans-1,4-Dichloro-2-buten	11.94	75	329	1007.469	ug/l #	33
82) 4-Chlorotoluene	11.93	91	135	40.338	ug/l	80
83) tert-Butylbenzene	12.18	119	121	25.959	ug/l #	1
84) 1,2,4-Trimethylbenzene	12.26	105	212	48.605	ug/l #	29
85) sec-Butylbenzene	12.49	105	379	62.630	ug/l	98
86) p-Isopropyltoluene	12.50	119	99	19.286	ug/l #	1
87) 1,3-Dichlorobenzene	12.30	146	60	27.890	ug/l #	46
89) n-Butylbenzene	12.90	91	428	107.335	ug/l #	28
90) Hexachloroethane	12.91	117	60	45.518	ug/l #	12
92) 1,2-Dibromo-3-Chloropropan	13.67	75	92	503.129	ug/l #	12
93) 1,2,4-Trichlorobenzene	14.24	180	59	33.946	ug/l #	11
95) Naphthalene	14.48	128	87	28.628	ug/l #	1


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA F\DATA\VF042219\
Data File : VF062251.D
Acq On : 22 Apr 2019 18:14
Operator : FY/SY
Sample : K2514-02
Misc : 4.70µL/MSVOA F/SOIL
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_F
ClientSampleId :
FK-SF-01-007-ABCD

Quant Time: Apr 23 01:08:30 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82F041719S.M
Quant Title : SW846 8260
QLast Update : Fri Apr 19 23:52:01 2019
Response via : Initial Calibration





#1

Pentafluorobenzene

Concen: 50.000 ug/l

RT: 5.07 min Scan# 461

Delta R.T. 0.04 min

Lab File: VF062251.D

Acq: 22 Apr 2019 18:14

Instrument :

MSVOA_F

ClientSampleId :


FK-SF-01-007-ABCD

Tot Ion:168 Resp: 2545




Ion Ratio Lower Upper

168 100

99 30.4 45.5 68.3#



Abundance Ion 167.90 (167.60 to 168.60): VF0



#2

Dichlorodifluoromethane

Concen: 2.611 ug/l

RT: 1.01 min Scan# 49

Delta R.T. 0.01 min

Lab File: VF062251.D


Acq: 22 Apr 2019 18:14

Tgt Ion: 85 Resp: 101



Ion Ratio Lower Upper


85 100

87 0.0 16.4 49.1#



Abundance Ion 84.90 (84.60 to 85.60): VF0





#3

Chloromethane

Concen: 5.929 ug/l

RT: 1.14 min Scan# 62

Delta R.T. -0.01 min

Lab File: VF062251.D

Acq: 22 Apr 2019 18:14

Instrument :

MSVOA_F

ClientSampleId :


FK-SF-01-007-ABCD

Tot Ion: 50 Resp: 237

Ion Ratio Lower Upper




50 100

52 58.6 27.0 40.4#



Abundance Ion 49.90 (49.60 to 50.60): VF0

Ion 51.90 (51.60 to 52.60): VF0



#4

Vinyl Chloride

Concen: 2.009 ug/l

RT: 1.17 min Scan# 65

Delta R.T. -0.02 min

Lab File: VF062251.D


Acq: 22 Apr 2019 18:14

Tgt Ion: 62 Resp: 69

Ion Ratio Lower Upper



62 100


64 0.0 26.2 39.4#



Abundance Ion 61.90 (61.60 to 62.60): VF0

Ion 63.90 (63.60 to 64.60): VF0





#5

Bromomethane

Concen: 21.151 ug/l

RT: 1.40 min Scan# 88

Delta R.T. -0.00 min

Lab File: VF062251.D

Acq: 22 Apr 2019 18:14

Instrument :

MSVOA_F

ClientSampleId :


FK-SF-01-007-ABCD

Tot Ion: 94 Resp: 422

Ion Ratio Lower Upper




94 100

96 48.3 70.2 105.4#



Abundance Ion 93.90 (93.60 to 94.60): VF0

Ion 95.90 (95.60 to 96.60): VF0



#6

Chloroethane

Concen: 30.646 ug/l

RT: 1.42 min Scan# 90

Delta R.T. -0.03 min

Lab File: VF062251.D


Acq: 22 Apr 2019 18:14

Tgt Ion: 64 Resp: 440

Ion Ratio Lower Upper



64 100


66 0.0 27.0 40.4#



Abundance Ion 63.90 (63.60 to 64.60): VF0

Ion 65.90 (65.60 to 66.60): VF0









#7

Trichlorofluoromethane
Concen: 1.902 ug/l
RT: 1.47 min Scan# 95
Delta R.T. -0.09 min
Lab File: VF062251.D
Acq: 22 Apr 2019 18:14

Instrument :
MSVOA_F
ClientSampleId :
FK-SF-01-007-ABCD




Tot Ion:101 Resp: 83
Ion Ratio Lower Upper
101 100
103 0.0 48.2 72.4#




#8

Diethyl Ether
Concen: 20.152 ug/l
RT: 1.50 min Scan# 98
Delta R.T. -0.21 min
Lab File: VF062251.D
Acq: 22 Apr 2019 18:14

Tgt Ion: 74 Resp: 183
Ion Ratio Lower Upper
74 100
45 170.5 70.0 210.2





#9

1,1,2-Trichlorotrifluoroethane

Concen: 2.147 ug/l

RT: 1.81 min Scan# 130

Delta R.T. -0.10 min

Lab File: VF062251.D

Acq: 22 Apr 2019 18:14

Instrument :

MSVOA_F

ClientSampleId :

FK-SF-01-007-ABCD


Tot Ion:101 Resp: 64

Ion Ratio Lower Upper

101 100

85 0.0 35.3 52.9#




151 0.0 57.5 86.3#



Abundance Ion 100.90 (100.60 to 101.60): V

Ion 84.90 (84.60 to 85.60): VF0

Ion 150.80 (150.50 to 151.50): V



#10

Methyl Iodide

Concen: 1.171 ug/l

RT: 2.10 min Scan# 159

Delta R.T. 0.12 min

Lab File: VF062251.D

Acq: 22 Apr 2019 18:14


Tgt Ion:142 Resp: 62

Ion Ratio Lower Upper

142 100

127 0.0 33.0 49.4#



141 0.0 12.7 19.1#




Abundance Ion 141.80 (141.50 to 142.50): V

Ion 126.80 (126.50 to 127.50): V

Ion 140.80 (140.50 to 141.50): V









#11

Tert butyl alcohol
 Concen: 96.786 ug/l
 RT: 2.72 min Scan# 222
 Delta R.T. 0.03 min
 Lab File: VF062251.D
 Acq: 22 Apr 2019 18:14

Instrument :
 MSVOA_F
 ClientSampleId :
 FK-SF-01-007-ABCD




Tot Ion: 59 Resp: 157
 Ion Ratio Lower Upper
 59 100
 57 463.1 9.3 13.9#




#12

1,1-Dichloroethene
 Concen: 17.672 ug/l
 RT: 1.88 min Scan# 137
 Delta R.T. 0.01 min
 Lab File: VF062251.D
 Acq: 22 Apr 2019 18:14

Tgt Ion: 96 Resp: 501
 Ion Ratio Lower Upper
 96 100
 61 0.0 129.1 193.7#
 98 0.0 50.0 75.0#





#13

Acrolein

Concen: 438.698 ug/l

RT: 2.14 min Scan# 163

Delta R.T. 0.05 min

Lab File: VF062251.D

Acq: 22 Apr 2019 18:14

Instrument :

MSVOA_F

ClientSampleId :





FK-SF-01-007-ABCD

Tot Ion: 56 Resp: 322

Ion Ratio Lower Upper

56 100

55 130.4 57.4 86.0#



#14

Allyl chloride

Concen: 17.150 ug/l

RT: 2.18 min Scan# 167

Delta R.T. -0.01 min

Lab File: VF062251.D

Acq: 22 Apr 2019 18:14




Tgt Ion: 41 Resp: 558


Ion Ratio Lower Upper

41 100

39 127.6 60.0 90.0#

76 13.8 20.6 30.8#





#15

Acrylonitrile

Concen: 46.553 ug/l

RT: 3.08 min Scan# 259

Delta R.T. 0.01 min

Lab File: VF062251.D

Acq: 22 Apr 2019 18:14

Instrument :

MSVOA_F

ClientSampleId :

FK-SF-01-007-ABCD





Tot Ion: 53 Resp: 230

Ion Ratio Lower Upper

53 100

52 0.0 69.5 104.3#

51 0.0 31.8 47.8#



#16

Acetone

Concen: 173.181 ug/l

RT: 2.38 min Scan# 188

Delta R.T. 0.04 min

Lab File: VF062251.D




Acq: 22 Apr 2019 18:14


Tgt Ion: 43 Resp: 966

Ion Ratio Lower Upper

43 100

58 0.0 18.8 28.2#









#17
Carbon Disulfide
Concen: 2.691 ug/l
RT: 1.93 min Scan# 142
Delta R.T. 0.02 min
Lab File: VF062251.D
Acq: 22 Apr 2019 18:14




Instrument :
MSVOA_F
ClientSampleId :
FK-SF-01-007-ABCD


Tot Ion: 76 Resp: 212
Ion Ratio Lower Upper
76 100
78 124.1 8.7 13.1#



#18
Methyl Acetate
Concen: 110.132 ug/l
RT: 2.46 min Scan# 196
Delta R.T. 0.01 min
Lab File: VF062251.D
Acq: 22 Apr 2019 18:14

Tgt Ion: 43 Resp: 1239
Ion Ratio Lower Upper
43 100
74 0.0 14.5 21.7#









#19

Methvl tert-butvl Ether
 Concen: 8.790 ug/l
 RT: 2.54 min Scan# 204
 Delta R.T. -0.00 min
 Lab File: VF062251.D
 Acq: 22 Apr 2019 18:14

Instrument :
 MSVOA_F
 ClientSampleId :
 FK-SF-01-007-ABCD




Tot Ion: 73 Resp: 440
 Ion Ratio Lower Upper
 73 100
 57 4.3 15.1 22.7#




#21

trans-1,2-Dichloroethene
 Concen: 7.098 ug/l
 RT: 2.46 min Scan# 196
 Delta R.T. 0.05 min
 Lab File: VF062251.D
 Acq: 22 Apr 2019 18:14

Tgt Ion: 96 Resp: 199
 Ion Ratio Lower Upper
 96 100
 61 0.0 114.6 172.0#
 98 0.0 51.2 76.8#





#22

Diisopropyl ether

Concen: 4.806 ug/l

RT: 2.95 min Scan# 246

Delta R.T. 0.03 min

Lab File: VF062251.D

Acq: 22 Apr 2019 18:14

Instrument :

MSVOA_F

ClientSampleId :

FK-SF-01-007-ABCD

Tot Ion: 45 Resp: 400


Ion Ratio Lower Upper

45 100

43 0.0 32.8 49.2#

87 0.0 15.2 22.8#

59 0.0 7.8 11.6#



Abundance Ion 45.00 (44.70 to 45.70): VF0



Ion 43.00 (42.70 to 43.70): VF0

Ion 87.00 (86.70 to 87.70): VF0

Ion 59.00 (58.70 to 59.70): VF0

2.95

Time-->



#23

Vinyl Acetate

Concen: 25.622 ug/l

RT: 3.34 min Scan# 285

Delta R.T. 0.01 min

Lab File: VF062251.D


Acq: 22 Apr 2019 18:14

Tgt Ion: 43 Resp: 1016

Ion Ratio Lower Upper

43 100

86 0.0 5.9 8.9#





Abundance Ion 43.00 (42.70 to 43.70): VF0

Ion 86.00 (85.70 to 86.70): VF0

3.34

Time-->





#24

1,1-Dichloroethane

Concen: 1.318 ug/l

RT: 3.00 min Scan# 251

Delta R.T. -0.00 min

Lab File: VF062251.D

Acq: 22 Apr 2019 18:14

Instrument :

MSVOA_F

ClientSampleId :

FK-SF-01-007-ABCD


Tot Ion: 63 Resp: 60

Ion Ratio Lower Upper

63 100

98 134.3 2.3 6.8#




100 0.0 1.7 5.0#



Abundance Ion 62.90 (62.60 to 63.60): VF0

Ion 97.90 (97.60 to 98.60): VF0

Ion 99.90 (99.60 to 100.60): VF0



#25

2-Butanone

Concen: 203.345 ug/l

RT: 4.51 min Scan# 404

Delta R.T. -0.00 min

Lab File: VF062251.D


Acq: 22 Apr 2019 18:14

Tgt Ion: 43 Resp: 2268

Ion Ratio Lower Upper



43 100


72 0.0 16.1 24.1#



Abundance Ion 43.00 (42.70 to 43.70): VF0

Ion 72.00 (71.70 to 72.70): VF0





#26

2,2-Dichloropropane

Concen: 11.050 ug/l

RT: 3.57 min Scan# 308

Delta R.T. -0.19 min

Lab File: VF062251.D

Acq: 22 Apr 2019 18:14

Instrument :

MSVOA_F

ClientSampleId :


FK-SF-01-007-ABCD

Tot Ion: 77 Resp: 254

Ion Ratio Lower Upper

77 100

97 29.9 11.1 33.3



Abundance Ion 76.90 (76.60 to 77.60): VF0

Ion 96.90 (96.60 to 97.60): VF0

3.57

3.57

3.57

3.57

3.57

3.57

3.57

3.57

3.57

3.57

3.57

3.57

3.57

3.57

3.57

3.57

3.57

3.57

3.57

3.57

3.57

3.57

3.57

3.57

3.57

3.57

3.57

3.57

3.57

3.57

3.57

3.57

3.57

3.57

3.57

3.57

3.57

3.57

3.57

3.57

3.57

3.57

3.57

3.57

3.57

3.57

3.57

3.57

3.57


3.57

3.57


3.57

3.57

3.57



Time-->



#27

cis-1,2-Dichloroethene

Concen: 7.685 ug/l

RT: 3.67 min Scan# 319

Delta R.T. 0.04 min

Lab File: VF062251.D

Acq: 22 Apr 2019 18:14


Tgt Ion: 96 Resp: 279

Ion Ratio Lower Upper

96 100

61 0.0 0.0 253.6

98 0.0 0.0 124.8



Abundance Ion 95.90 (95.60 to 96.60): VF0

Ion 60.90 (60.60 to 61.60): VF0

Ion 97.90 (97.60 to 98.60): VF0

3.67

3.67

3.67

3.67

3.67

3.67

3.67

3.67

3.67

3.67

3.67

3.67

3.67

3.67

3.67

3.67

3.67

3.67

3.67

3.67


3.67

3.67


3.67

3.67

3.67



Time-->



#28

Bromochloromethane

Concen: 2.705 ug/l

RT: 3.93 min Scan# 345

Delta R.T. 0.05 min

Lab File: VF062251.D

Acq: 22 Apr 2019 18:14

Instrument :

MSVOA_F

ClientSampleId :

FK-SF-01-007-ABCD


Tot Ion: 49 Resp: 59

Ion Ratio Lower Upper

49 100

129 0.0 0.0 0.0




130 0.0 66.6 100.0#



Abundance Ion 48.90 (48.60 to 49.60): VF0

Ion 128.80 (128.50 to 129.50): V

Ion 129.80 (129.50 to 130.50): V



#29

Tetrahydrofuran

Concen: 65.169 ug/l

RT: 4.27 min Scan# 379

Delta R.T. 0.01 min

Lab File: VF062251.D

Acq: 22 Apr 2019 18:14


Tgt Ion: 42 Resp: 329

Ion Ratio Lower Upper

42 100

72 0.0 28.6 43.0#



71 46.5 29.2 43.8#




Abundance Ion 42.00 (41.70 to 42.70): VF0

Ion 72.00 (71.70 to 72.70): VF0

Ion 71.00 (70.70 to 71.70): VF0









#30
Chloroform
Concen: 3.146 ug/l
RT: 4.04 min Scan# 356
Delta R.T. -0.00 min
Lab File: VF062251.D
Acq: 22 Apr 2019 18:14




Instrument :
MSVOA_F
ClientSampleId :
FK-SF-01-007-ABCD


Tot Ion: 83 Resp: 176
Ion Ratio Lower Upper
83 100
85 68.3 51.8 77.8



#31
Cyclohexane
Concen: 1.796 ug/l
RT: 3.84 min Scan# 336
Delta R.T. -0.02 min
Lab File: VF062251.D
Acq: 22 Apr 2019 18:14

Tgt Ion: 56 Resp: 98
Ion Ratio Lower Upper
56 100
69 0.0 25.1 37.7#
84 0.0 67.8 101.6#





#32

1,1,1-Trichloroethane

Concen: 5.030 ug/l

RT: 4.32 min Scan# 385

Delta R.T. 0.06 min

Lab File: VF062251.D

Acq: 22 Apr 2019 18:14

Instrument :

MSVOA_F

ClientSampleId :

FK-SF-01-007-ABCD


Tot Ion: 97 Resp: 206

Ion Ratio Lower Upper

97 100

99 0.0 52.7 79.1#



61 0.0 53.0 79.4#



Abundance Ion 96.90 (96.60 to 97.60): VF0

Ion 98.90 (98.60 to 99.60): VF0

Ion 60.90 (60.60 to 61.60): VF0



#33

1,2-Dichloroethane-d4

Concen: 5.030 ug/l

RT: 5.06 min Scan# 460

Delta R.T. 0.05 min

Lab File: VF062251.D



Acq: 22 Apr 2019 18:14

Tgt Ion: 65 Resp: 193

Ion Ratio Lower Upper



65 100


67 76.7 0.0 109.4



Abundance Ion 64.90 (64.60 to 65.60): VF0

Ion 66.90 (66.60 to 67.60): VF0





#34

1,4-Difluorobenzene

Concen: 50.000 ug/l

RT: 5.78 min Scan# 533

Delta R.T. 0.02 min

Lab File: VF062251.D

Acq: 22 Apr 2019 18:14

Instrument :

MSVOA_F

ClientSampleId :

FK-SF-01-007-ABCD


Tot Ion:114 Resp: 1400

Ion Ratio Lower Upper

114 100

63 23.7 0.0 36.6




88 0.0 0.0 35.8



Abundance Ion 113.90 (113.60 to 114.60): V

Ion 63.00 (62.70 to 63.70): VF0

Ion 87.90 (87.60 to 88.60): VF0



#35

Dibromofluoromethane

Concen: 50.000 ug/l

RT: 4.30 min Scan# 382

Delta R.T. -0.00 min

Lab File: VF062251.D

Acq: 22 Apr 2019 18:14


Tgt Ion:113 Resp: 147

Ion Ratio Lower Upper

113 100

111 257.1 81.1 121.7#



192 41.5 14.7 22.1#




Abundance Ion 112.80 (112.50 to 113.50): V

Ion 110.80 (110.50 to 111.50): V

Ion 191.70 (191.40 to 192.40): V





#36

1,1-Dichloropropene

Concen: 3.813 ug/l

RT: 4.43 min Scan# 396

Delta R.T. -0.02 min

Lab File: VF062251.D

Acq: 22 Apr 2019 18:14

Instrument :

MSVOA_F

ClientSampleId :

FK-SF-01-007-ABCD


Tot Ion: 75 Resp: 64

Ion Ratio Lower Upper

75 100

110 114.1 18.4 55.2#


77 740.6 25.4 38.2#




Abundance Ion 74.90 (74.60 to 75.60): VF0

Ion 109.90 (109.60 to 110.60): V

Ion 76.90 (76.60 to 77.60): VF0



Time-->



#37

Ethyl Acetate

Concen: 238.239 ug/l

RT: 4.31 min Scan# 384

Delta R.T. 0.03 min

Lab File: VF062251.D

Acq: 22 Apr 2019 18:14



Tgt Ion: 43 Resp: 1887

Ion Ratio Lower Upper

43 100

61 0.0 0.0 0.0


70 0.0 7.8 11.6#




Abundance Ion 43.00 (42.70 to 43.70): VF0


Ion 60.90 (60.60 to 61.60): VF0

Ion 70.00 (69.70 to 70.70): VF0



Time-->





#38

Carbon Tetrachloride

Concen: 4.913 ug/l

RT: 4.19 min Scan# 371

Delta R.T. 0.02 min

Lab File: VF062251.D

Acq: 22 Apr 2019 18:14

Instrument :

MSVOA_F

ClientSampleId :

FK-SF-01-007-ABCD


Tot Ion:117 Resp: 73

Ion Ratio Lower Upper

117 100

119 0.0 75.0 112.4#




121 0.0 26.9 40.3#



Abundance Ion 116.80 (116.50 to 117.50): V

Ion 118.80 (118.50 to 119.50): V

Ion 120.80 (120.50 to 121.50): V



#39

Methylcyclohexane

Concen: 11.544 ug/l

RT: 5.66 min Scan# 520

Delta R.T. 0.04 min

Lab File: VF062251.D

Acq: 22 Apr 2019 18:14


Tgt Ion: 83 Resp: 232

Ion Ratio Lower Upper

83 100

55 201.2 71.0 106.6#



98 85.3 36.7 55.1#




Abundance Ion 83.00 (82.70 to 83.70): VF0

Ion 55.00 (54.70 to 55.70): VF0

Ion 98.00 (97.70 to 98.70): VF0





#40

Benzene

Concen: 13.265 ug/l

RT: 4.83 min Scan# 436

Delta R.T. 0.02 min

Lab File: VF062251.D

Acq: 22 Apr 2019 18:14

Instrument :

MSVOA_F

ClientSampleId :





FK-SF-01-007-ABCD

Tot Ion: 78 Resp: 584

Ion Ratio Lower Upper

78 100

77 0.0 18.9 28.3#



#41

Methacrylonitrile

Concen: 158.684 ug/l

RT: 4.90 min Scan# 443

Delta R.T. -0.02 min

Lab File: VF062251.D

Acq: 22 Apr 2019 18:14

Tgt Ion: 41 Resp: 725




Ion Ratio Lower Upper


41 100

39 111.7 44.7 67.1#

67 27.3 39.1 58.7#

52 9.7 49.8 74.8#





#42

1,2-Dichloroethane

Concen: 6.014 ug/l

RT: 5.19 min Scan# 473

Delta R.T. 0.08 min

Lab File: VF062251.D

Acq: 22 Apr 2019 18:14

Instrument :

MSVOA_F

ClientSampleId :


FK-SF-01-007-ABCD

Tot Ion: 62 Resp: 70

Ion Ratio Lower Upper

62 100

98 114.3 0.0 19.0#



Abundance Ion 61.90 (61.60 to 62.60): VF0

Ion 97.90 (97.60 to 98.60): VF0


150

100


50

0

5.16 5.18 5.20 5.22



Time-->



#43

Isopropyl Acetate

Concen: 225.915 ug/l

RT: 7.12 min Scan# 669

Delta R.T. 0.02 min

Lab File: VF062251.D

Acq: 22 Apr 2019 18:14


Tgt Ion: 43 Resp: 1882

Ion Ratio Lower Upper

43 100

61 0.0 24.4 36.6#

87 0.0 0.4 0.6#



Abundance Ion 43.00 (42.70 to 43.70): VF0

Ion 61.00 (60.70 to 61.70): VF0

Ion 87.00 (86.70 to 87.70): VF0

800


600

400


200

0

7.10 7.12 7.15



Time-->



#44

Trichloroethene

Concen: 5.548 ug/l

RT: 5.68 min Scan# 523

Delta R.T. 0.02 min

Lab File: VF062251.D

Acq: 22 Apr 2019 18:14

Instrument :

MSVOA_F

ClientSampleId :


FK-SF-01-007-ABCD

Tot Ion:130 Resp: 66




Ion Ratio Lower Upper

130 100

95 0.0 0.0 191.4



Abundance Ion 129.80 (129.50 to 130.50): VF0



#45

1,2-Dichloropropane

Concen: 6.806 ug/l

RT: 6.40 min Scan# 596

Delta R.T. 0.01 min

Lab File: VF062251.D


Acq: 22 Apr 2019 18:14

Tgt Ion: 63 Resp: 66



Ion Ratio Lower Upper


63 100

65 90.1 25.9 38.9#



Abundance Ion 62.90 (62.60 to 63.60): VF0





#46

Dibromomethane

Concen: 21.543 ug/l

RT: 6.39 min Scan# 595

Delta R.T. 0.15 min

Lab File: VF062251.D

Acq: 22 Apr 2019 18:14

Instrument :

MSVOA_F

ClientSampleId :

FK-SF-01-007-ABCD


Tot Ion: 93 Resp: 139

Ion Ratio Lower Upper

93 100

95 0.0 65.1 97.7#




174 0.0 81.8 122.6#



Abundance Ion 92.80 (92.50 to 93.50): VF0

Ion 94.80 (94.50 to 95.50): VF0

Ion 173.70 (173.40 to 174.40): V



#47

Bromodichloromethane

Concen: 12.019 ug/l

RT: 6.59 min Scan# 615

Delta R.T. 0.06 min

Lab File: VF062251.D

Acq: 22 Apr 2019 18:14


Tgt Ion: 83 Resp: 175

Ion Ratio Lower Upper

83 100

85 0.0 52.2 78.4#



127 0.0 6.7 10.1#




Abundance Ion 82.90 (82.60 to 83.60): VF0

Ion 84.90 (84.60 to 85.60): VF0

Ion 126.80 (126.50 to 127.50): V









#48
Methvl methacrylate
Concen: 91.877 ug/l
RT: 6.90 min Scan# 646
Delta R.T. 0.04 min
Lab File: VF062251.D
Acq: 22 Apr 2019 18:14




Instrument :
MSVOA_F
ClientSampleId :
FK-SF-01-007-ABCD


Tot Ion: 41 Resp: 503
Ion Ratio Lower Upper
41 100
69 37.8 54.9 82.3#
39 157.7 43.6 65.4#



#50
Toluene-d8
Concen: N.D. ug/l
RT: 7.73 min Scan# 730
Delta R.T. 0.03 min
Lab File: VF062251.D
Acq: 22 Apr 2019 18:14

Tgt Ion: 98 Resp: 778
Ion Ratio Lower Upper
98 100
100 30.8 56.2 84.2#





#51

4-Methyl-2-Pentanone

Concen: 100.385 ug/l

RT: 8.45 min Scan# 803

Delta R.T. 0.06 min

Lab File: VF062251.D

Acq: 22 Apr 2019 18:14

Instrument :

MSVOA_F

ClientSampleId :


FK-SF-01-007-ABCD

Tot Ion: 43 Resp: 622

Ion Ratio Lower Upper




43 100

58 19.6 29.3 43.9#



Abundance Ion 43.00 (42.70 to 43.70): VF0

Ion 58.00 (57.70 to 58.70): VF0



#52

Toluene

Concen: 11.634 ug/l

RT: 7.77 min Scan# 735

Delta R.T. 0.01 min

Lab File: VF062251.D


Acq: 22 Apr 2019 18:14

Tgt Ion: 92 Resp: 270

Ion Ratio Lower Upper



92 100


91 313.7 130.1 195.1#



Abundance Ion 92.00 (91.70 to 92.70): VF0

Ion 91.00 (90.70 to 91.70): VF0









#53
 t-1,3-Dichloropropene
 Concen: 5.578 ug/l
 RT: 8.47 min Scan# 806
 Delta R.T. 0.07 min
 Lab File: VF062251.D
 Acq: 22 Apr 2019 18:14




Instrument :
 MSVOA_F
 ClientSampleId :
 FK-SF-01-007-ABCD


Tot Ion: 75 Resp: 61
 Ion Ratio Lower Upper
 75 100
 77 0.0 24.6 36.8#



#54
 cis-1,3-Dichloropropene
 Concen: 4.867 ug/l
 RT: 7.46 min Scan# 703
 Delta R.T. 0.02 min
 Lab File: VF062251.D
 Acq: 22 Apr 2019 18:14

Tgt Ion: 75 Resp: 72
 Ion Ratio Lower Upper
 75 100
 77 0.0 26.4 39.6#
 39 12.3 46.7 70.1#





#55

1,1,2-Trichloroethane

Concen: 11.228 ug/l

RT: 8.57 min Scan# 816

Delta R.T. -0.05 min

Lab File: VF062251.D

Acq: 22 Apr 2019 18:14

Instrument :

MSVOA_F

ClientSampleId :

FK-SF-01-007-ABCD

Tot Ion: 97 Resp: 73


Ion Ratio Lower Upper

97 100

83 114.6 70.2 105.2#

85 0.0 46.2 69.4#

99 0.0 52.2 78.4#






Abundance Ion 96.90 (96.60 to 97.60): VF0

Ion 82.90 (82.60 to 83.60): VF0

Ion 84.90 (84.60 to 85.60): VF0

Ion 98.90 (98.60 to 99.60): VF0



#56

Ethyl methacrylate

Concen: 38.962 ug/l

RT: 8.75 min Scan# 834

Delta R.T. 0.01 min

Lab File: VF062251.D

Acq: 22 Apr 2019 18:14


Tgt Ion: 69 Resp: 288

Ion Ratio Lower Upper

69 100

41 78.1 61.5 92.3



39 0.0 33.3 49.9#




Abundance Ion 69.00 (68.70 to 69.70): VF0

Ion 41.00 (40.70 to 41.70): VF0

Ion 39.00 (38.70 to 39.70): VF0





#57

1,3-Dichloropropane

Concen: 7.239 ug/l

RT: 8.95 min Scan# 854

Delta R.T. -0.03 min

Lab File: VF062251.D

Acq: 22 Apr 2019 18:14

Instrument :

MSVOA_F

ClientSampleId :


FK-SF-01-007-ABCD

Tot Ion: 76 Resp: 76

Ion Ratio Lower Upper




76 100

78 325.0 26.5 39.7#



Abundance Ion 75.90 (75.60 to 76.60): VF0

Ion 77.90 (77.60 to 78.60): VF0



#58

2-Chloroethyl Vinyl ether

Concen: 37.743 ug/l

RT: 7.30 min Scan# 687

Delta R.T. -0.13 min

Lab File: VF062251.D


Acq: 22 Apr 2019 18:14

Tgt Ion: 63 Resp: 81

Ion Ratio Lower Upper



63 100


106 0.0 21.4 32.0#



Abundance Ion 62.90 (62.60 to 63.60): VF0

Ion 105.90 (105.60 to 106.60): V









#59
2-Hexanone
Concen: 88.083 ug/l
RT: 9.45 min Scan# 905
Delta R.T. -0.16 min
Lab File: VF062251.D
Acq: 22 Apr 2019 18:14




Instrument :
MSVOA_F
ClientSampleId :
FK-SF-01-007-ABCD


Tot Ion: 43 Resp: 340
Ion Ratio Lower Upper
43 100
58 45.9 24.6 74.0



#60
Dibromochloromethane
Concen: 8.895 ug/l
RT: 9.01 min Scan# 860
Delta R.T. 0.17 min
Lab File: VF062251.D
Acq: 22 Apr 2019 18:14

Tgt Ion:129 Resp: 86
Ion Ratio Lower Upper
129 100
127 95.3 39.0 117.0





#61

1,2-Dibromoethane

Concen: 9.966 ug/l

RT: 8.88 min Scan# 847

Delta R.T. -0.24 min

Lab File: VF062251.D

Acq: 22 Apr 2019 18:14

Instrument :

MSVOA_F

ClientSampleId :


FK-SF-01-007-ABCD

Tot Ion:107 Resp: 73

Ion Ratio Lower Upper

107 100

109 220.5 74.3 111.5#




Abundance Ion 106.90 (106.60 to 107.60): V


Ion 108.80 (108.50 to 109.50): V

8.88

Time-->



Time-->



#62

4-Bromofluorobenzene

Concen: 9.966 ug/l

RT: 11.55 min Scan# 1118

Delta R.T. 0.07 min

Lab File: VF062251.D

Acq: 22 Apr 2019 18:14


Tgt Ion: 95 Resp: 749

Ion Ratio Lower Upper

95 100

174 0.0 0.0 159.4

176 10.0 0.0 155.4




Abundance Ion 94.90 (94.60 to 95.60): VF0

Ion 173.80 (173.50 to 174.50): V


Ion 175.80 (175.50 to 176.50): V

11.55

Time-->







Time-->



#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 9.90 min Scan# 951
Delta R.T. 0.01 min
Lab File: VF062251.D
Acq: 22 Apr 2019 18:14




Instrument :
MSVOA_F
ClientSampleId :
FK-SF-01-007-ABCD


Tot Ion:117 Resp: 399
Ion Ratio Lower Upper
117 100
82 51.3 43.2 64.8
119 44.9 26.2 39.4#



#64
Tetrachloroethene
Concen: 19.435 ug/l
RT: 8.25 min Scan# 783
Delta R.T. -0.04 min
Lab File: VF062251.D
Acq: 22 Apr 2019 18:14

Tgt Ion:164 Resp: 70
Ion Ratio Lower Upper
164 100
166 0.0 104.7 157.1#
129 0.0 72.6 109.0#
131 0.0 71.6 107.4#









#65
Chlorobenzene
Concen: 7.287 ug/l
RT: 9.68 min Scan# 928
Delta R.T. -0.24 min
Lab File: VF062251.D
Acq: 22 Apr 2019 18:14

Instrument :
MSVOA_F
ClientSampleId :
FK-SF-01-007-ABCD

Tot Ion:112 Resp: 62
Ion Ratio Lower Upper
112 100
114 0.0 26.1 39.1#




Abundance Ion 111.90 (111.60 to 112.60): V
Ion 113.90 (113.60 to 114.60): V






#66
1,1,1,2-Tetrachloroethane
Concen: 15.060 ug/l
RT: 9.81 min Scan# 941
Delta R.T. -0.24 min
Lab File: VF062251.D
Acq: 22 Apr 2019 18:14

Tgt Ion:131 Resp: 60
Ion Ratio Lower Upper
131 100
133 583.3 47.7 143.1#
119 230.0 34.6 103.9#



Abundance Ion 130.80 (130.50 to 131.50): V
Ion 132.80 (132.50 to 133.50): V
Ion 118.80 (118.50 to 119.50): V





#67

Ethyl Benzene

Concen: 13.833 ug/l

RT: 10.02 min Scan# 963

Delta R.T. 0.01 min

Lab File: VF062251.D

Acq: 22 Apr 2019 18:14

Instrument :

MSVOA_F

ClientSampleId :


FK-SF-01-007-ABCD

Tot Ion: 91 Resp: 214

Ion Ratio Lower Upper




91 100

106 0.0 24.9 37.3#



Abundance Ion 91.00 (90.70 to 91.70): VF0

Ion 106.00 (105.70 to 106.70): V



#68

m/p-Xylenes

Concen: 13.345 ug/l

RT: 10.30 min Scan# 991

Delta R.T. 0.06 min

Lab File: VF062251.D


Acq: 22 Apr 2019 18:14

Tgt Ion:106 Resp: 76

Ion Ratio Lower Upper



106 100


91 357.9 159.1 238.7#



Abundance Ion 106.00 (105.70 to 106.70): V

Ion 91.00 (90.70 to 91.70): VF0









#69
o-Xylene
Concen: 9.805 ug/l
RT: 10.90 min Scan# 1052
Delta R.T. 0.10 min
Lab File: VF062251.D
Acq: 22 Apr 2019 18:14




Instrument :
MSVOA_F
ClientSampleId :
FK-SF-01-007-ABCD


Tot Ion:106 Resp: 62
Ion Ratio Lower Upper
106 100
91 269.4 107.5 322.4



#70
Styrene
Concen: 7.172 ug/l
RT: 10.85 min Scan# 1047
Delta R.T. -0.03 min
Lab File: VF062251.D
Acq: 22 Apr 2019 18:14

Tgt Ion:104 Resp: 64
Ion Ratio Lower Upper
104 100
78 0.0 34.5 51.7#
103 103.1 37.2 55.8#





#72

1,4-Dichlorobenzene-d4

Concen: 50.000 ug/l

RT: 12.64 min Scan# 1229

Delta R.T. 0.04 min

Lab File: VF062251.D

Acq: 22 Apr 2019 18:14

Instrument :

MSVOA_F

ClientSampleId :

FK-SF-01-007-ABCD


Tot Ion:152 Resp: 63

Ion Ratio Lower Upper

152 100

115 123.8 27.8 83.3#




150 0.0 0.0 341.8



Abundance Ion 151.90 (151.60 to 152.60): V

Ion 114.90 (114.60 to 115.60): V

Ion 149.90 (149.60 to 150.60): V



#73

Isopropylbenzene

Concen: 18.109 ug/l

RT: 11.24 min Scan# 1087

Delta R.T. 0.04 min

Lab File: VF062251.D


Acq: 22 Apr 2019 18:14

Tgt Ion:105 Resp: 98

Ion Ratio Lower Upper



105 100


120 0.0 13.0 38.9#



Abundance Ion 105.00 (104.70 to 105.70): V

Ion 120.00 (119.70 to 120.70): V





#74

N-amvl acetate

Concen: 1482.714 ug/l

RT: 11.39 min Scan# 1102

Delta R.T. -0.04 min

Lab File: VF062251.D

Acq: 22 Apr 2019 18:14

Instrument :

MSVOA_F

ClientSampleId :

FK-SF-01-007-ABCD

Tot Ion: 43 Resp: 1823


Ion Ratio Lower Upper

43 100

70 13.1 27.0 40.4#

55 13.6 15.7 23.5#

61 0.0 20.8 31.2#






Abundance Ion 43.00 (42.70 to 43.70): VF0

Ion 70.00 (69.70 to 70.70): VF0

Ion 55.00 (54.70 to 55.70): VF0

Ion 61.00 (60.70 to 61.70): VF0



#75

1,1,2,2-Tetrachloroethane

Concen: 150.426 ug/l

RT: 11.75 min Scan# 1138

Delta R.T. -0.01 min

Lab File: VF062251.D

Acq: 22 Apr 2019 18:14


Tgt Ion: 83 Resp: 140

Ion Ratio Lower Upper

83 100

131 0.0 4.5 13.5#



85 97.9 34.7 104.1




Abundance Ion 82.90 (82.60 to 83.60): VF0

Ion 130.80 (130.50 to 131.50): V

Ion 84.90 (84.60 to 85.60): VF0





#76

1,2,3-Trichloropropane

Concen: 288.919 ug/l

RT: 11.88 min Scan# 1151

Delta R.T. 0.02 min

Lab File: VF062251.D

Acq: 22 Apr 2019 18:14

Instrument :

MSVOA_F

ClientSampleId :


FK-SF-01-007-ABCD

Tot Ion: 75 Resp: 185

Ion Ratio Lower Upper




75 100

77 148.6 7.4 22.2#



Abundance Ion 74.90 (74.60 to 75.60): VF0

Ion 77.00 (76.70 to 77.70): VF0



#77

Bromobenzene

Concen: 55.206 ug/l

RT: 11.53 min Scan# 1116

Delta R.T. -0.04 min

Lab File: VF062251.D

Acq: 22 Apr 2019 18:14


Tgt Ion: 156 Resp: 64

Ion Ratio Lower Upper

156 100

77 453.1 80.9 242.7#



158 0.0 49.8 149.4#




Abundance Ion 155.80 (155.50 to 156.50): V

Ion 77.00 (76.70 to 77.70): VF0

Ion 157.80 (157.50 to 158.50): V





#78

n-propylbenzene

Concen: 42.565 ug/l

RT: 11.66 min Scan# 1129

Delta R.T. -0.00 min

Lab File: VF062251.D

Acq: 22 Apr 2019 18:14

Instrument :

MSVOA_F

ClientSampleId :





FK-SF-01-007-ABCD

Tot Ion: 91 Resp: 243

Ion Ratio Lower Upper

91 100

120 0.0 11.2 33.6#



#79

2-Chlorotoluene

Concen: 83.453 ug/l

RT: 11.72 min Scan# 1135

Delta R.T. -0.07 min

Lab File: VF062251.D




Acq: 22 Apr 2019 18:14


Tgt Ion: 91 Resp: 289

Ion Ratio Lower Upper

91 100

126 21.5 16.9 50.6





#80

1,3,5-Trimethylbenzene

Concen: 54.134 ug/l

RT: 11.96 min Scan# 1160

Delta R.T. 0.08 min

Lab File: VF062251.D

Acq: 22 Apr 2019 18:14

Instrument :

MSVOA_F

ClientSampleId :


FK-SF-01-007-ABCD

Tot Ion:105 Resp: 244

Ion Ratio Lower Upper




105 100

120 27.0 26.2 78.5



Abundance Ion 105.00 (104.70 to 105.70): V

Ion 120.00 (119.70 to 120.70): V



#81

trans-1,4-Dichloro-2-butene

Concen: 1007.469 ug/l

RT: 11.94 min Scan# 1158

Delta R.T. 0.02 min

Lab File: VF062251.D

Acq: 22 Apr 2019 18:14


Tgt Ion: 75 Resp: 329

Ion Ratio Lower Upper

75 100

53 114.6 50.2 75.4#



89 27.1 0.0 0.0#




Abundance Ion 74.90 (74.60 to 75.60): VF0

Ion 53.00 (52.70 to 53.70): VF0

Ion 88.90 (88.60 to 89.60): VF0









#82
4-Chlorotoluene
Concen: 40.338 ug/l
RT: 11.93 min Scan# 1157
Delta R.T. -0.03 min
Lab File: VF062251.D
Acq: 22 Apr 2019 18:14




Instrument :
MSVOA_F
ClientSampleId :
FK-SF-01-007-ABCD


Tot Ion: 91 Resp: 135
Ion Ratio Lower Upper
91 100
126 45.2 16.9 50.6



#83
tert-Butylbenzene
Concen: 25.959 ug/l
RT: 12.18 min Scan# 1182
Delta R.T. -0.01 min
Lab File: VF062251.D
Acq: 22 Apr 2019 18:14

Tgt Ion: 119 Resp: 121
Ion Ratio Lower Upper
119 100
91 210.7 29.4 88.2#
134 0.0 11.8 35.3#









#84
 1,2,4-Trimethylbenzene
 Concen: 48.605 ug/l
 RT: 12.26 min Scan# 1190
 Delta R.T. -0.00 min
 Lab File: VF062251.D
 Acq: 22 Apr 2019 18:14




Instrument :
 MSVOA_F
 ClientSampleId :
 FK-SF-01-007-ABCD


Tot Ion:105 Resp: 212
 Ion Ratio Lower Upper
 105 100
 120 0.0 24.0 72.0#



#85
 sec-Butylbenzene
 Concen: 62.630 ug/l
 RT: 12.49 min Scan# 1213
 Delta R.T. 0.13 min
 Lab File: VF062251.D
 Acq: 22 Apr 2019 18:14

Tgt Ion:105 Resp: 379
 Ion Ratio Lower Upper
 105 100
 134 18.5 9.7 29.1





#86

p-Isopropyltoluene

Concen: 19.286 ug/l

RT: 12.50 min Scan# 1214

Delta R.T. -0.01 min

Lab File: VF062251.D

Acq: 22 Apr 2019 18:14


Instrument :

MSVOA_F

ClientSampleId :

FK-SF-01-007-ABCD




Tot Ion:	119	Resp:	99
Ion	Ratio	Lower	Upper
119	100		
134	70.7	13.2	39.5#
91	93.9	11.7	35.1#



Abundance Ion 119.00 (118.70 to 119.70): V

Ion 134.00 (133.70 to 134.70): V

Ion 91.00 (90.70 to 91.70): V



#87

1,3-Dichlorobenzene

Concen: 27.890 ug/l


RT: 12.30 min Scan# 1194

Delta R.T. -0.23 min

Lab File: VF062251.D

Acq: 22 Apr 2019 18:14



Tgt Ion:	146	Resp:	60
Ion	Ratio	Lower	Upper
146	100		
111	0.0	19.7	59.0#
148	103.3	32.9	98.6#




Abundance Ion 145.90 (145.60 to 146.60): V

Ion 110.90 (110.60 to 111.60): V

Ion 147.90 (147.60 to 148.60): V





#89

n-Butylbenzene

Concen: 107.335 ug/l

RT: 12.90 min Scan# 1255

Delta R.T. 0.01 min

Lab File: VF062251.D

Acq: 22 Apr 2019 18:14

Instrument :

MSVOA_F

ClientSampleId :

FK-SF-01-007-ABCD


Tot Ion: 91 Resp: 428

Ion Ratio Lower Upper

91 100

92 0.0 31.2 93.6#


134 0.0 13.1 39.1#





Abundance Ion 91.00 (90.70 to 91.70): VF0

Ion 92.00 (91.70 to 92.70): VF0

Ion 134.00 (133.70 to 134.70): V



Time-->



#90

Hexachloroethane

Concen: 45.518 ug/l

RT: 12.91 min Scan# 1256

Delta R.T. -0.05 min

Lab File: VF062251.D


Acq: 22 Apr 2019 18:14

Tgt Ion: 117 Resp: 60

Ion Ratio Lower Upper


117 100

201 0.0 37.7 113.1#





Abundance Ion 116.80 (116.50 to 117.50): V

Ion 200.70 (200.40 to 201.40): V



Time-->





#92

1,2-Dibromo-3-Chloropropane

Concen: 503.129 ug/l

RT: 13.67 min Scan# 1333

Delta R.T. -0.01 min

Lab File: VF062251.D

Acq: 22 Apr 2019 18:14

Instrument :

MSVOA_F

ClientSampleId :

FK-SF-01-007-ABCD


Tot Ion: 75 Resp: 92

Ion Ratio Lower Upper

75 100

155 155.4 47.6 142.8#

157 0.0 61.3 183.9#





Abundance Ion 74.90 (74.60 to 75.60): VF0

Ion 154.80 (154.50 to 155.50): V

Ion 156.80 (156.50 to 157.50): V

13.67

Time-->



#93

1,2,4-Trichlorobenzene

Concen: 33.946 ug/l

RT: 14.24 min Scan# 1391

Delta R.T. 0.01 min

Lab File: VF062251.D

Acq: 22 Apr 2019 18:14


Tgt Ion:180 Resp: 59

Ion Ratio Lower Upper

180 100

182 0.0 49.3 147.9#

145 0.0 18.1 54.4#




Abundance Ion 179.80 (179.50 to 180.50): V


Ion 181.80 (181.50 to 182.50): V

Ion 144.80 (144.50 to 145.50): V

14.24

Time-->





#95

Naphthalene

Concen: 28.628 ug/l

RT: 14.48 min Scan# 1415

Delta R.T. -0.01 min

Lab File: VF062251.D

Acq: 22 Apr 2019 18:14

Instrument :

MSVOA_F

ClientSampleId :

FK-SF-01-007-ABCD

Tot Ion:128 Resp: 87

Ion Ratio Lower Upper

128 100

127 0.0 10.4 15.6#

129 87.4 9.2 13.8#

