

Data Path : Z:\VOASRV\HPCHEM1\MSVOA F\DATA\VF042219\
 Data File : VF062248.D
 Acq On : 22 Apr 2019 16:48
 Operator : FY/SY
 Sample : K2475-05RE
 Misc : 4.56g/5mL/MSVOA F/SOIL
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_F
 ClientSampleId :
 TP-TRACK-74-75RE

Manual Integrations
 APPROVED

MMDadoda
 4/23/2019 9:22:27 AM

Quant Time: Apr 23 01:44:46 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82F041719S.M
 Quant Title : SW846 8260
 QLast Update : Fri Apr 19 23:52:01 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.06	168	455894	50.00	ug/l	0.02
34) 1,4-Difluorobenzene	5.78	114	671476	50.00	ug/l	0.01
63) Chlorobenzene-d5	9.91	117	512911	50.00	ug/l	0.01
72) 1,4-Dichlorobenzene-d4	12.61	152	303897	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.03	65	173399	41.48	ug/l	0.01
Spiked Amount 50.000			Recovery	=	82.96%	
35) Dibromofluoromethane	4.31	113	242416	41.27	ug/l	0.01
Spiked Amount 50.000			Recovery	=	82.54%	
50) Toluene-d8	7.71	98	485805	32.67	ug/l	0.01
Spiked Amount 50.000			Recovery	=	65.34%	
62) 4-Bromofluorobenzene	11.49	95	220162	38.49	ug/l	0.01
Spiked Amount 50.000			Recovery	=	76.98%	
Target Compounds						
16) Acetone	2.35	43	64912	64.964	ug/l	94
20) Methylene Chloride	2.34	84	49532m	4.380	ug/l	

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA F\DATA\VF042219\
Data File : VF062248.D
Acq On : 22 Apr 2019 16:48
Operator : FY/SY
Sample : K2475-05RE
Misc : 4.56g/5mL/MSVOA F/SOIL
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_F
ClientSampleId :
TP-TRACK-74-75RE

Manual Integrations
APPROVED

MMDadoda
4/23/2019 9:22:27 AM

Quant Time: Apr 23 01:44:46 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82F041719S.M
Quant Title : SW846 8260
QLast Update : Fri Apr 19 23:52:01 2019
Response via : Initial Calibration

