


Data Path : Z:\voasrv\HPCHEM1\MSVOA F\Data\VF042619\
 Data File : VF062302.D
 Acq On : 26 Apr 2019 12:22
 Operator : FY/SY
 Sample : BFB
 Misc : 5.00µ/5mL/MSVOA F/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_F
 ClientSampleId :
 BFB

Integration File: RTEINT.P

Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82F042419S.M
 Title : SW846 8260
 Last Update : Thu Apr 25 08:40:59 2019



AutoFind: Scans 1113, 1114, 1115; Background Corrected with Scan 1107

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	17.6	19438	PASS
75	95	30	60	42.8	47229	PASS
95	95	100	100	100.0	110378	PASS
96	95	5	9	6.4	7081	PASS
173	174	0.00	2	0.0	0	PASS
174	95	50	100	97.5	107600	PASS
175	174	5	9	7.7	8246	PASS
176	174	95	101	97.5	104901	PASS
177	176	5	9	6.9	7190	PASS