

Data Path : Z:\VOASRV\HPCHEM1\MSVOA F\DATA\VF042919\
 Data File : VF062324.D
 Acq On : 29 Apr 2019 20:33
 Operator : FY/SY
 Sample : K2194-07RE
 Misc : 6.26g/5mL/MSVOA F/SOIL
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_F
 ClientSampleId :
 JC-03-042919-DRE

Quant Time: Apr 30 04:55:02 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82F042419S.M
 Quant Title : SW846 8260
 QLast Update : Thu Apr 25 08:40:59 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.04	168	232492	50.00	ug/l	-0.03
34) 1,4-Difluorobenzene	5.76	114	328513	50.00	ug/l	-0.03
63) Chlorobenzene-d5	9.90	117	269000	50.00	ug/l	-0.02
72) 1,4-Dichlorobenzene-d4	12.60	152	153940	50.00	ug/l	-0.02
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.02	65	123935	65.81	ug/l	-0.02
Spiked Amount 50.000			Recovery	=	131.62%	
35) Dibromofluoromethane	4.30	113	145610	59.78	ug/l	-0.02
Spiked Amount 50.000			Recovery	=	119.56%	
50) Toluene-d8	7.70	98	283464	47.64	ug/l	-0.02
Spiked Amount 50.000			Recovery	=	95.28%	
62) 4-Bromofluorobenzene	11.49	95	139392	56.84	ug/l	-0.02
Spiked Amount 50.000			Recovery	=	113.68%	
Target Compounds						
16) Acetone	2.34	43	36897	87.507	ug/l #	82
20) Methylene Chloride	2.32	84	14294	2.978	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA F\DATA\VF042919\
Data File : VF062324.D
Acq On : 29 Apr 2019 20:33
Operator : FY/SY
Sample : K2194-07RE
Misc : 6.26g/5mL/MSVOA F/SOIL
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_F
ClientSampleId :
JC-03-042919-DRE

Quant Time: Apr 30 04:55:02 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82F042419S.M
Quant Title : SW846 8260
QLast Update : Thu Apr 25 08:40:59 2019
Response via : Initial Calibration

