

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA F\DATA\VF050719\
 Data File : VF062362.D
 Acq On : 7 May 2019 17:25
 Operator : FY/SY
 Sample : K2635-07
 Misc : 6.26g/5mL/MSVOA F/SOIL
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_F
 ClientSampleId :
 AU-05-050219-D

Manual Integrations
 APPROVED

MMDadoda
 5/8/2019 11:10:51 AM

Quant Time: May 07 17:50:12 2019
 Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82F050619S.M
 Quant Title : SW846 8260
 QLast Update : Tue May 07 04:34:57 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.04	168	315572	50.00	ug/l	-0.01
34) 1,4-Difluorobenzene	5.76	114	434284	50.00	ug/l	-0.01
63) Chlorobenzene-d5	9.90	117	333209	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.60	152	189680	50.00	ug/l	-0.01
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.02	65	145104	48.66	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.32%	
35) Dibromofluoromethane	4.29	113	163259	47.19	ug/l	-0.01
Spiked Amount 50.000			Recovery	=	94.38%	
50) Toluene-d8	7.70	98	326800	39.02	ug/l	0.00
Spiked Amount 50.000			Recovery	=	78.04%	
62) 4-Bromofluorobenzene	11.49	95	145946	38.62	ug/l	0.00
Spiked Amount 50.000			Recovery	=	77.24%	
Target Compounds						
20) Methylene Chloride	2.35	84	16834m	7.712	ug/l	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA F\DATA\VF050719\
Data File : VF062362.D
Acq On : 7 May 2019 17:25
Operator : FY/SY
Sample : K2635-07
Misc : 6.26g/5mL/MSVOA F/SOIL
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_F
ClientSampleId :
AU-05-050219-D

Manual Integrations
APPROVED

MMDadoda
5/8/2019 11:10:51 AM

Quant Time: May 07 17:50:12 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82F050619S.M
Quant Title : SW846 8260
QLast Update : Tue May 07 04:34:57 2019
Response via : Initial Calibration

