

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA F\DATA\VF050819\
 Data File : VF062380.D
 Acq On : 8 May 2019 16:36
 Operator : FY/SY
 Sample : K2673-11RE
 Misc : 5.08g/5mL/MSVOA F/SOIL
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_F
 ClientSampleId :
 SB-16-2.0-2.5RE

Quant Time: May 09 05:22:20 2019
 Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82F050619S.M
 Quant Title : SW846 8260
 QLast Update : Tue May 07 04:34:57 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.03	168	267911	50.00	ug/l	-0.02
34) 1,4-Difluorobenzene	5.76	114	363612	50.00	ug/l	0.00
63) Chlorobenzene-d5	9.89	117	245924	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.60	152	94895	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.01	65	125639	49.62	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.24%	
35) Dibromofluoromethane	4.29	113	150378	51.91	ug/l	0.00
Spiked Amount 50.000			Recovery	=	103.82%	
50) Toluene-d8	7.70	98	255443	36.43	ug/l	0.00
Spiked Amount 50.000			Recovery	=	72.86%	
62) 4-Bromofluorobenzene	11.48	95	91217	28.83	ug/l	0.00
Spiked Amount 50.000			Recovery	=	57.66%	
Target Compounds						
20) Methylene Chloride	2.33	84	24762m	13.363	ug/l	Qvalue
95) Naphthalene	14.49	128	23366	6.639	ug/l	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA F\DATA\VF050819\
Data File : VF062380.D
Acq On : 8 May 2019 16:36
Operator : FY/SY
Sample : K2673-11RE
Misc : 5.08g/5mL/MSVOA F/SOIL
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_F
ClientSampleId :
SB-16-2.0-2.5RE

Quant Time: May 09 05:22:20 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82F050619S.M
Quant Title : SW846 8260
QLast Update : Tue May 07 04:34:57 2019
Response via : Initial Calibration

