

Data Path : Z:\VOASRV\HPCHEM1\MSVOA F\DATA\VF050919\
 Data File : VF062402.D
 Acq On : 9 May 2019 18:14
 Operator : FY/SY
 Sample : K2641-03
 Misc : 6.95g/5mL/MSVOA F/SOIL
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_F
 ClientSampleId :
 CL-02-050719-B

Manual Integrations
 APPROVED

MMDadoda
 5/10/2019 11:03:48 AM

Quant Time: May 10 08:17:17 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82F050619S.M
 Quant Title : SW846 8260
 QLast Update : Tue May 07 04:34:57 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.05	168	273559	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	5.77	114	344204	50.00	ug/l	0.00
63) Chlorobenzene-d5	9.90	117	180165	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.61	152	107452	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.02	65	213261	82.49	ug/l	0.00
Spiked Amount 50.000			Recovery	=	164.98%	
35) Dibromofluoromethane	4.30	113	217959	79.48	ug/l	0.00
Spiked Amount 50.000			Recovery	=	158.96%	
50) Toluene-d8	7.70	98	127510	19.21	ug/l	0.00
Spiked Amount 50.000			Recovery	=	38.42%	
62) 4-Bromofluorobenzene	11.48	95	113632	37.94	ug/l	0.00
Spiked Amount 50.000			Recovery	=	75.88%	
Target Compounds						
16) Acetone	2.34	43	58611	112.072	ug/l #	92
20) Methylene Chloride	2.34	84	21320m	11.268	ug/l	

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA F\DATA\VF050919\
Data File : VF062402.D
Acq On : 9 May 2019 18:14
Operator : FY/SY
Sample : K2641-03
Misc : 6.95g/5mL/MSVOA F/SOIL
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_F
ClientSampleId :
CL-02-050719-B

Manual Integrations
APPROVED

MMDadoda
5/10/2019 11:03:48 AM

Quant Time: May 10 08:17:17 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82F050619S.M
Quant Title : SW846 8260
QLast Update : Tue May 07 04:34:57 2019
Response via : Initial Calibration

