

Data Path : Z:\VOASRV\HPCHEM1\MSVOA F\DATA\VF050919\
 Data File : VF062404.D
 Acq On : 9 May 2019 19:10
 Operator : FY/SY
 Sample : K2641-07
 Misc : 6.01g/5mL/MSVOA F/SOIL
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_F
 ClientSampleId :
 CL-02-050719-D

Manual Integrations
 APPROVED

MMDadoda
 5/10/2019 11:17:20 AM

Quant Time: May 10 03:43:37 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82F050619S.M
 Quant Title : SW846 8260
 QLast Update : Tue May 07 04:34:57 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.05	168	291457	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	5.77	114	313938	50.00	ug/l	0.00
63) Chlorobenzene-d5	9.91	117	102313	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.61	152	152320	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.02	65	200329	72.73	ug/l	0.00
Spiked Amount 50.000			Recovery	=	145.46%	
35) Dibromofluoromethane	4.30	113	233547	93.38	ug/l	0.00
Spiked Amount 50.000			Recovery	=	186.76%	
50) Toluene-d8	7.70	98	42340	6.99	ug/l	0.00
Spiked Amount 50.000			Recovery	=	13.98%	
62) 4-Bromofluorobenzene	11.48	95	150206	54.99	ug/l	0.00
Spiked Amount 50.000			Recovery	=	109.98%	
Target Compounds						
16) Acetone	2.34	43	24291	43.595	ug/l	94
20) Methylene Chloride	2.35	84	30437m	15.098	ug/l	

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA F\DATA\VF050919\
Data File : VF062404.D
Acq On : 9 May 2019 19:10
Operator : FY/SY
Sample : K2641-07
Misc : 6.01q/5mL/MSVOA F/SOIL
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_F
ClientSampleId :
CL-02-050719-D

Manual Integrations
APPROVED

MMDadoda
5/10/2019 11:17:20 AM

Quant Time: May 10 03:43:37 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82F050619S.M
Quant Title : SW846 8260
QLast Update : Tue May 07 04:34:57 2019
Response via : Initial Calibration

