


Data Path : Z:\voasrv\HPCHEM1\MSVOA F\Data\VF050919\
Data File : VF062392.D
Acq On : 9 May 2019 12:06
Operator : FY/SY
Sample : BFB
Misc : 5.00µ/5mL/MSVOA F/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_F
ClientSampleId :
BFB

Integration File: RTEINT.P

Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82F050619S.M
Title : SW846 8260
Last Update : Tue May 07 04:34:57 2019



AutoFind: Scans 1110, 1111, 1112; Background Corrected with Scan 1102

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	19.2	22925	PASS
75	95	30	60	44.6	53334	PASS
95	95	100	100	100.0	119501	PASS
96	95	5	9	6.5	7781	PASS
173	174	0.00	2	0.0	0	PASS
174	95	50	100	95.9	114634	PASS
175	174	5	9	8.1	9290	PASS
176	174	95	101	100.4	115122	PASS
177	176	5	9	6.5	7461	PASS