

Data File : VF062469.D
Acq On : 11 May 2019 19:27
Operator : FY/SY
Sample : K2734-17
Misc : 5.93/5mL/MSVOA_F/SOIL
ALS Vial : 20 Sample Multiplier: 1

Instrument :
MSVOA_F
ClientSampleId :
P026-SS014-0206-01

Quant Time: May 12 04:42:30 2019
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82F050619S.M
Quant Title : SW846 8260
QLast Update : Tue May 07 04:34:57 2019
Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.06	168	284292	50.00	ug/l	0.01
34) 1,4-Difluorobenzene	5.78	114	361021	50.00	ug/l	0.01
63) Chlorobenzene-d5	9.91	117	282339	50.00	ug/l	0.01
72) 1,4-Dichlorobenzene-d4	12.61	152	157488	50.00	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.03	65	122099	45.45	ug/l	0.01
Spiked Amount 50.000			Recovery	=	90.90%	
35) Dibromofluoromethane	4.32	113	134678	46.83	ug/l	0.01
Spiked Amount 50.000			Recovery	=	93.66%	
50) Toluene-d8	7.72	98	257302	36.96	ug/l	0.01
Spiked Amount 50.000			Recovery	=	73.92%	
62) 4-Bromofluorobenzene	11.50	95	106226	33.82	ug/l	0.01
Spiked Amount 50.000			Recovery	=	67.64%	

Target Compounds			Qvalue			
16) Acetone	2.34	43	28584	52.593	ug/l #	92

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data File : VF062469.D
Acq On : 11 May 2019 19:27
Operator : FY/SY
Sample : K2734-17
Misc : 5.93/5mL/MSVOA_F/SOIL
ALS Vial : 20 Sample Multiplier: 1

Instrument :
MSVOA_F
ClientSampleId :
P026-SS014-0206-01

Quant Time: May 12 04:42:30 2019
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82F050619S.M
Quant Title : SW846 8260
QLast Update : Tue May 07 04:34:57 2019
Response via : Initial Calibration

