


Data Path : Z:\voasrv\HPCHEM1\MSVOA F\Data\VF051619\
 Data File : VF062617.D
 Acq On : 16 May 2019 9:33
 Operator : FY/SY
 Sample : BFB
 Misc : 5.00µ/5mL/MSVOA F/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_F
 ClientSampleId :
 BFB

Integration File: RTEINT.P

Method : Z:\VOASRV\HPCHEM1\MSVOA_F\METHODS\82F051619S.M
 Title : SW846 8260
 Last Update : Fri May 17 05:49:09 2019



AutoFind: Scans 1111, 1112, 1113; Background Corrected with Scan 1106

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	16.1	56306	PASS
75	95	30	60	38.7	135315	PASS
95	95	100	100	100.0	349954	PASS
96	95	5	9	6.5	22577	PASS
173	174	0.00	2	0.4	1033	PASS
174	95	50	100	81.2	284117	PASS
175	174	5	9	7.4	20920	PASS
176	174	95	101	98.4	279637	PASS
177	176	5	9	6.6	18328	PASS