

Data Path : Z:\VOASRV\HPCHEM1\MSVOA F\DATA\VF052219\
 Data File : VF062704.D
 Acq On : 22 May 2019 14:29
 Operator : FY/SY
 Sample : K2642-05
 Misc : 6.81g/5mL/MSVOA F/SOIL
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_F
 ClientSampleId :
 OK-01-052119

Quant Time: May 23 02:01:03 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_F\METHODS\82F051619S.M
 Quant Title : SW846 8260
 QLast Update : Fri May 17 05:49:09 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.05	168	565635	50.00	ug/l	-0.01
34) 1,4-Difluorobenzene	5.77	114	922254	50.00	ug/l	-0.01
63) Chlorobenzene-d5	9.90	117	542336	50.00	ug/l	-0.01
72) 1,4-Dichlorobenzene-d4	12.60	152	138766	50.00	ug/l	-0.01
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.03	65	160081	40.05	ug/l	0.00
Spiked Amount 50.000			Recovery	=	80.10%	
35) Dibromofluoromethane	4.30	113	297495	42.16	ug/l	-0.01
Spiked Amount 50.000			Recovery	=	84.32%	
50) Toluene-d8	7.71	98	408704	24.26	ug/l	-0.01
Spiked Amount 50.000			Recovery	=	48.52%	
62) 4-Bromofluorobenzene	11.49	95	113445	17.75	ug/l	-0.01
Spiked Amount 50.000			Recovery	=	35.50%	
Target Compounds						
16) Acetone	2.34	43	22684	21.875	ug/l #	79
20) Methylene Chloride	2.32	84	53612	10.315	ug/l	89

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA F\DATA\VF052219\
Data File : VF062704.D
Acq On : 22 May 2019 14:29
Operator : FY/SY
Sample : K2642-05
Misc : 6.81g/5mL/MSVOA F/SOIL
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_F
ClientSampleId :
OK-01-052119

Quant Time: May 23 02:01:03 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_F\METHODS\82F051619S.M
Quant Title : SW846 8260
QLast Update : Fri May 17 05:49:09 2019
Response via : Initial Calibration

