


Data Path : Z:\voasrv\HPCHEM1\MSVOA F\Data\VF053119\
 Data File : VF062800.D
 Acq On : 31 May 2019 9:32
 Operator : FY/SY
 Sample : BFB
 Misc : 5.00µ/5mL/MSVOA F/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_F
 ClientSampleId :
 BFB

Integration File: RTEINT.P

Method : Z:\VOASRV\HPCHEM1\MSVOA_F\METHODS\82F053019S.M
 Title : SW846 8260
 Last Update : Fri May 31 04:18:30 2019



AutoFind: Scans 1112, 1113, 1114; Background Corrected with Scan 1107

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	16.4	40834	PASS
75	95	30	60	39.7	98749	PASS
95	95	100	100	100.0	248448	PASS
96	95	5	9	6.1	15210	PASS
173	174	0.00	2	0.2	488	PASS
174	95	50	100	86.1	213830	PASS
175	174	5	9	7.5	16068	PASS
176	174	95	101	99.2	212068	PASS
177	176	5	9	6.3	13304	PASS