

Data Path : Z:\VOASRV\HPCHEM1\MSVOA F\DATA\VF060419\
 Data File : VF062845.D
 Acq On : 4 Jun 2019 18:10
 Operator : FY/SY
 Sample : K3172-02
 Misc : 3.93g/5mL/MSVOA F/SOIL
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_F
 ClientSampleId :
 FK-SF-01-005-A

Quant Time: Jun 05 06:32:07 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_F\METHODS\82F053019S.M
 Quant Title : SW846 8260
 QLast Update : Fri May 31 04:18:30 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.05	168	576299	50.00	ug/l	-0.02
34) 1,4-Difluorobenzene	5.77	114	1062373	50.00	ug/l	-0.03
63) Chlorobenzene-d5	9.91	117	707450	50.00	ug/l	-0.02
72) 1,4-Dichlorobenzene-d4	12.61	152	196412	50.00	ug/l	-0.01
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.03	65	210259	52.76	ug/l	-0.02
Spiked Amount 50.000			Recovery	=	105.52%	
35) Dibromofluoromethane	4.31	113	19336	2.41	ug/l	-0.02
Spiked Amount 50.000			Recovery	=	4.82%	
50) Toluene-d8	7.71	98	727798	35.85	ug/l	-0.02
Spiked Amount 50.000			Recovery	=	71.70%	
62) 4-Bromofluorobenzene	11.49	95	232955	31.90	ug/l	-0.02
Spiked Amount 50.000			Recovery	=	63.80%	
Target Compounds						
					Qvalue	
16) Acetone	2.35	43	33725	36.975	ug/l	97
68) m/p-Xylenes	10.26	106	12548	1.399	ug/l #	71
70) Styrene	10.89	104	86539	6.464	ug/l	92
80) 1,3,5-Trimethylbenzene	11.89	105	13806	1.232	ug/l	80
84) 1,2,4-Trimethylbenzene	12.27	105	35640	3.354	ug/l	93
95) Naphthalene	14.49	128	275369	45.862	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA F\DATA\VF060419\
Data File : VF062845.D
Acq On : 4 Jun 2019 18:10
Operator : FY/SY
Sample : K3172-02
Misc : 3.93g/5mL/MSVOA F/SOIL
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_F
ClientSampleId :
FK-SF-01-005-A

Quant Time: Jun 05 06:32:07 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_F\METHODS\82F053019S.M
Quant Title : SW846 8260
QLast Update : Fri May 31 04:18:30 2019
Response via : Initial Calibration

