

Data Path : Z:\VOASRV\HPCHEM1\MSVOA F\DATA\VF060519\
 Data File : VF062869.D
 Acq On : 5 Jun 2019 17:02
 Operator : FY/SY
 Sample : K3172-02RE
 Misc : 4.22g/5mL/MSVOA F/SOIL
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_F
 ClientSampleId :
 FK-SF-01-005-ARE

Quant Time: Jun 06 08:23:26 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_F\METHODS\82F053019S.M
 Quant Title : SW846 8260
 QLast Update : Fri May 31 04:18:30 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.08	168	614241	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	5.80	114	1068270	50.00	ug/l	0.00
63) Chlorobenzene-d5	9.93	117	783471	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.63	152	208353	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.06	65	202008	47.56	ug/l	0.00
Spiked Amount 50.000			Recovery	=	95.12%	
35) Dibromofluoromethane	4.33	113	20484	2.54	ug/l	0.00
Spiked Amount 50.000			Recovery	=	5.08%	
50) Toluene-d8	7.73	98	742250	36.36	ug/l	0.00
Spiked Amount 50.000			Recovery	=	72.72%	
62) 4-Bromofluorobenzene	11.50	95	225608	30.72	ug/l	0.00
Spiked Amount 50.000			Recovery	=	61.44%	
Target Compounds						
16) Acetone	2.37	43	26191	26.941	ug/l	96
68) m/p-Xylenes	10.28	106	12667	1.275	ug/l	94
70) Styrene	10.90	104	72149	4.866	ug/l	92
80) 1,3,5-Trimethylbenzene	11.91	105	12823	1.079	ug/l	93
84) 1,2,4-Trimethylbenzene	12.28	105	36767	3.262	ug/l	87
95) Naphthalene	14.50	128	298014	46.789	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA F\DATA\VF060519\
Data File : VF062869.D
Acq On : 5 Jun 2019 17:02
Operator : FY/SY
Sample : K3172-02RE
Misc : 4.22g/5mL/MSVOA F/SOIL
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_F
ClientSampleId :
FK-SF-01-005-ARE

Quant Time: Jun 06 08:23:26 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_F\METHODS\82F053019S.M
Quant Title : SW846 8260
QLast Update : Fri May 31 04:18:30 2019
Response via : Initial Calibration

