


Data Path : Z:\voasrv\HPCHEM1\MSVOA F\Data\VF060519\
Data File : VF062858.D
Acq On : 5 Jun 2019 9:42
Operator : FY/SY
Sample : BFB
Misc : 5.00µ/5mL/MSVOA F/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_F
ClientSampleId :
BFB

Integration File: RTEINT.P

Method : Z:\VOASRV\HPCHEM1\MSVOA_F\METHODS\82F053019S.M
Title : SW846 8260
Last Update : Fri May 31 04:18:30 2019



AutoFind: Scans 1111, 1112, 1113; Background Corrected with Scan 1103

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	16.1	40896	PASS
75	95	30	60	39.6	100416	PASS
95	95	100	100	100.0	253347	PASS
96	95	5	9	6.9	17423	PASS
173	174	0.00	2	0.0	0	PASS
174	95	50	100	82.1	207893	PASS
175	174	5	9	6.7	14006	PASS
176	174	95	101	99.1	205930	PASS
177	176	5	9	6.5	13348	PASS