

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_F\DATA\VF061818\
 Data File : VF059186.D
 Acq On : 18 Jun 2018 20:07
 Operator : VA/AP
 Sample : J3556-01RE
 Misc : 5.99g/5mL/MSVOA-F/SOIL
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 MSVOA_F
 ClientSampleId :
 MW286-18-SRI3RE

Quant Time: Jun 19 02:58:00 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_F\METHODS\82F060618S.M
 Quant Title : SW846 8260
 QLast Update : Thu Jun 07 10:52:00 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.04	168	226010	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	5.76	114	361160	50.00	ug/l	0.00
63) Chlorobenzene-d5	9.91	117	303015	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.63	152	168119	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	5.01	65	156638	57.15	ug/l	0.00
Spiked Amount 50.000			Recovery	=	114.30%	
35) Dibromofluoromethane	4.28	113	141210	55.72	ug/l	-0.02
Spiked Amount 50.000			Recovery	=	111.44%	
50) Toluene-d8	7.70	98	325399	43.71	ug/l	0.00
Spiked Amount 50.000			Recovery	=	87.42%	
62) 4-Bromofluorobenzene	11.51	95	190360	49.85	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.70%	


Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
3) Chloromethane	1.11	50	2071	1.045	ug/l #	72
11) Tert butyl alcohol	2.67	59	447	2.545	ug/l #	1
13) Acrolein	2.09	56	2801	18.426	ug/l #	1
16) Acetone	2.33	43	39708	60.674	ug/l	99
18) Methyl Acetate	2.48	43	8489	7.545	ug/l #	79
20) Methylene Chloride	2.27	84	5095	3.251	ug/l #	94
25) 2-Butanone	4.51	43	6532	6.438	ug/l	99
29) Tetrahydrofuran	4.25	42	1330	2.975	ug/l #	46
39) Methylcyclohexane	5.65	83	3823	1.071	ug/l #	81
49) 1,4-Dioxane	6.99	88	83	18.764	ug/l #	1
51) 4-Methyl-2-Pentanone	8.42	43	63490	31.977	ug/l #	44
52) Toluene	7.77	92	14250	2.435	ug/l	91
56) Ethyl methacrylate	8.82	69	23316	8.722	ug/l	94
58) 2-Chloroethyl Vinyl ether	7.77	63	2152	11.258	ug/l	100
59) 2-Hexanone	9.67	43	8292	5.280	ug/l #	26
73) Isopropylbenzene	11.23	105	16463	1.414	ug/l	85
74) N-amyl acetate	11.49	43	23905	5.749	ug/l #	11
75) 1,1,2,2-Tetrachloroethane	11.76	83	40063	14.518	ug/l #	39
83) tert-Butylbenzene	12.21	119	43049	4.500	ug/l	83
84) 1,2,4-Trimethylbenzene	12.38	105	144220	14.448	ug/l #	28
85) sec-Butylbenzene	12.38	105	144294	11.101	ug/l #	75
90) Hexachloroethane	13.00	117	3932	1.496	ug/l #	24
92) 1,2-Dibromo-3-Chloropropan	13.70	75	1124	2.077	ug/l #	40
93) 1,2,4-Trichlorobenzene	14.34	180	17223	4.726	ug/l #	1
95) Naphthalene	14.52	128	26055	3.347	ug/l #	1
96) 1,2,3-Trichlorobenzene	14.70	180	23782	7.062	ug/l #	1


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA_F\DATA\VF061818\
Data File : VF059186.D
Acq On : 18 Jun 2018 20:07
Operator : VA/AP
Sample : J3556-01RE
Misc : 5.99g/5mL/MSVOA-F/SOIL
ALS Vial : 20 Sample Multiplier: 1

Instrument :
MSVOA_F
ClientSampleId :
MW286-18-SRI3RE

Quant Time: Jun 19 02:58:00 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_F\METHODS\82F060618S.M
Quant Title : SW846 8260
QLast Update : Thu Jun 07 10:52:00 2018
Response via : Initial Calibration





#1

Pentafluorobenzene

Concen: 50.000 ug/l

RT: 5.04 min Scan# 457

Delta R.T. -0.01 min

Lab File: VF059186.D

Acq: 18 Jun 2018 20:07

Instrument :

MSVOA_F

ClientSampleId :


MW286-18-SRI3RE

Tgt Ion:168 Resp: 226010

Ion Ratio Lower Upper




168 100

99 64.4 52.9 79.3



Abundance Ion 168.00 (167.70 to 168.70): V

Ion 99.00 (98.70 to 99.70): VF0



#3

Chloromethane

Concen: 1.045 ug/l

RT: 1.11 min Scan# 58

Delta R.T. 0.02 min

Lab File: VF059186.D


Acq: 18 Jun 2018 20:07

Tgt Ion: 50 Resp: 2071

Ion Ratio Lower Upper



50 100


52 18.1 27.3 40.9#



Abundance Ion 50.00 (49.70 to 50.70): VF0

Ion 52.00 (51.70 to 52.70): VF0









#11
Tert butyl alcohol
Concen: 2.545 ug/l
RT: 2.67 min Scan# 217
Delta R.T. -0.01 min
Lab File: VF059186.D
Acq: 18 Jun 2018 20:07




Instrument :
MSVOA_F
ClientSampleId :
MW286-18-SRI3RE


Tgt Ion: 59 Resp: 447
Ion Ratio Lower Upper
59 100
57 52.0 8.7 13.1#



#13
Acrolein
Concen: 18.426 ug/l
RT: 2.09 min Scan# 158
Delta R.T. 0.00 min
Lab File: VF059186.D
Acq: 18 Jun 2018 20:07

Tgt Ion: 56 Resp: 2801
Ion Ratio Lower Upper
56 100
55 298.3 60.0 90.0#









#16
Acetone
Concen: 60.674 ug/l
RT: 2.33 min Scan# 182
Delta R.T. -0.01 min
Lab File: VF059186.D
Acq: 18 Jun 2018 20:07




Instrument :
MSVOA_F
ClientSampleId :
MW286-18-SRI3RE


Tgt Ion: 43 Resp: 39708
Ion Ratio Lower Upper
43 100
58 18.5 14.4 21.6



#18
Methyl Acetate
Concen: 7.545 ug/l
RT: 2.48 min Scan# 197
Delta R.T. 0.03 min
Lab File: VF059186.D
Acq: 18 Jun 2018 20:07

Tgt Ion: 43 Resp: 8489
Ion Ratio Lower Upper
43 100
74 5.1 10.9 16.3#









#20
Methylene Chloride
Concen: 3.251 ug/l
RT: 2.27 min Scan# 176
Delta R.T. -0.01 min
Lab File: VF059186.D
Acq: 18 Jun 2018 20:07




Instrument :
MSVOA_F
ClientSampleId :
MW286-18-SRI3RE


Tgt Ion:	84	Resp:	5095
Ion	Ratio	Lower	Upper
84	100		
49	149.7	121.0	181.4
51	62.5	35.8	53.6#
86	63.2	53.2	79.8



#25
2-Butanone
Concen: 6.438 ug/l
RT: 4.51 min Scan# 403
Delta R.T. 0.00 min
Lab File: VF059186.D
Acq: 18 Jun 2018 20:07

Tgt Ion:	43	Resp:	6532
Ion	Ratio	Lower	Upper
43	100		
72	19.7	16.1	24.1





#29

Tetrahydrofuran

Concen: 2.975 ug/l

RT: 4.25 min Scan# 377

Delta R.T. 0.00 min

Lab File: VF059186.D

Acq: 18 Jun 2018 20:07

Instrument :

MSVOA_F

ClientSampleId :

MW286-18-SRI3RE





Tgt Ion: 42 Resp: 1330

Ion Ratio Lower Upper

42 100

72 0.0 30.5 45.7#

71 11.7 31.5 47.3#



#33

1,2-Dichloroethane-d4

Concen: 57.145 ug/l

RT: 5.01 min Scan# 454

Delta R.T. -0.01 min

Lab File: VF059186.D




Acq: 18 Jun 2018 20:07


Tgt Ion: 65 Resp: 156638

Ion Ratio Lower Upper

65 100

67 45.4 0.0 97.0






#34
1,4-Difluorobenzene
Concen: 50.000 ug/l
RT: 5.76 min Scan# 530
Delta R.T. -0.01 min
Lab File: VF059186.D
Acq: 18 Jun 2018 20:07

Instrument :
MSVOA_F
ClientSampleId :
MW286-18-SRI3RE




Tgt Ion	Ratio	Lower	Upper
114	100		
63	20.9	0.0	41.6
88	18.3	0.0	38.8



Abundance Ion 114.00 (113.70 to 114.70): V


Ion 63.00 (62.70 to 63.70): VF0

Ion 88.00 (87.70 to 88.70): VF0



#35
Dibromofluoromethane
Concen: 55.723 ug/l
RT: 4.28 min Scan# 380
Delta R.T. -0.02 min
Lab File: VF059186.D
Acq: 18 Jun 2018 20:07



Tgt Ion	Ratio	Lower	Upper
113	100		
111	103.1	81.3	121.9
192	15.9	12.7	19.1




Abundance Ion 113.00 (112.70 to 113.70): V

Ion 111.00 (110.70 to 111.70): V

Ion 192.00 (191.70 to 192.70): V









#39
Methylcyclohexane
Concen: 1.071 ug/l
RT: 5.65 min Scan# 519
Delta R.T. 0.02 min
Lab File: VF059186.D
Acq: 18 Jun 2018 20:07




Instrument :
MSVOA_F
ClientSampleId :
MW286-18-SRI3RE


Tgt Ion: 83 Resp: 3823
Ion Ratio Lower Upper
83 100
55 84.9 71.0 106.4
98 16.2 38.1 57.1#



#49
1,4-Dioxane
Concen: 18.764 ug/l
RT: 6.99 min Scan# 655
Delta R.T. 0.12 min
Lab File: VF059186.D
Acq: 18 Jun 2018 20:07

Tgt Ion: 88 Resp: 83
Ion Ratio Lower Upper
88 100
43 1321.3 47.8 71.8#
58 0.0 59.4 89.0#









#50
Toluene-d8
Concen: 43.708 ug/l
RT: 7.70 min Scan# 727
Delta R.T. -0.01 min
Lab File: VF059186.D
Acq: 18 Jun 2018 20:07




Instrument :
MSVOA_F
ClientSampleId :
MW286-18-SRI3RE


Tgt Ion: 98 Resp: 325399
Ion Ratio Lower Upper
98 100
100 64.6 55.6 83.4



#51
4-Methyl-2-Pentanone
Concen: 31.977 ug/l
RT: 8.42 min Scan# 800
Delta R.T. 0.01 min
Lab File: VF059186.D
Acq: 18 Jun 2018 20:07

Tgt Ion: 43 Resp: 63490
Ion Ratio Lower Upper
43 100
58 3.5 29.4 44.0#





#52

Toluene

Concen: 2.435 ug/l

RT: 7.77 min Scan# 734

Delta R.T. -0.01 min

Lab File: VF059186.D

Acq: 18 Jun 2018 20:07

Instrument :

MSVOA_F

ClientSampleId :





MW286-18-SRI3RE

Tgt Ion: 92 Resp: 14250

Ion Ratio Lower Upper

92 100

91 154.5 132.9 199.3



#56

Ethyl methacrylate

Concen: 8.722 ug/l

RT: 8.82 min Scan# 841

Delta R.T. 0.06 min

Lab File: VF059186.D

Acq: 18 Jun 2018 20:07




Tgt Ion: 69 Resp: 23316


Ion Ratio Lower Upper

69 100

41 65.7 56.9 85.3

39 43.6 38.0 57.0









#58
 2-Chloroethyl Vinyl ether
 Concen: 11.258 ug/l
 RT: 7.77 min Scan# 734
 Delta R.T. -0.01 min
 Lab File: VF059186.D
 Acq: 18 Jun 2018 20:07




Instrument :
 MSVOA_F
 ClientSampleId :
 MW286-18-SRI3RE


Tgt Ion: 63 Resp: 2152
 Ion Ratio Lower Upper
 63 100
 106 0.0 0.0 0.0



#59
 2-Hexanone
 Concen: 5.280 ug/l
 RT: 9.67 min Scan# 927
 Delta R.T. 0.04 min
 Lab File: VF059186.D
 Acq: 18 Jun 2018 20:07

Tgt Ion: 43 Resp: 8292
 Ion Ratio Lower Upper
 43 100
 58 0.0 26.3 78.8#





#62

4-Bromofluorobenzene

Concen: 49.852 ug/l

RT: 11.51 min Scan# 1113

Delta R.T. -0.01 min

Lab File: VF059186.D

Acq: 18 Jun 2018 20:07

Instrument :

MSVOA_F

ClientSampleId :

MW286-18-SRI3RE


Tgt Ion: 95 Resp: 190360

Ion Ratio Lower Upper

95 100

174 72.0 0.0 157.8




176 69.1 0.0 155.8



Abundance Ion 95.00 (94.70 to 95.70): VF0

Ion 174.00 (173.70 to 174.70): V

Ion 176.00 (175.70 to 176.70): V



#63

Chlorobenzene-d5

Concen: 50.000 ug/l

RT: 9.91 min Scan# 951

Delta R.T. -0.01 min

Lab File: VF059186.D

Acq: 18 Jun 2018 20:07


Tgt Ion: 117 Resp: 303015

Ion Ratio Lower Upper

117 100

82 61.4 45.2 67.8



119 33.3 27.5 41.3




Abundance Ion 117.00 (116.70 to 117.70): V

Ion 82.00 (81.70 to 82.70): VF0

Ion 119.00 (118.70 to 119.70): V





#72

1,4-Dichlorobenzene-d4

Concen: 50.000 ug/l

RT: 12.63 min Scan# 1227

Delta R.T. -0.01 min

Lab File: VF059186.D

Acq: 18 Jun 2018 20:07

Instrument :

MSVOA_F

ClientSampleId :

MW286-18-SRI3RE


Tgt Ion:152 Resp: 168119

Ion Ratio Lower Upper

152 100

115 57.5 28.5 85.5




150 154.9 0.0 323.8



Abundance Ion 152.00 (151.70 to 152.70): V

Ion 115.00 (114.70 to 115.70): V

Ion 150.00 (149.70 to 150.70): V



#73

Isopropylbenzene

Concen: 1.414 ug/l

RT: 11.23 min Scan# 1085

Delta R.T. 0.00 min

Lab File: VF059186.D


Acq: 18 Jun 2018 20:07

Tgt Ion:105 Resp: 16463

Ion Ratio Lower Upper



105 100


120 33.1 12.7 38.1



Abundance Ion 105.00 (104.70 to 105.70): V

Ion 120.00 (119.70 to 120.70): V





#74

N-ethyl acetate

Concen: 5.749 ug/l

RT: 11.49 min Scan# 1111

Delta R.T. 0.02 min

Lab File: VF059186.D

Acq: 18 Jun 2018 20:07

Instrument :

MSVOA_F

ClientSampleId :

MW286-18-SRI3RE

Tgt Ion: 43 Resp: 23905


Ion Ratio Lower Upper

43 100

70 47.7 29.2 43.8#

55 150.1 15.8 23.6#

61 6.7 21.3 31.9#






Abundance Ion 43.00 (42.70 to 43.70): VF0

Ion 70.00 (69.70 to 70.70): VF0

Ion 55.00 (54.70 to 55.70): VF0

Ion 61.00 (60.70 to 61.70): VF0



#75

1,1,2,2-Tetrachloroethane

Concen: 14.518 ug/l

RT: 11.76 min Scan# 1139

Delta R.T. -0.03 min

Lab File: VF059186.D

Acq: 18 Jun 2018 20:07


Tgt Ion: 83 Resp: 40063

Ion Ratio Lower Upper

83 100

131 0.0 4.5 13.7#



85 13.5 32.8 98.3#




Abundance Ion 83.00 (82.70 to 83.70): VF0

Ion 131.00 (130.70 to 131.70): V

Ion 85.00 (84.70 to 85.70): VF0









#83
 tert-Butylbenzene
 Concen: 4.500 ug/l
 RT: 12.21 min Scan# 1185
 Delta R.T. 0.00 min
 Lab File: VF059186.D
 Acq: 18 Jun 2018 20:07




Instrument :
 MSVOA_F
 ClientSampleId :
 MW286-18-SRI3RE


Tgt Ion:119 Resp: 43049
 Ion Ratio Lower Upper
 119 100
 91 76.9 30.4 91.3
 134 24.3 11.2 33.5



#84
 1,2,4-Trimethylbenzene
 Concen: 14.448 ug/l
 RT: 12.38 min Scan# 1202
 Delta R.T. 0.09 min
 Lab File: VF059186.D
 Acq: 18 Jun 2018 20:07

Tgt Ion:105 Resp: 144220
 Ion Ratio Lower Upper
 105 100
 120 0.2 24.7 74.1#





#85

sec-Butylbenzene

Concen: 11.101 ug/l

RT: 12.38 min Scan# 1202

Delta R.T. -0.01 min

Lab File: VF059186.D

Acq: 18 Jun 2018 20:07

Instrument :

MSVOA_F

ClientSampleId :





MW286-18-SRI3RE

Tgt Ion:105 Resp: 144294

Ion Ratio Lower Upper

105 100

134 30.4 9.6 28.6#



#90

Hexachloroethane

Concen: 1.496 ug/l

RT: 13.00 min Scan# 1265

Delta R.T. 0.01 min

Lab File: VF059186.D




Acq: 18 Jun 2018 20:07


Tgt Ion:117 Resp: 3932

Ion Ratio Lower Upper

117 100

201 0.0 27.5 82.4#





#92

1,2-Dibromo-3-Chloropropane

Concen: 2.077 ug/l

RT: 13.70 min Scan# 1336

Delta R.T. -0.01 min

Lab File: VF059186.D

Acq: 18 Jun 2018 20:07

Instrument :

MSVOA_F

ClientSampleId :

MW286-18-SRI3RE





Tgt Ion: 75 Resp: 1124

Ion Ratio Lower Upper

75 100

155 60.2 38.0 114.0

157 0.0 45.1 135.5#



#93

1,2,4-Trichlorobenzene

Concen: 4.726 ug/l

RT: 14.34 min Scan# 1401

Delta R.T. 0.07 min

Lab File: VF059186.D

Acq: 18 Jun 2018 20:07




Tgt Ion: 180 Resp: 17223


Ion Ratio Lower Upper

180 100

182 64.6 49.3 147.9

145 441.8 14.9 44.9#





#95

Naphthalene

Concen: 3.347 ug/l

RT: 14.52 min Scan# 1419

Delta R.T. -0.01 min

Lab File: VF059186.D

Acq: 18 Jun 2018 20:07

Instrument :

MSVOA_F

ClientSampleId :

MW286-18-SRI3RE


Tgt Ion:128 Resp: 26055

Ion Ratio Lower Upper

128 100

127 43.0 9.9 14.9#




129 81.5 8.7 13.1#



Abundance Ion 128.00 (127.70 to 128.70): V

Ion 127.00 (126.70 to 127.70): V

Ion 129.00 (128.70 to 129.70): V



#96

1,2,3-Trichlorobenzene

Concen: 7.062 ug/l

RT: 14.70 min Scan# 1437

Delta R.T. 0.03 min

Lab File: VF059186.D

Acq: 18 Jun 2018 20:07


Tgt Ion:180 Resp: 23782

Ion Ratio Lower Upper

180 100

182 54.7 48.9 146.6

145 375.1 17.3 51.7#



Abundance Ion 180.00 (179.70 to 180.70): V

Ion 182.00 (181.70 to 182.70): V

Ion 145.00 (144.70 to 145.70): V

