

Data Path : Z:\VOASRV\HPCHEM1\MSVOA F\DATA\VF061818\
 Data File : VF059176.D
 Acq On : 18 Jun 2018 15:29
 Operator : VA/AP
 Sample : J3566-03
 Misc : 5.47g/5mL/MSVOA-F/SOIL
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_F
 ClientSampleId :
 TP-8-S

Manual Integrations
 APPROVED

apatel
 6/19/2018 3:47:19 PM

Quant Time: Jun 19 03:58:24 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_F\METHODS\82F060618S.M
 Quant Title : SW846 8260
 QLast Update : Thu Jun 07 10:52:00 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.04	168	290389	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	5.76	114	424413	50.00	ug/l	0.00
63) Chlorobenzene-d5	9.92	117	403387	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.63	152	243931	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.01	65	187933	53.36	ug/l	0.00
Spiked Amount 50.000			Recovery	=	106.72%	
35) Dibromofluoromethane	4.28	113	166468	55.90	ug/l	-0.02
Spiked Amount 50.000			Recovery	=	111.80%	
50) Toluene-d8	7.70	98	427961	48.92	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.84%	
62) 4-Bromofluorobenzene	11.50	95	216624	48.28	ug/l	0.00
Spiked Amount 50.000			Recovery	=	96.56%	
Target Compounds						
16) Acetone	2.34	43	14981	17.816	ug/l #	88
20) Methylene Chloride	2.32	84	12941m	6.428	ug/l	

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA F\DATA\VF061818\
Data File : VF059176.D
Acq On : 18 Jun 2018 15:29
Operator : VA/AP
Sample : J3566-03
Misc : 5.47g/5mL/MSVOA-F/SOIL
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_F
ClientSampleId :
TP-8-S

Quant Time: Jun 19 03:58:24 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_F\METHODS\82F060618S.M
Quant Title : SW846 8260
QLast Update : Thu Jun 07 10:52:00 2018
Response via : Initial Calibration

Manual Integrations
APPROVED

apatel
6/19/2018 3:47:19 PM

